

7471A


LB 88313

INSTRUMENT ID: C11

Linear

 μ Abs.:

46915



A= 0.0000e+000

B= 2.1256e-004

C= -5.1643e-002

Rho= 0.9998077

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.048	-0.048	19	0.000	19				
0.2	0.200	0.150	-0.050	951	0.0 %	951				
2.5	2.500	2.542	0.042	12202	0.0 %	12202				
5.0	5.000	5.136	0.136	24408	0.0 %	24408				
7.5	7.500	7.498	-0.002	35520	0.0 %	35520	0			
10.0	10.000	9.920	-0.080	46915	0.0 %	46915				

LB88313 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 23 Jun 2017 17:00:41					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	19	-	0.0000	7471A	PPB	23 Jun 2017 17:06:25	S	Std
0.2 - 1	50.2	951	-	0.2000	7471A	PPB	23 Jun 2017 17:08:43	S	Std
2.5 - 1	52.5	12202	-	2.5000	7471A	PPB	23 Jun 2017 17:11:00	S	Std
5.0 - 1	55	24408	-	5.0000	7471A	PPB	23 Jun 2017 17:13:18	S	Std
7.5 - 1	57.5	35520	-	7.5000	7471A	PPB	23 Jun 2017 17:15:35	S	Std
10.0 - 1	510	46915	-	10.0000	7471A	PPB	23 Jun 2017 17:17:52	S	Std
ICV13 - 1	ICV13	18796	3.9436	-	7471A	PPB	23 Jun 2017 17:20:42	U	SMPL
ICB13 - 1	ICB13	-104	-0.0737	-	7471A	PPB	23 Jun 2017 17:22:57	U	SMPL
CCV34 - 1	CCV34	24364	5.1271	-	7471A	PPB	23 Jun 2017 17:25:13	U	SMPL
CCB34 - 1	CCB34	-97	-0.0723	-	7471A	PPB	23 Jun 2017 17:27:29	U	SMPL
CRA - 1	CRA	1249	0.2138	-	7471A	PPB	23 Jun 2017 17:29:45	U	SMPL
HighStd - 1	HighStd	46564	9.8458	-	7471A	PPB	23 Jun 2017 17:32:00	U	SMPL
ChkStd - 1	ChkStd	32745	6.9085	-	7471A	PPB	23 Jun 2017 17:34:17	U	SMPL
I3852-01 - 1	G-33-0-1.5	3644	0.7229	-	7471A	PPB	23 Jun 2017 17:36:35	U	SMPL
I3852-01DUP - 1	G-33-0-1.5DUP	3406	0.6723	-	7471A	PPB	23 Jun 2017 17:38:52	U	SMPL
I3852-01MS - 1	G-33-0-1.5MS	20756	4.3602	-	7471A	PPB	23 Jun 2017 17:41:09	U	SMPL
I3852-01MSD - 1	G-33-0-1.5MSD	20858	4.3818	-	7471A	PPB	23 Jun 2017 17:45:09	U	SMPL
I3852-01LX5 - 1		631	0.0825	-	7471A	PPB	23 Jun 2017 17:47:26	U	SMPL
I3852-01A - 1		23011	4.8395	-	7471A	PPB	23 Jun 2017 17:49:44	U	SMPL
CCV35 - 1	G-33-0-1.5	23847	5.0172	-	7471A	PPB	23 Jun 2017 17:52:01	U	SMPL
CCB35 - 1	CCB35	-82	-0.0691	-	7471A	PPB	23 Jun 2017 17:54:17	U	SMPL
I3844-01 - 1	FK-PS-01-008	1038	0.1690	-	7471A	PPB	23 Jun 2017 17:56:33	U	SMPL
I3844-01DUP - 1	FK-PS-01-008DUP	982	0.1571	-	7471A	PPB	23 Jun 2017 17:58:49	U	SMPL
I3844-01MS - 1	FK-PS-01-008MS	21850	4.5927	-	7471A	PPB	23 Jun 2017 18:01:04	U	SMPL
I3844-01MSD - 1	FK-PS-01-008MSD	21716	4.5642	-	7471A	PPB	23 Jun 2017 18:03:20	U	SMPL
I3844-01LX5 - 1		137	-0.0225	-	7471A	PPB	23 Jun 2017 18:05:36	U	SMPL
I3844-01A - 1		19954	4.1897	-	7471A	PPB	23 Jun 2017 18:07:52	U	SMPL
I3849-01 - 1	SB-01	52587	11.1260	-	7471A	PPB	23 Jun 2017 18:10:09	U	SMPL
I3849-02 - 1	SB-02	3149	0.6177	-	7471A	PPB	23 Jun 2017 18:12:26	U	SMPL
I3868-01 - 1	CL-01-062217-A	2596	0.5002	-	7471A	PPB	23 Jun 2017 18:14:45	U	SMPL
I3868-04 - 1	CL-01-062217-B	2007	0.3750	-	7471A	PPB	23 Jun 2017 18:17:03	U	SMPL
CCV36 - 1	CCV36	23912	5.0310	-	7471A	PPB	23 Jun 2017 18:19:19	U	SMPL
CCB36 - 1	CCB36	-44	-0.0610	-	7471A	PPB	23 Jun 2017 18:21:36	U	SMPL
I3868-07 - 1	CL-01-062217-C	2302	0.4377	-	7471A	PPB	23 Jun 2017 18:23:52	U	SMPL
I3868-10 - 1	CL-01-062217-D	2955	0.5765	-	7471A	PPB	23 Jun 2017 18:26:08	U	SMPL
I3868-13 - 1	CL-01-062217-E	8205	1.6924	-	7471A	PPB	23 Jun 2017 18:28:25	U	SMPL
I3872-01 - 1	HR-05-062217-A	5276	1.0698	-	7471A	PPB	23 Jun 2017 18:30:41	U	SMPL
I3872-01DUP - 1	HR-05-062217-ADUP	5273	1.0692	-	7471A	PPB	23 Jun 2017 18:32:58	U	SMPL
I3872-01MS - 1	HR-05-062217-AMS	22184	4.6637	-	7471A	PPB	23 Jun 2017 18:35:15	U	SMPL
I3872-01MSD - 1	HR-05-062217-AMSD	22321	4.6928	-	7471A	PPB	23 Jun 2017 18:37:31	U	SMPL
I3872-01LX5 - 1		957	0.1518	-	7471A	PPB	23 Jun 2017 18:39:49	U	SMPL
I3872-01A - 1		24325	5.1188	-	7471A	PPB	23 Jun 2017 18:42:06	U	SMPL
I3872-04 - 1	HR-05-062217-B	10912	2.2678	-	7471A	PPB	23 Jun 2017 18:44:24	U	SMPL
CCV37 - 1	CCV37	23667	4.9789	-	7471A	PPB	23 Jun 2017 18:46:40	U	SMPL
CCB37 - 1	CCB37	-58	-0.0640	-	7471A	PPB	23 Jun 2017 18:48:56	U	SMPL
I3872-07 - 1	HR-05-062217-C	2385	0.4553	-	7471A	PPB	23 Jun 2017 18:51:12	U	SMPL
I3849-01DLX2 - 1		26367	5.5528	-	7471A	PPB	23 Jun 2017 18:53:29	U	SMPL
PB100061BL - 1	PBS	-90	-0.0708	-	7471A	PPB	23 Jun 2017 18:55:45	U	SMPL
PB100061BS - 1	LCSS	17011	3.5641	-	7471A	PPB	23 Jun 2017 18:58:01	U	SMPL
CCV38 - 1	CCV38	23399	4.9219	-	7471A	PPB	23 Jun 2017 19:00:17	U	SMPL
CCB38 - 1	CCB38	-36	-0.0593	-	7471A	PPB	23 Jun 2017 19:02:34	U	SMPL