



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88315

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/23/2017	12:30
2	CAL2	1	Water			4.00	06/23/2017	12:31
3	CAL3	1	Water			10.02	06/23/2017	12:33
4	ICV	1	Water			7.00	06/23/2017	12:34
5	CCV1	1	Water			2.01	06/23/2017	12:35
6	I3879-01	1	Solid	20.02	20	7.41	06/23/2017	12:36
7	I3879-01DUP	1	Solid	20.03	20	7.42	06/23/2017	12:37
8	CCV2	1	Water			12.01	06/23/2017	12:39

WORKLIST(Hardcopy Internal Chain)

2B88315

WorkList Name : PH I3879

WorkList ID : 99997

Date : 6/23/2017 12:03:59 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3879-01	pH	Cool 4 deg C	NAT109	N43	I16-005-MH	06/23/2017	9045C

Date/Time 06/23/17 12:03 PM
 Received by: JA
 Relinquished by: [Signature]

Date/Time 06/23/17 12:03 PM
 Received by: CP
 Relinquished by: JP