



Analytical Summary Report

Analysis Method: 1498

Reviewed By: JIGNESH

Parameter: Redox Potential

Supervisor Review By: apatel

Run Number: LB88317

Reagent/Standard	Lot/Log #
ZOBELLS SOLUTION ORP STD 500ML	W2189

Seq	LabID	ClientID	True value	DL	Initial Wt (g)	Final Vol (mL)	Temp (°C)	Result (mV)	Anal Date	Anal Time
1	LB88317BL	LB88317BL		1			20.1	0.00	06/23/2017	17:15
2	LB88317BS	LB88317BS	228	1			20.2	225.70	06/23/2017	17:16
3	I3900-01	MW-1		1			20.2	165.60	06/23/2017	17:17
4	I3900-01DUP	MW-1DUP		1			20.1	167.20	06/23/2017	17:19
5	I3900-02	MW-3		1			20.0	162.70	06/23/2017	17:22
6	I3900-03	MW-4		1			20.2	159.20	06/23/2017	17:24
7	I3900-04	MW-2		1			20.4	155.10	06/23/2017	17:25

LB 88317

WORKLIST(Hardcopy Internal Chain)

WorkList Name : RP I3900

WorkList ID : 100021

Date : 6/23/2017 4:59:52 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3900-01	Redox Potential	Cool 4 deg C	LOCK01	N61	MW-1	06/22/2017	1498
	Water	I3900-02	Redox Potential	Cool 4 deg C	LOCK01	N61	MW-3	06/22/2017	1498
	Water	I3900-03	Redox Potential	Cool 4 deg C	LOCK01	N61	MW-4	06/22/2017	1498
	Water	I3900-04	Redox Potential	Cool 4 deg C	LOCK01	N61	MW-2	06/22/2017	1498

Date/Time

Received by:

Relinquished by:

06/23/17 4:59 PM

JP

CR

Date/Time

Received by:

Relinquished by:

06/23/17

JP

JP