


7470A

LB88319
INSTRUMENT ID: CV1

Linear

 μ Abs.:

39633



A= 0.0000e+000

B= 2.5354e-004

C= -3.6573e-002

Rho= 0.9999863

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	145	0.000	145				
0.2	0.200	0.190	-0.010	895	0.0 %	895				
2.5	2.500	2.507	0.007	10033	0.0 %	10033				
5.0	5.000	5.026	0.026	19968	0.0 %	19968				
7.5	7.500	7.464	-0.036	29585	0.0 %	29585				
10.0	10.000	10.012	0.012	39633	0.0 %	39633				

LB88319 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Jun 2017 19:18:05

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	145	-	0.0007470A	PPB		23 Jun 2017 19:18:15	S	Std
0.2 - 1	50.2	895	-	0.20007470A	PPB		23 Jun 2017 19:20:24	S	Std
2.5 - 1	52.5	10033	-	2.50007470A	PPB		23 Jun 2017 19:22:32	S	Std
5.0 - 1	55	19968	-	5.00007470A	PPB		23 Jun 2017 19:24:40	S	Std
7.5 - 1	57.5	29585	-	7.50007470A	PPB		23 Jun 2017 19:26:48	S	Std
10.0 - 1	58	39633	-	10.00007470A	PPB		23 Jun 2017 19:28:56	S	Std
ICV14 - 1	ICV14	15465	3.8844	-7470A	PPB		23 Jun 2017 19:31:35	U	SMPL
ICB14 - 1	ICB14	-31	-0.0444	-7470A	PPB		23 Jun 2017 19:33:41	U	SMPL
CCV39 - 1	CCV39	19689	4.9553	-7470A	PPB		23 Jun 2017 19:35:47	U	SMPL
CCB39 - 1	CCB39	53	-0.0231	-7470A	PPB		23 Jun 2017 19:37:54	U	SMPL
CRA - 1	CRA	795	0.1650	-7470A	PPB		23 Jun 2017 19:40:00	U	SMPL
HighStd - 1	HighStd	38747	9.7873	-7470A	PPB		23 Jun 2017 19:42:07	U	SMPL
ChkStd - 1	ChkStd	27739	6.9963	-7470A	PPB		23 Jun 2017 19:44:13	U	SMPL
PB100049BL - 1	PBW	-199	-0.0870	-7470A	PPB		23 Jun 2017 19:46:22	U	SMPL
PB100049BS - 1	LCSW	15835	3.9782	-7470A	PPB		23 Jun 2017 19:48:31	U	SMPL
PB100049TB - 1	PB100003TB	-138	-0.0716	-7470A	PPB		23 Jun 2017 19:50:39	U	SMPL
I3808-05 - 1	DRUM-001-003-004-COMP	-69	-0.0541	-7470A	PPB		23 Jun 2017 19:52:47	U	SMPL
I3808-05DUP - 1	DRUM-001-003-004-COMPDUP	-50	-0.0492	-7470A	PPB		23 Jun 2017 19:54:55	U	SMPL
I3808-05MS - 1	DRUM-001-003-004-COMPMS	15009	3.7688	-7470A	PPB		23 Jun 2017 19:57:04	U	SMPL
I3808-05MSD - 1	DRUM-001-003-004-COMPMSD	14730	3.6980	-7470A	PPB		23 Jun 2017 19:59:11	U	SMPL
CCV40 - 1	CCV40	19676	4.9520	-7470A	PPB		23 Jun 2017 20:01:17	U	SMPL
CCB40 - 1	CCB40	277	0.0337	-7470A	PPB		23 Jun 2017 20:08:21	U	SMPL
I3808-09 - 1	DRUM-002-COMP	116	-0.0072	-7470A	PPB		23 Jun 2017 20:10:33	U	SMPL
I3840-09 - 1	WC-B48-B49-002	-120	-0.0670	-7470A	PPB		23 Jun 2017 20:12:36	U	SMPL
I3840-10 - 1	WC-B53-B54-002	-108	-0.0640	-7470A	PPB		23 Jun 2017 20:14:40	U	SMPL
I3840-11 - 1	WC-B45-B44-002	-185	-0.0835	-7470A	PPB		23 Jun 2017 20:16:46	U	SMPL
I3840-12 - 1	WC-B56-B57-002	-26	-0.0432	-7470A	PPB		23 Jun 2017 20:18:53	U	SMPL
PB100050BL - 1	PBW	-134	-0.0705	-7470A	PPB		23 Jun 2017 20:21:00	U	SMPL
PB100050BS - 1	LCSW	15004	3.7675	-7470A	PPB		23 Jun 2017 20:23:07	U	SMPL
I3808-04 - 1	DRUM-001-003-004-COMP	476	0.0841	-7470A	PPB		23 Jun 2017 20:25:15	U	SMPL
I3841-01 - 1	VL-WWTP-003	398	0.0643	-7470A	PPB		23 Jun 2017 20:27:24	U	SMPL
I3841-01DUP - 1	VL-WWTP-003DUP	450	0.0775	-7470A	PPB		23 Jun 2017 20:29:31	U	SMPL
CCV41 - 1	CCV41	19584	4.9287	-7470A	PPB		23 Jun 2017 20:31:38	U	SMPL
CCB41 - 1	CCB41	-115	-0.0657	-7470A	PPB		23 Jun 2017 20:33:44	U	SMPL
I3841-01MS - 1	VL-WWTP-003MS	15615	3.9224	-7470A	PPB		23 Jun 2017 20:35:52	U	SMPL
I3841-01MSD - 1	VL-WWTP-003MSD	15444	3.8791	-7470A	PPB		23 Jun 2017 20:37:59	U	SMPL
I3867-01 - 1	A6229	127	-0.0044	-7470A	PPB		23 Jun 2017 20:40:06	U	SMPL
I3808-05LX5 - 1		-36	-0.0457	-7470A	PPB		23 Jun 2017 20:42:13	U	SMPL
I3808-05A - 1		14949	3.7536	-7470A	PPB		23 Jun 2017 20:44:20	U	SMPL
I3841-01LX5 - 1		29	-0.0292	-7470A	PPB		23 Jun 2017 20:46:28	U	SMPL
I3841-01A - 1		14866	3.7325	-7470A	PPB		23 Jun 2017 20:48:36	U	SMPL
CCV42 - 1	CCV42	19251	4.8443	-7470A	PPB		23 Jun 2017 20:50:39	U	SMPL
CCB42 - 1	CCB42	-98	-0.0614	-7470A	PPB		23 Jun 2017 20:52:47	U	SMPL