


ISM02.4

LB 88325
INSTRUMENT ID: C41

Linear

 μ Abs.:

45794



A= 0.0000e+000

B= 2.1942e-004

C= -3.7387e-002

Rho= 0.9998481

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.029	-0.029	38	0.000	38					-
0.2	0.200	0.186	-0.014	1020	0.0 %	1020					-7
2.5	2.500	2.516	0.016	11635	0.0 %	11635					1
5.0	5.000	5.116	0.116	23488	0.0 %	23488					2
7.5	7.500	7.400	-0.100	33898	0.0 %	33898					-1
10.0	10.000	10.011	0.011	45794	0.0 %	45794					0

AB
6/23/17

INSTRUMENT IS : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	38	-		0 ISM02.4	PPB	6/23/2017 21:06	Std
	0.2 S0.2	1020	-		0.2 ISM02.4	PPB	6/23/2017 21:08	Std
	2.5 S2.5	11635	-		2.5 ISM02.4	PPB	6/23/2017 21:10	Std
	5 S5	23488	-		5 ISM02.4	PPB	6/23/2017 21:13	Std
	7.5 S7.5	33898	-		7.5 ISM02.4	PPB	6/23/2017 21:15	Std
	10 S10	45794	-		10 ISM02.4	PPB	6/23/2017 21:17	Std
ICV	ICV	18587	4.0409 -		ISM02.4	PPB	6/23/2017 21:20	SMPL
ICB	ICB	-43	-0.0468 -		ISM02.4	PPB	6/23/2017 21:22	SMPL
CCV043	CCV043	23289	5.0726 -		ISM02.4	PPB	6/23/2017 21:24	SMPL
CCB043	CCB043	-54	-0.0492 -		ISM02.4	PPB	6/23/2017 21:27	SMPL
PB100062BL	PBS004	156	-0.0032 -		ISM02.4	PPB	6/23/2017 21:29	SMPL
I3859-01	MH2TR7	473	0.0664 -		ISM02.4	PPB	6/23/2017 21:36	SMPL
I3859-02	MH2TR7D	469	0.0655 -		ISM02.4	PPB	6/23/2017 21:38	SMPL
I3859-03	MH2TR7S	12981	2.8108 -		ISM02.4	PPB	6/23/2017 21:40	SMPL
I3859-04	MH2TR8	453	0.062 -		ISM02.4	PPB	6/23/2017 21:43	SMPL
CCV044	CCV044	23153	5.0427 -		ISM02.4	PPB	6/23/2017 21:45	SMPL
CCB044	CCB044	-48	-0.0479 -		ISM02.4	PPB	6/23/2017 21:47	SMPL