



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88329

Temp.: 23.5 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	06/24/2017	12:15
2	CAL2	1	Water			7.01	06/24/2017	12:16
3	CAL3	1	Water			10.03	06/24/2017	12:17
4	ICV	1	Water			7.00	06/24/2017	12:19
5	CCV1	1	Water			2.02	06/24/2017	12:20
6	I3882-15	1	Solid	20.02	20	9.13	06/24/2017	12:21
7	I3882-15DUP	1	Solid	20.03	20	9.14	06/24/2017	12:23
8	I3882-16	1	Solid	20.02	20	8.90	06/24/2017	12:24
9	I3882-17	1	Solid	20.04	20	10.05	06/24/2017	12:25
10	I3882-18	1	Solid	20.01	20	9.31	06/24/2017	12:26
11	I3882-19	1	Solid	20.03	20	10.70	06/24/2017	12:29
12	I3882-20	1	Solid	20.02	20	9.38	06/24/2017	12:30
13	I3882-21	1	Solid	20.01	20	9.90	06/24/2017	12:33
14	I3882-22	1	Solid	20.05	20	8.61	06/24/2017	12:36
15	I3882-23	1	Solid	20.02	20	9.80	06/24/2017	12:37
16	CCV2	1	Water			12.01	06/24/2017	12:39
17	I3882-24	1	Solid	20.01	20	9.30	06/24/2017	12:40
18	I3882-25	1	Solid	20.03	20	10.67	06/24/2017	12:41
19	I3882-26	1	Solid	20.02	20	10.36	06/24/2017	12:42
20	I3882-27	1	Solid	20.01	20	10.76	06/24/2017	12:43

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I3882-28	1	Solid	20.03	20	10.65	06/24/2017	12:45
22	CCV3	1	Water			12.02	06/24/2017	12:46

LB 88329

WORKLIST(Hardcopy Internal Chain)

Date : 6/24/2017 8:20:27 AM

WorkList ID : 100045

WorkList Name : CORROSIONITY I3882

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3882-15	Corrosivity	Cool 4 deg C	TECT01	N53	G51-1.5-3	06/22/2017	9045C
	Solid	I3882-16	Corrosivity	Cool 4 deg C	TECT01	N53	G51-0-1.5	06/22/2017	9045C
	Solid	I3882-17	Corrosivity	Cool 4 deg C	TECT01	N53	G46-0-1.5	06/22/2017	9045C
	Solid	I3882-18	Corrosivity	Cool 4 deg C	TECT01	N53	G46-1.5-3	06/22/2017	9045C
	Solid	I3882-19	Corrosivity	Cool 4 deg C	TECT01	N53	G52-0-1.5	06/22/2017	9045C
	Solid	I3882-20	Corrosivity	Cool 4 deg C	TECT01	N53	G52-1.5-3	06/22/2017	9045C
	Solid	I3882-21	Corrosivity	Cool 4 deg C	TECT01	N53	G43-0-1.5	06/22/2017	9045C
	Solid	I3882-22	Corrosivity	Cool 4 deg C	TECT01	N53	G43-1.5-3	06/22/2017	9045C
	Solid	I3882-23	Corrosivity	Cool 4 deg C	TECT01	N53	G47-0-1.5	06/22/2017	9045C
	Solid	I3882-24	Corrosivity	Cool 4 deg C	TECT01	N53	G47-1.5-3	06/22/2017	9045C
	Solid	I3882-25	Corrosivity	Cool 4 deg C	TECT01	N53	G56-0-1.5	06/22/2017	9045C
	Solid	I3882-26	Corrosivity	Cool 4 deg C	TECT01	N53	G56-1.5-3	06/22/2017	9045C
	Solid	I3882-27	Corrosivity	Cool 4 deg C	TECT01	N53	G54-0-1.5	06/22/2017	9045C
	Solid	I3882-28	Corrosivity	Cool 4 deg C	TECT01	N53	G54-1.5-3	06/22/2017	9045C

Date/Time 06/24/17 8:20:27 AM
 Received by: JP
 Relinquished by: JP

Date/Time 06/24/17 8:20:27 AM
 Received by: JP
 Relinquished by: JP