



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88336

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/23/2017	17:30
2	CAL2	1	Water			7.00	06/23/2017	17:31
3	CAL3	1	Water			10.03	06/23/2017	17:33
4	ICV	1	Water			7.01	06/23/2017	17:35
5	CCV1	1	Water			2.01	06/23/2017	17:36
6	I3874-04	1	Solid	20.03	20	8.39	06/23/2017	17:39
7	I3874-05	1	Solid	20.01	20	8.33	06/23/2017	17:40
8	I3874-06	1	Solid	20.04	20	7.23	06/23/2017	17:41
9	I3897-03	1	Solid	20.00	20	6.77	06/23/2017	17:42
10	I3898-09	1	Solid	20.04	20	8.00	06/23/2017	17:44
11	I3899-07	1	Solid	20.01	20	9.20	06/23/2017	17:45
12	I3906-03	1	Solid	20.02	20	8.32	06/23/2017	17:46
13	I3906-06	1	Solid	20.04	20	9.20	06/23/2017	17:47
14	I3906-09	1	Solid	20.02	20	8.16	06/23/2017	17:49
15	I3911-03	1	Solid	20.04	20	7.77	06/23/2017	17:50
16	CCV2	1	Water			12.01	06/23/2017	17:51
17	I3912-03	1	Solid	20.03	20	8.74	06/23/2017	17:52
18	I3913-03	1	Solid	20.05	20	8.60	06/23/2017	18:00
19	I3913-03DUP	1	Solid	20.06	20	8.61	06/23/2017	18:01
20	CCV3	1	Water			12.02	06/23/2017	18:03

LB88336

WORKLIST(Hardcopy Internal Chain)

WorkList Name: CORROSIONITY I3874

WorkList ID: 100000

Date: 6/23/2017 12:11:09 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3874-04	Corrosivity	Cool 4 deg C	GUAR01	N41	TL-WTP-01	06/21/2017	9045C
	Solid	I3874-05	Corrosivity	Cool 4 deg C	GUAR01	N41	TL-WTP-02	06/22/2017	9045C
	Solid	I3874-06	Corrosivity	Cool 4 deg C	GUAR01	N41	TL-DR-01	06/22/2017	9045C
06/28/2017	Solid	I3897-03	Corrosivity	Cool 4 deg C	PSEG01	O52	72-11931	06/23/2017	9045C
06/28/2017	Solid	I3898-09	Corrosivity	Cool 4 deg C	PSEG01	L51	VNJ-72-12002-VNJ-201-COM	06/23/2017	9045C
06/28/2017	Solid	I3899-07	Corrosivity	Cool 4 deg C	PSEG01	O53	72-11940	06/23/2017	9045C
06/28/2017	Solid	I3906-03	Corrosivity	Cool 4 deg C	PSEG05	O52	EO-02-062317-A	06/23/2017	9045C
06/28/2017	Solid	I3906-06	Corrosivity	Cool 4 deg C	PSEG05	O52	EO-02-062317-B	06/23/2017	9045C
06/28/2017	Solid	I3906-09	Corrosivity	Cool 4 deg C	PSEG05	O52	EO-02-062317-C	06/23/2017	9045C
06/28/2017	Solid	I3911-03	Corrosivity	Cool 4 deg C	PSEG05	M51	SU-04-062317	06/23/2017	9045C
06/28/2017	Solid	I3912-03	Corrosivity	Cool 4 deg C	PSEG05	N33	OK-03-062317	06/23/2017	9045C
06/28/2017	Solid	I3913-03	Corrosivity	Cool 4 deg C	PSEG05	M31	TR-04-062317	06/23/2017	9045C

Date/Time 06/23/17 12:11:09
 Received by: JB
 Relinquished by: CB

Date/Time 06/23/17 7:31M
 Received by: LB
 Relinquished by: JP