



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88337

Temp.: 20.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/23/2017	17:20
2	CAL2	1	Water			7.00	06/23/2017	17:21
3	CAL3	1	Water			10.03	06/23/2017	17:22
4	ICV	1	Water			7.00	06/23/2017	17:24
5	CCV1	1	Water			2.01	06/23/2017	17:25
6	I3898-02	1	Water			7.46	06/23/2017	17:26
7	I3898-02DUP	1	Water			7.47	06/23/2017	17:28
8	CCV2	1	Water			12.01	06/23/2017	17:30

2B 88337

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3898 WorkList ID : 100053 Date : 6/23/2017 4:49:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/23/2017	Water	I3898-02-A	Corrosivity	Cool 4 deg C	PSEG01	L51	A6987MDW	06/23/2017	9040C

Date/Time 06/23/17 11:14 AM
Received by: JP
Relinquished by: S

Date/Time 06/23/17 7:20 PM
Received by: CR
Relinquished by: Jo