



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88340

Temp.: 22.1 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/23/2017	16:50
2	CAL2	1	Water			6.99	06/23/2017	16:51
3	CAL3	1	Water			10.02	06/23/2017	16:52
4	ICV	1	Water			7.00	06/23/2017	16:53
5	CCV1	1	Water			2.01	06/23/2017	17:00
6	I3885-02	1	Water			7.16	06/23/2017	17:01
7	I3885-02DUP	1	Water			7.17	06/23/2017	17:02
8	I3900-01	1	Water			6.06	06/23/2017	17:03
9	I3900-02	1	Water			6.48	06/23/2017	17:05
10	I3900-03	1	Water			6.51	06/23/2017	17:06
11	I3900-04	1	Water			6.63	06/23/2017	17:07
12	CCV2	1	Water			12.01	06/23/2017	17:09

WORKLIST(Hardcopy Internal Chain)

LB88340

WorkList Name : PH 13900

WorkList ID : 100019

Date : 6/23/2017 4:44:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I3900-01-H	pH	Cool 4 deg C	LOCK01	N61	MW-1	06/22/2017	SM 4500-PH B
	Water	I3900-02-G	pH	Cool 4 deg C	LOCK01	N61	MW-3	06/22/2017	SM 4500-PH B
	Water	I3900-03-H	pH	Cool 4 deg C	LOCK01	N61	MW-4	06/22/2017	SM 4500-PH B
	Water	I3900-04-L	pH	Cool 4 deg C	LOCK01	N61	MW-2	06/22/2017	SM 4500-PH B
	Water	I3885-02-L	pH	Cool 4 deg C	PARS15	N53	CAN-WC-L-1	06/22/2017	SM 4500-PH B

Date/Time 06/23/17 4:44 PM

Received by: JB

Relinquished by: CG

Date/Time 06/23/17 7:00 PM

Received by: JC

Relinquished by: JB