

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB88353BALANCE ID:

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP55256
calibration std. phosphate 1 ppm	WP55934
calibration std. phosphate 0.5 ppm	WP55933
calibration std. phosphate 0.3 ppm	WP55932
calibration std. 0 ppm	WP55929
phosphate ICV std.	WP55936
phosphate CCV std.	WP55935
calibration std. phosphate 0.1 ppm	WP55931
calibration std. phosphate 0.05 ppm	WP55930
Combined reagent	WP55938
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749

Intercept: -0.0014Slope: 0.6527Regression: 0.999837

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	06/26/2017	10:56
2	CAL2	0.05	1	50	50	0.032	06/26/2017	10:56
3	CAL3	0.10	1	50	50	0.065	06/26/2017	10:57
4	CAL4	0.30	1	50	50	0.186	06/26/2017	10:57
5	CAL5	0.50	1	50	50	0.330	06/26/2017	10:58
6	CAL6	1.00	1	50	50	0.651	06/26/2017	10:58



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB88353

BALANCE ID:

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.494	06/26/2017	10:59
2	ICB		1	50	50	0.000	0.002	06/26/2017	10:59
3	CCV1	0.50	1	50	50	0.329	0.506	06/26/2017	11:00
4	CCB1		1	50	50	0.000	0.002	06/26/2017	11:00
5	PB100166BL		1	1.00	50	0.000	0.002	06/26/2017	11:01
6	PB100166BS	0.50	1	1.02	50	0.320	0.492	06/26/2017	11:01
7	I3879-01		1	1.04	50	2.658	4.074	06/26/2017	11:02
8	I3879-01DUP		1	1.01	50	2.661	4.079	06/26/2017	11:02
9	I3879-01MS	0.50	1	1.06	50	2.673	4.097	06/26/2017	11:03
10	I3879-01MSD	0.50	1	1.06	50	2.670	4.093	06/26/2017	11:03
11	I3879-01		200	1.04	50	0.394	0.606	06/26/2017	11:20
12	I3879-01DUP		200	1.01	50	0.390	0.600	06/26/2017	11:21
13	I3879-01MS	0.50	200	1.06	50	0.571	0.877	06/26/2017	11:21
14	I3879-01MSD	0.50	200	1.06	50	0.568	0.872	06/26/2017	11:22
15	CCV2	0.50	1	50	50	0.331	0.509	06/26/2017	11:22
16	CCB2		1	50	50	0.000	0.002	06/26/2017	11:23