



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88354

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/26/2017	17:10
2	CAL2	1	Water			7.00	06/26/2017	17:11
3	CAL3	1	Water			10.03	06/26/2017	17:12
4	ICV	1	Water			6.99	06/26/2017	17:13
5	CCV1	1	Water			2.02	06/26/2017	17:15
6	I3437-08	1	Solid	20.03	20	11.51	06/26/2017	17:17
7	I3437-08DUP	1	Solid	20.05	20	11.52	06/26/2017	17:19
8	CCV2	1	Water			12.01	06/26/2017	17:20

1988354

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i3457

WorkList ID : 100099

Date : 6/26/2017 5:05:06 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2017	Solid	I3437-08	Corrosivity	Cool 4 deg C	TECT01	J51	COMP-1	06/01/2017	9045C

Date/Time 6/26/17 5:05 PM
Received by: 31
Relinquished by: [Signature]

Date/Time 6/26/17 9:00 AM
Received by: [Signature]
Relinquished by: [Signature]