



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88357

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/26/2017	16:25
2	CAL2	1	Water			7.01	06/26/2017	16:26
3	CAL3	1	Water			10.03	06/26/2017	16:27
4	ICV	1	Water			7.00	06/26/2017	16:30
5	CCV1	1	Water			2.01	06/26/2017	16:31
6	I3908-13	1	Solid	20.03	20	9.07	06/26/2017	16:32
7	I3908-13DUP	1	Solid	20.04	20	9.08	06/26/2017	16:33
8	I3908-14	1	Solid	20.05	20	8.42	06/26/2017	16:35
9	I3908-15	1	Solid	20.04	20	10.41	06/26/2017	16:36
10	I3908-16	1	Solid	20.00	20	7.10	06/26/2017	16:37
11	I3908-17	1	Solid	20.03	20	8.58	06/26/2017	16:39
12	I3908-18	1	Solid	20.05	20	7.73	06/26/2017	16:40
13	I3908-19	1	Solid	20.03	20	10.23	06/26/2017	16:41
14	I3908-20	1	Solid	20.02	20	8.93	06/26/2017	16:42
15	I3908-21	1	Solid	20.02	20	8.60	06/26/2017	16:43
16	CCV2	1	Water			12.01	06/26/2017	16:45
17	I3908-22	1	Solid	20.02	20	7.75	06/26/2017	16:46
18	I3908-23	1	Solid	20.03	20	8.14	06/26/2017	16:47
19	I3908-24	1	Solid	20.01	20	8.89	06/26/2017	16:49
20	I3925-03	1	Solid	20.03	20	8.74	06/26/2017	16:50

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I3926-05	1	Solid	20.05	20	8.63	06/26/2017	16:51
22	I3926-10	1	Solid	20.01	20	7.92	06/26/2017	16:52
23	I3931-03	1	Solid	20.04	20	7.70	06/26/2017	16:53
24	I3932-05	1	Solid	20.02	20	5.53	06/26/2017	16:54
25	I3932-05DUP	1	Solid	20.01	20	5.54	06/26/2017	17:00
26	CCV3	1	Water			12.01	06/26/2017	17:01

1B8357

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3925

WorkList ID : 100092

Date : 6/26/2017 2:36:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3908-13	Corrosivity	Cool 4 deg C	TECT01	N62	G-55-0-1.5	06/23/2017	9045C
	Solid	I3908-14	Corrosivity	Cool 4 deg C	TECT01	N62	G-55-1.5-3	06/23/2017	9045C
	Solid	I3908-15	Corrosivity	Cool 4 deg C	TECT01	N62	G-57-0-1.5	06/23/2017	9045C
	Solid	I3908-16	Corrosivity	Cool 4 deg C	TECT01	N62	G-57-1.5-3	06/23/2017	9045C
	Solid	I3908-17	Corrosivity	Cool 4 deg C	TECT01	N62	G-60-0-1.5	06/23/2017	9045C
	Solid	I3908-18	Corrosivity	Cool 4 deg C	TECT01	N62	G-60-1.5-3	06/23/2017	9045C
	Solid	I3908-19	Corrosivity	Cool 4 deg C	TECT01	N62	G-59-0-1.5	06/23/2017	9045C
	Solid	I3908-20	Corrosivity	Cool 4 deg C	TECT01	N62	G-59-1.5-3	06/23/2017	9045C
	Solid	I3908-21	Corrosivity	Cool 4 deg C	TECT01	N62	G-58-0-1.5	06/23/2017	9045C
	Solid	I3908-22	Corrosivity	Cool 4 deg C	TECT01	N62	G-58-1.5-3	06/23/2017	9045C
	Solid	I3908-23	Corrosivity	Cool 4 deg C	TECT01	N62	G-61-0-1.5	06/23/2017	9045C
	Solid	I3908-24	Corrosivity	Cool 4 deg C	TECT01	N62	G-61-1.5-3	06/23/2017	9045C
06/29/2017	Solid	I3925-03	Corrosivity	Cool 4 deg C	PSEG05	O63	HD-01-062617	06/26/2017	9045C
06/29/2017	Solid	I3926-05	Corrosivity	Cool 4 deg C	PSEG01	O63	SP1-A-COMP	06/26/2017	9045C
06/29/2017	Solid	I3926-10	Corrosivity	Cool 4 deg C	PSEG01	O63	SP1-B-COMP	06/26/2017	9045C
06/29/2017	Solid	I3931-03	Corrosivity	Cool 4 deg C	PSEG05	P11	OR-01-062617	06/26/2017	9045C
06/29/2017	Solid	I3932-05	Corrosivity	Cool 4 deg C	PSEG01	P13	TP-C1-(0-8)	06/26/2017	9045C

06/26/17 2:36 PM

Date/Time

Received by: JP

JP

Relinquished by:

JP

Date/Time

Received by: JP

JP

Relinquished by:

JP