


7470A

LB 88360
INSTRUMENT ID: CV1 μ Abs.:

41222



Linear

A= 0.0000e+000

B= 2.4254e-004

C= -5.3426e-003

Rho= 0.9999722

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.009	-0.009	-14	0.000	-14				
0.2	0.200	0.198	-0.002	840	0.0 %	840				
2.5	2.500	2.536	0.036	10480	0.0 %	10480				
5.0	5.000	4.952	-0.048	20441	0.0 %	20441				
7.5	7.500	7.529	0.029	31063	0.0 %	31063				
10.0	10.000	9.993	-0.007	41222	0.0 %	41222				

LB88360 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Jun 2017 18:10:23

Sample ID	Extended ID	µ Abs.	Conc.	Std. Conc/Method	Units	Date	Type	Type
0.0 - 1	50	-14	-	0.00007470A	PPB	26 Jun 2017 18:33:19	S	Std
0.2 - 1	50.2	840	-	0.20007470A	PPB	26 Jun 2017 18:35:27	S	Std
2.5 - 1	52.5	10480	-	2.50007470A	PPB	26 Jun 2017 18:37:36	S	Std
5.0 - 1	55	20441	-	5.00007470A	PPB	26 Jun 2017 18:39:44	S	Std
7.5 - 1	57.5	31063	-	7.50007470A	PPB	26 Jun 2017 18:41:52	S	Std
10.0 - 1	510	41222	-	10.00007470A	PPB	26 Jun 2017 18:43:59	S	Std
ICV40 - 1	ICV40	16801	4.0696	-7470A	PPB	26 Jun 2017 18:46:42	U	SMPL
ICB40 - 1	ICB40	338	0.0766	-7470A	PPB	26 Jun 2017 18:48:49	U	SMPL
CCV45 - 1	CCV45	19920	4.8261	-7470A	PPB	26 Jun 2017 18:50:56	U	SMPL
CCB45 - 1	CCB45	-94	-0.0281	-7470A	PPB	26 Jun 2017 18:53:02	U	SMPL
CRA - 1	CRA	841	0.1986	-7470A	PPB	26 Jun 2017 18:55:10	U	SMPL
HighStd - 1	HighStd	41506	10.0616	-7470A	PPB	26 Jun 2017 18:57:16	U	SMPL
ChkStd - 1	ChkStd	31191	7.5598	-7470A	PPB	26 Jun 2017 18:59:23	U	SMPL
PB100151BL - 1	PBW	-168	-0.0461	-7470A	PPB	26 Jun 2017 19:01:32	U	SMPL
PB100151BS - 1	LCSW	14422	3.4926	-7470A	PPB	26 Jun 2017 19:03:41	U	SMPL
PB100151TB - 1	PB100048TB	-129	-0.0366	-7470A	PPB	26 Jun 2017 19:05:49	U	SMPL
I3851-13 - 1	G-26-0-1.5	-57	-0.0192	-7470A	PPB	26 Jun 2017 19:07:58	U	SMPL
I3851-14 - 1	G-26-1.5-3	-59	-0.0197	-7470A	PPB	26 Jun 2017 19:10:06	U	SMPL
I3851-15 - 1	G-30-0-1.5	-54	-0.0184	-7470A	PPB	26 Jun 2017 19:12:15	U	SMPL
I3851-16 - 1	G-30-1.5-3	-44	-0.0160	-7470A	PPB	26 Jun 2017 19:14:21	U	SMPL
CCV46 - 1	CCV46	19946	4.8324	-7470A	PPB	26 Jun 2017 19:16:28	U	SMPL
CCB46 - 1	CCB46	57	0.0085	-7470A	PPB	26 Jun 2017 19:18:35	U	SMPL
I3851-17 - 1	G-36-0-1.5	-41	-0.0153	-7470A	PPB	26 Jun 2017 19:20:43	U	SMPL
I3851-18 - 1	G-36-1.5-3	-42	-0.0155	-7470A	PPB	26 Jun 2017 19:22:50	U	SMPL
I3851-19 - 1	G-34-0-1.5	-52	-0.0180	-7470A	PPB	26 Jun 2017 19:24:56	U	SMPL
I3851-20 - 1	G-34-1.5-3	-54	-0.0184	-7470A	PPB	26 Jun 2017 19:27:03	U	SMPL
I3851-21 - 1	G-37-0-1.5	-51	-0.0177	-7470A	PPB	26 Jun 2017 19:29:10	U	SMPL
I3851-22 - 1	G-37-1.5-3	-45	-0.0163	-7470A	PPB	26 Jun 2017 19:31:18	U	SMPL
I3851-23 - 1	G-35-0-1.5	-33	-0.0133	-7470A	PPB	26 Jun 2017 19:33:26	U	SMPL
I3851-24 - 1	G-35-1.5-3	-35	-0.0138	-7470A	PPB	26 Jun 2017 19:35:34	U	SMPL
I3867-02 - 1	A6229	-47	-0.0167	-7470A	PPB	26 Jun 2017 19:37:43	U	SMPL
I3868-03 - 1	CL-01-062217-A	-46	-0.0165	-7470A	PPB	26 Jun 2017 19:39:50	U	SMPL
CCV47 - 1	CCV47	19982	4.8411	-7470A	PPB	26 Jun 2017 19:41:57	U	SMPL
CCB47 - 1	CCB47	334	0.0757	-7470A	PPB	26 Jun 2017 19:44:04	U	SMPL
I3868-06 - 1	CL-01-062217-B	-73	-0.0230	-7470A	PPB	26 Jun 2017 19:46:11	U	SMPL
I3868-09 - 1	CL-01-062217-C	-28	-0.0121	-7470A	PPB	26 Jun 2017 19:48:18	U	SMPL
I3868-12 - 1	CL-01-062217-D	-34	-0.0136	-7470A	PPB	26 Jun 2017 19:50:25	U	SMPL
I3868-15 - 1	CL-01-062217-E	-27	-0.0119	-7470A	PPB	26 Jun 2017 19:52:32	U	SMPL
I3872-03 - 1	HR-05-062217-A	-27	-0.0119	-7470A	PPB	26 Jun 2017 19:54:39	U	SMPL
I3872-06 - 1	HR-05-062217-B	-35	-0.0138	-7470A	PPB	26 Jun 2017 19:56:47	U	SMPL
I3872-09 - 1	HR-05-062217-C	-36	-0.0141	-7470A	PPB	26 Jun 2017 19:58:55	U	SMPL
I3872-09DUP - 1	HR-05-062217-CDUP	-40	-0.0150	-7470A	PPB	26 Jun 2017 20:01:03	U	SMPL
I3872-09MS - 1	HR-05-062217-CMS	14162	3.4295	-7470A	PPB	26 Jun 2017 20:03:11	U	SMPL
I3872-09MSD - 1	HR-05-062217-CMSD	13624	3.2990	-7470A	PPB	26 Jun 2017 20:05:18	U	SMPL
CCV48 - 1	CCV48	19475	4.7181	-7470A	PPB	26 Jun 2017 20:09:58	U	SMPL
CCB48 - 1	CCB48	39	0.0041	-7470A	PPB	26 Jun 2017 20:12:05	U	SMPL
PB100152BL - 1	PBW	-23	-0.0109	-7470A	PPB	26 Jun 2017 20:14:13	U	SMPL
PB100152BS - 1	LCSW	14599	3.5355	-7470A	PPB	26 Jun 2017 20:16:20	U	SMPL
PB100152TB - 1	PB100081TB	-84	-0.0257	-7470A	PPB	26 Jun 2017 20:18:27	U	SMPL
I3852-19 - 1	G-33-0-1.5	-41	-0.0153	-7470A	PPB	26 Jun 2017 20:20:35	U	SMPL
I3852-20 - 1	G-33-1.5-3	-29	-0.0124	-7470A	PPB	26 Jun 2017 20:22:43	U	SMPL
I3852-21 - 1	G-32-0-1.5	-38	-0.0146	-7470A	PPB	26 Jun 2017 20:24:51	U	SMPL
I3852-22 - 1	G-32-1.5-3	-37	-0.0143	-7470A	PPB	26 Jun 2017 20:26:58	U	SMPL
I3852-23 - 1	G-31-0-1.5	-26	-0.0116	-7470A	PPB	26 Jun 2017 20:29:06	U	SMPL
I3852-24 - 1	G-31-1.5-3	-20	-0.0102	-7470A	PPB	26 Jun 2017 20:31:15	U	SMPL
I3852-25 - 1	G-28-0-1.5	-29	-0.0124	-7470A	PPB	26 Jun 2017 20:33:22	U	SMPL
CCV49 - 1	CCV49	19889	4.8186	-7470A	PPB	26 Jun 2017 20:35:29	U	SMPL
CCB49 - 1	CCB49	47	0.0061	-7470A	PPB	26 Jun 2017 20:37:36	U	SMPL
I3852-26 - 1	G-28-1.5-3	-25	-0.0114	-7470A	PPB	26 Jun 2017 20:39:44	U	SMPL
I3852-27 - 1	G-27-0-1.5	-18	-0.0097	-7470A	PPB	26 Jun 2017 20:41:51	U	SMPL
I3852-28 - 1	G-27-1.5-3	-29	-0.0124	-7470A	PPB	26 Jun 2017 20:43:58	U	SMPL
I3852-29 - 1	G-39-0-1.5	-25	-0.0114	-7470A	PPB	26 Jun 2017 20:46:05	U	SMPL
I3852-30 - 1	G-39-1.5-3	-34	-0.0136	-7470A	PPB	26 Jun 2017 20:48:12	U	SMPL
I3852-31 - 1	G-48-0-1.5	-28	-0.0121	-7470A	PPB	26 Jun 2017 20:50:19	U	SMPL
I3852-32 - 1	G-48-1.5-3	-24	-0.0112	-7470A	PPB	26 Jun 2017 20:52:28	U	SMPL
I3852-33 - 1	G-38-0-1.5	-30	-0.0126	-7470A	PPB	26 Jun 2017 20:54:36	U	SMPL
I3852-34 - 1	G-38-1.5-3	-32	-0.0131	-7470A	PPB	26 Jun 2017 20:56:45	U	SMPL
I3852-35 - 1	G-50-0-1.5	-31	-0.0129	-7470A	PPB	26 Jun 2017 20:58:52	U	SMPL
CCV50 - 1	CCV50	20140	4.8794	-7470A	PPB	26 Jun 2017 21:01:00	U	SMPL
CCB50 - 1	CCB50	63	0.0099	-7470A	PPB	26 Jun 2017 21:03:07	U	SMPL
I3852-36 - 1	G-50-1.5-3	-32	-0.0131	-7470A	PPB	26 Jun 2017 21:05:15	U	SMPL
I3852-36DUP - 1	G-50-1.5-3DUP	-37	-0.0143	-7470A	PPB	26 Jun 2017 21:07:23	U	SMPL
I3852-36MS - 1	G-50-1.5-3MS	14450	3.4994	-7470A	PPB	26 Jun 2017 21:09:30	U	SMPL
I3852-36MSD - 1	G-50-1.5-3MSD	14109	3.4167	-7470A	PPB	26 Jun 2017 21:11:37	U	SMPL
PB100174BL - 1	PBW	-95	-0.0284	-7470A	PPB	26 Jun 2017 21:13:44	U	SMPL
PB100174BS - 1	LCSW	14922	3.6139	-7470A	PPB	26 Jun 2017 21:15:52	U	SMPL
PB100174TB - 1	PB100097TB	-46	-0.0165	-7470A	PPB	26 Jun 2017 21:18:00	U	SMPL
I3874-04 - 1	TL-WTP-01	-7	-0.0070	-7470A	PPB	26 Jun 2017 21:20:08	U	SMPL
I3874-04DUP - 1	TL-WTP-01DUP	0	-0.0053	-7470A	PPB	26 Jun 2017 21:22:17	U	SMPL
I3874-04MS - 1	TL-WTP-01MS	14961	3.6233	-7470A	PPB	26 Jun 2017 21:24:24	U	SMPL
CCV51 - 1	CCV51	20339	4.9277	-7470A	PPB	26 Jun 2017 21:26:32	U	SMPL
CCB51 - 1	CCB51	46	0.0058	-7470A	PPB	26 Jun 2017 21:28:39	U	SMPL
I3874-04MSD - 1	TL-WTP-01MSD	15047	3.6442	-7470A	PPB	26 Jun 2017 21:30:48	U	SMPL
I3874-05 - 1	TL-WTP-02	-74	-0.0233	-7470A	PPB	26 Jun 2017 21:32:56	U	SMPL
I3874-06 - 1	TL-DR-01	7	-0.0036	-7470A	PPB	26 Jun 2017 21:35:03	U	SMPL
I3897-03 - 1	72-11931	8	-0.0034	-7470A	PPB	26 Jun 2017 21:37:11	U	SMPL
I3898-02 - 1	A6987MDW	4	-0.0044	-7470A	PPB	26 Jun 2017 21:39:18	U	SMPL
I3898-09 - 1	VNJ-72-12002-VNJ-201-COMP	-3	-0.0061	-7470A	PPB	26 Jun 2017 21:41:26	U	SMPL
I3899-07 - 1	72-11940	-10	-0.0078	-7470A	PPB	26 Jun 2017 21:43:35	U	SMPL
I3906-03 - 1	EO-02-062317-A	-2	-0.0058	-7470A	PPB	26 Jun 2017 21:45:43	U	SMPL
I3906-06 - 1	EO-02-062317-B	-6	-0.0068	-7470A	PPB	26 Jun 2017 21:47:51	U	SMPL
I3906-09 - 1	EO-02-062317-C	-11	-0.0080	-7470A	PPB	26 Jun 2017 21:49:58	U	SMPL
CCV52 - 1	CCV52	20038	4.8547	-7470A	PPB	26 Jun 2017 21:52:06	U	SMPL

LB88360 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Jun 2017 18:10:23

Sample ID	Extended ID	µ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
CCB52 - 1	CCB52	55	0.0080	-7470A	PPB		26 Jun 2017 21:54:14	U	SMPL
I3911-03 - 1	SU-04-062317	-18	-0.0097	-7470A	PPB		26 Jun 2017 21:56:23	U	SMPL
I3912-03 - 1	OK-03-062317	-14	-0.0087	-7470A	PPB		26 Jun 2017 21:58:30	U	SMPL
I3913-03 - 1	TR-04-062317	13	-0.0022	-7470A	PPB		26 Jun 2017 22:00:38	U	SMPL
PB100175BL - 1	PBW	-11	-0.0080	-7470A	PPB		26 Jun 2017 22:02:46	U	SMPL
PB100175BS - 1	LCSW	15353	3.7184	-7470A	PPB		26 Jun 2017 22:04:54	U	SMPL
PB100175TB - 1	PB100133TB	-64	-0.0209	-7470A	PPB		26 Jun 2017 22:07:01	U	SMPL
I3882-15 - 1	G51-1.5-3	-15	-0.0090	-7470A	PPB		26 Jun 2017 22:09:10	U	SMPL
I3882-15DUP - 1	G51-1.5-3DUP	-10	-0.0078	-7470A	PPB		26 Jun 2017 22:11:18	U	SMPL
I3882-15MS - 1	G51-1.5-3MS	15503	3.7548	-7470A	PPB		26 Jun 2017 22:13:26	U	SMPL
I3882-15MSD - 1	G51-1.5-3MSD	14907	3.6102	-7470A	PPB		26 Jun 2017 22:15:34	U	SMPL
CCV53 - 1	CCV53	20054	4.8586	-7470A	PPB		26 Jun 2017 22:17:42	U	SMPL
CCB53 - 1	CCB53	12	-0.0024	-7470A	PPB		26 Jun 2017 22:19:50	U	SMPL
I3882-16 - 1	G51-0-1.5	-22	-0.0107	-7470A	PPB		26 Jun 2017 22:21:58	U	SMPL
I3882-17 - 1	G46-0-1.5	-17	-0.0095	-7470A	PPB		26 Jun 2017 22:24:06	U	SMPL
I3882-18 - 1	G46-1.5-3	-12	-0.0083	-7470A	PPB		26 Jun 2017 22:26:14	U	SMPL
I3882-19 - 1	G52-0-1.5	-10	-0.0078	-7470A	PPB		26 Jun 2017 22:28:22	U	SMPL
I3882-20 - 1	G52-1.5-3	-14	-0.0087	-7470A	PPB		26 Jun 2017 22:30:30	U	SMPL
I3882-21 - 1	G43-0-1.5	-10	-0.0078	-7470A	PPB		26 Jun 2017 22:32:37	U	SMPL
I3882-22 - 1	G43-1.5-3	-10	-0.0078	-7470A	PPB		26 Jun 2017 22:34:45	U	SMPL
I3882-23 - 1	G47-0-1.5	-4	-0.0063	-7470A	PPB		26 Jun 2017 22:36:53	U	SMPL
I3882-24 - 1	G47-1.5-3	-7	-0.0070	-7470A	PPB		26 Jun 2017 22:39:02	U	SMPL
I3882-25 - 1	G56-0-1.5	-27	-0.0119	-7470A	PPB		26 Jun 2017 22:41:10	U	SMPL
CCV54 - 1	CCV54	20560	4.9813	-7470A	PPB		26 Jun 2017 22:43:18	U	SMPL
CCB54 - 1	CCB54	35	0.0031	-7470A	PPB		26 Jun 2017 22:45:26	U	SMPL
I3882-26 - 1	G56-1.5-3	-19	-0.0100	-7470A	PPB		26 Jun 2017 22:47:35	U	SMPL
I3882-27 - 1	G54-0-1.5	-9	-0.0075	-7470A	PPB		26 Jun 2017 22:49:43	U	SMPL
I3882-28 - 1	G54-1.5-3	-35	-0.0138	-7470A	PPB		26 Jun 2017 22:51:51	U	SMPL
PB100176BL - 1	PBW	-23	-0.0109	-7470A	PPB		26 Jun 2017 22:53:58	U	SMPL
PB100176BS - 1	LCSW	14667	3.5520	-7470A	PPB		26 Jun 2017 22:56:06	U	SMPL
PB100176TB - 1	PB100137TB	-81	-0.0250	-7470A	PPB		26 Jun 2017 22:58:14	U	SMPL
I3883-17 - 1	G45-0-1.5	-36	-0.0141	-7470A	PPB		26 Jun 2017 23:00:22	U	SMPL
I3883-17DUP - 1	G45-0-1.5DUP	-15	-0.0090	-7470A	PPB		26 Jun 2017 23:02:31	U	SMPL
I3883-17MS - 1	G45-0-1.5MS	13663	3.3085	-7470A	PPB		26 Jun 2017 23:04:39	U	SMPL
I3883-17MSD - 1	G45-0-1.5MSD	15587	3.7751	-7470A	PPB		26 Jun 2017 23:06:47	U	SMPL
CCV55 - 1	CCV55	19935	4.8297	-7470A	PPB		26 Jun 2017 23:08:55	U	SMPL
CCB55 - 1	CCB55	23	0.0002	-7470A	PPB		26 Jun 2017 23:11:04	U	SMPL
I3883-18 - 1	G45-1.5-3	-19	-0.0100	-7470A	PPB		26 Jun 2017 23:13:13	U	SMPL
I3883-19 - 1	G41-0-1.5	-37	-0.0143	-7470A	PPB		26 Jun 2017 23:15:21	U	SMPL
I3883-20 - 1	G41-1.5-3	-26	-0.0116	-7470A	PPB		26 Jun 2017 23:17:29	U	SMPL
I3883-21 - 1	G49-0-1.5	-27	-0.0119	-7470A	PPB		26 Jun 2017 23:19:37	U	SMPL
I3883-22 - 1	G49-1.5-3	7	-0.0036	-7470A	PPB		26 Jun 2017 23:21:46	U	SMPL
I3883-23 - 1	G42-0-1.5	-5	-0.0066	-7470A	PPB		26 Jun 2017 23:23:54	U	SMPL
I3883-24 - 1	G42-1.5-3	-31	-0.0129	-7470A	PPB		26 Jun 2017 23:26:02	U	SMPL
I3883-25 - 1	G44-0-1.5	-34	-0.0136	-7470A	PPB		26 Jun 2017 23:28:11	U	SMPL
I3883-26 - 1	G44-1.5-3	-23	-0.0109	-7470A	PPB		26 Jun 2017 23:30:19	U	SMPL
I3883-27 - 1	G53-0-1.5	-19	-0.0100	-7470A	PPB		26 Jun 2017 23:32:27	U	SMPL
CCV56 - 1	CCV56	19903	4.8220	-7470A	PPB		26 Jun 2017 23:34:36	U	SMPL
CCB56 - 1	CCB56	26	0.0010	-7470A	PPB		26 Jun 2017 23:36:44	U	SMPL
I3883-28 - 1	G53-1.5-3	-2	-0.0058	-7470A	PPB		26 Jun 2017 23:38:53	U	SMPL
I3883-29 - 1	G40-0-1.5	-20	-0.0102	-7470A	PPB		26 Jun 2017 23:41:01	U	SMPL
I3883-30 - 1	G40-1.5-3	-27	-0.0119	-7470A	PPB		26 Jun 2017 23:43:09	U	SMPL
I3883-31 - 1	G29-0-1.5	-29	-0.0124	-7470A	PPB		26 Jun 2017 23:45:17	U	SMPL
I3883-32 - 1	G29-1.5-3	-15	-0.0090	-7470A	PPB		26 Jun 2017 23:47:25	U	SMPL
I3885-01 - 1	CAN-WC-S-1-0-11	-2	-0.0058	-7470A	PPB		26 Jun 2017 23:49:34	U	SMPL
I3904-01 - 1	CAN-WC-S-2-11-50	-3	-0.0061	-7470A	PPB		26 Jun 2017 23:51:42	U	SMPL
PB100180BL - 1	PBW	-14	-0.0087	-7470A	PPB		26 Jun 2017 23:53:50	U	SMPL
PB100180BS - 1	LCSW	14824	3.5901	-7470A	PPB		26 Jun 2017 23:55:58	U	SMPL
I3876-03 - 1	CAN-FB062217	-150	-0.0417	-7470A	PPB		26 Jun 2017 23:58:06	U	SMPL
CCV57 - 1	CCV57	19757	4.7865	-7470A	PPB		27 Jun 2017 00:00:14	U	SMPL
CCB57 - 1	CCB57	19	-0.0007	-7470A	PPB		27 Jun 2017 00:02:23	U	SMPL
I3884-01 - 1	CAN-GW-05	276151	66.9727	-7470A	PPB		27 Jun 2017 00:04:33	U	SMPL
I3884-01DUP - 1	CAN-GW-05DUP	279546	67.7961	-7470A	PPB		27 Jun 2017 00:06:41	U	SMPL
I3884-01MS - 1	CAN-GW-05MS	284026	68.8827	-7470A	PPB		27 Jun 2017 00:08:58	U	SMPL
I3884-01MSD - 1	CAN-GW-05MSD	283726	68.8099	-7470A	PPB		27 Jun 2017 00:11:16	U	SMPL
I3885-02 - 1	CAN-WC-L-1	143333	34.7588	-7470A	PPB		27 Jun 2017 00:13:32	U	SMPL
I3898-01 - 1	A6987MDW	41297	10.0109	-7470A	PPB		27 Jun 2017 00:15:49	U	SMPL
I3872-09LX5 - 1		-208	-0.0558	-7470A	PPB		27 Jun 2017 00:18:00	U	SMPL
I3872-09A - 1		14996	3.6318	-7470A	PPB		27 Jun 2017 00:20:10	U	SMPL
I3852-36LX5 - 1		-70	-0.0223	-7470A	PPB		27 Jun 2017 00:22:18	U	SMPL
I3852-36A - 1		15129	3.6641	-7470A	PPB		27 Jun 2017 00:24:28	U	SMPL
CCV58 - 1	CCV58	19821	4.8021	-7470A	PPB		27 Jun 2017 00:26:36	U	SMPL
CCB58 - 1	CCB58	16	-0.0015	-7470A	PPB		27 Jun 2017 00:28:45	U	SMPL
I3874-04LX5 - 1		-31	-0.0129	-7470A	PPB		27 Jun 2017 00:30:55	U	SMPL
I3874-04A - 1		16176	3.9180	-7470A	PPB		27 Jun 2017 00:33:04	U	SMPL
I3882-15LX5 - 1		-70	-0.0223	-7470A	PPB		27 Jun 2017 00:35:13	U	SMPL
I3882-15A - 1		15657	3.7921	-7470A	PPB		27 Jun 2017 00:37:21	U	SMPL
I3883-17LX5 - 1		-103	-0.0303	-7470A	PPB		27 Jun 2017 00:39:30	U	SMPL
I3883-17A - 1		15658	3.7924	-7470A	PPB		27 Jun 2017 00:41:38	U	SMPL
I3884-01LX5 - 1		55782	13.5241	-7470A	PPB		27 Jun 2017 00:43:46	U	SMPL
I3884-01A - 1		292451	70.9261	-7470A	PPB		27 Jun 2017 00:45:55	U	SMPL
CCV59 - 1	CCV59	19547	4.7356	-7470A	PPB		27 Jun 2017 00:48:06	U	SMPL
CCB59 - 1	CCB59	-44	-0.0160	-7470A	PPB		27 Jun 2017 00:50:25	U	SMPL