


7471A

LB88366
INSTRUMENT ID: CV1

Linear

 μ Abs.:

49135



A= 0.0000e+000

B= 2.0454e-004

C= -1.0709e-002

Rho= 0.9999244

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-40	0.000	-40				
0.2	0.200	0.196	-0.004	1012	0.0 %	1012				
2.5	2.500	2.540	0.040	12470	0.0 %	12470				
5.0	5.000	5.032	0.032	24654	0.0 %	24654				
7.5	7.500	7.412	-0.088	36288	0.0 %	36288	0			
10.0	10.000	10.039	0.039	49135	0.0 %	49135				

LB88366 INSTRUMENT ID :CV1

Method: 7471A

Operator: Admin

Date of Analysis: 27 Jun 2017 08:47:08

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-40	-	0.0000	7471A	PPB	27 Jun 2017 08:51:33	S	Std
0.2 - 1	50-2	1012	-	0.2000	7471A	PPB	27 Jun 2017 08:53:50	S	Std
2.5 - 1	52-5	12470	-	2.5000	7471A	PPB	27 Jun 2017 08:56:08	S	Std
5.0 - 1	55	24654	-	5.0000	7471A	PPB	27 Jun 2017 08:58:25	S	Std
7.5 - 1	57-5	36288	-	7.5000	7471A	PPB	27 Jun 2017 09:00:42	S	Std
10.0 - 1	50	49135	-	10.0000	7471A	PPB	27 Jun 2017 09:02:59	S	Std
ICV41 - 1	ICV41	19398	3.9569	-	7471A	PPB	27 Jun 2017 09:05:54	U	SMPL
ICB41 - 1	ICB41	-98	-0.0308	-	7471A	PPB	27 Jun 2017 09:08:10	U	SMPL
CCV60 - 1	CCV60	24607	5.0223	-	7471A	PPB	27 Jun 2017 09:13:00	U	SMPL
CCB60 - 1	CCB60	-97	-0.0305	-	7471A	PPB	27 Jun 2017 09:15:15	U	SMPL
CRA - 1	CRA	1067	0.2075	-	7471A	PPB	27 Jun 2017 09:17:31	U	SMPL
HighStd - 1	HighStd	48790	9.9687	-	7471A	PPB	27 Jun 2017 09:19:47	U	SMPL
ChkStd - 1	ChkStd	33062	6.7517	-	7471A	PPB	27 Jun 2017 09:22:03	U	SMPL
PB100159BL - 1	PBS	-119	-0.0350	-	7471A	PPB	27 Jun 2017 09:24:21	U	SMPL
PB100159BS - 1	LCSS	17059	3.4785	-	7471A	PPB	27 Jun 2017 09:26:38	U	SMPL
I3879-01 - 1	I16-005-MH	53654	10.9635	-	7471A	PPB	27 Jun 2017 09:28:55	U	SMPL
I3880-01 - 1	BF-1	-13	-0.0134	-	7471A	PPB	27 Jun 2017 09:31:12	U	SMPL
I3880-02 - 1	TS-1	3250	0.6540	-	7471A	PPB	27 Jun 2017 09:33:31	U	SMPL
I3893-01 - 1	WC062317-1E	1612	0.3190	-	7471A	PPB	27 Jun 2017 09:35:49	U	SMPL
I3893-04 - 1	WC062317-1SE	1550	0.3063	-	7471A	PPB	27 Jun 2017 09:38:05	U	SMPL
CCV61 - 1	CCV61	24167	4.9323	-	7471A	PPB	27 Jun 2017 09:40:21	U	SMPL
CCB61 - 1	CCB61	-74	-0.0258	-	7471A	PPB	27 Jun 2017 09:42:37	U	SMPL
I3893-07 - 1	WC062317-1S	1320	0.2593	-	7471A	PPB	27 Jun 2017 09:44:53	U	SMPL
I3893-10 - 1	WC062317-1SW	13414	2.7330	-	7471A	PPB	27 Jun 2017 09:47:09	U	SMPL
I3897-01 - 1	72-11931	1449	0.2857	-	7471A	PPB	27 Jun 2017 09:49:26	U	SMPL
I3898-04 - 1	VNJ-72-12002-VNJ-201-COMP	589	0.1098	-	7471A	PPB	27 Jun 2017 09:51:42	U	SMPL
I3899-01 - 1	72-11940	226	0.0355	-	7471A	PPB	27 Jun 2017 09:53:59	U	SMPL
I3906-01 - 1	EO-02-062317-A	1456	0.2871	-	7471A	PPB	27 Jun 2017 09:56:16	U	SMPL
I3906-04 - 1	EO-02-062317-B	2514	0.5035	-	7471A	PPB	27 Jun 2017 09:58:33	U	SMPL
I3906-07 - 1	EO-02-062317-C	2647	0.5307	-	7471A	PPB	27 Jun 2017 10:00:51	U	SMPL
I3909-01 - 1	E2-P10-ESW	9191	1.8692	-	7471A	PPB	27 Jun 2017 10:03:09	U	SMPL
I3909-02 - 1	E2-Q6-BASE	3205	0.6448	-	7471A	PPB	27 Jun 2017 10:05:25	U	SMPL
CCV62 - 1	CCV62	24492	4.9988	-	7471A	PPB	27 Jun 2017 10:07:41	U	SMPL
CCB62 - 1	CCB62	-112	-0.0336	-	7471A	PPB	27 Jun 2017 10:09:58	U	SMPL
I3909-03 - 1	E2-Q4-BASE	375383	76.7692	-	7471A	PPB	27 Jun 2017 10:12:14	U	SMPL
I3909-04 - 1	E2-Q9-BASE	407764	HIGH	-	7471A	PPB	27 Jun 2017 10:14:30	X	
I3911-01 - 1	SU-04-062317	586	0.1091	-	7471A	PPB	27 Jun 2017 10:16:31	U	SMPL
I3912-01 - 1	OK-03-062317	662	0.1247	-	7471A	PPB	27 Jun 2017 10:18:49	U	SMPL
I3913-01 - 1	TR-04-062317	783	0.1494	-	7471A	PPB	27 Jun 2017 10:21:06	U	SMPL
I3913-01DUP - 1	TR-04-062317DUP	707	0.1339	-	7471A	PPB	27 Jun 2017 10:23:23	U	SMPL
I3913-01MS - 1	TR-04-062317MS	18232	3.7184	-	7471A	PPB	27 Jun 2017 10:25:40	U	SMPL
I3913-01MSD - 1	TR-04-062317MSD	18164	3.7045	-	7471A	PPB	27 Jun 2017 10:27:57	U	SMPL
PB100160BL - 1	PBS	12	-0.0083	-	7471A	PPB	27 Jun 2017 10:30:16	U	SMPL
PB100160BS - 1	LCSS	16918	3.4497	-	7471A	PPB	27 Jun 2017 10:32:32	U	SMPL
CCV63 - 1	CCV63	24346	4.9690	-	7471A	PPB	27 Jun 2017 10:34:48	U	SMPL
CCB63 - 1	CCB63	-81	-0.0273	-	7471A	PPB	27 Jun 2017 10:37:05	U	SMPL
I3874-02 - 1	TL-WTP-02	5660	1.1470	-	7471A	PPB	27 Jun 2017 10:39:21	U	SMPL
I3874-03 - 1	TL-DR-01	5311	1.0756	-	7471A	PPB	27 Jun 2017 10:41:37	U	SMPL
I3876-04 - 1	CAN-SB-0203040506-0-11	11315	2.3036	-	7471A	PPB	27 Jun 2017 10:43:54	U	SMPL
I3876-04DUP - 1	CAN-SB-0203040506-0-11DUP	10968	2.2327	-	7471A	PPB	27 Jun 2017 10:46:10	U	SMPL
I3876-05 - 1	I3876-04MS	28288	5.7752	-	7471A	PPB	27 Jun 2017 10:48:27	U	SMPL
I3876-06 - 1	I3876-04MSD	27246	5.5621	-	7471A	PPB	27 Jun 2017 10:50:44	U	SMPL
I3876-07 - 1	CAN-SB-10203040506-0-11	12430	2.5317	-	7471A	PPB	27 Jun 2017 10:53:01	U	SMPL
I3882-01 - 1	G51-1.5-3	1608	0.3182	-	7471A	PPB	27 Jun 2017 10:55:19	U	SMPL
I3882-02 - 1	G51-0-1.5	8786	1.7864	-	7471A	PPB	27 Jun 2017 10:57:37	U	SMPL
I3882-03 - 1	G46-0-1.5	2316	0.4630	-	7471A	PPB	27 Jun 2017 10:59:53	U	SMPL
CCV64 - 1	CCV64	24332	4.9661	-	7471A	PPB	27 Jun 2017 11:02:10	U	SMPL
CCB64 - 1	CCB64	-89	-0.0289	-	7471A	PPB	27 Jun 2017 11:04:26	U	SMPL
I3882-04 - 1	G46-1.5-3	4945	1.0007	-	7471A	PPB	27 Jun 2017 11:06:42	U	SMPL
I3882-05 - 1	G52-0-1.5	5703	1.1558	-	7471A	PPB	27 Jun 2017 11:08:59	U	SMPL
I3882-06 - 1	G52-1.5-3	2409	0.4820	-	7471A	PPB	27 Jun 2017 11:11:15	U	SMPL
I3882-07 - 1	G43-0-1.5	2744	0.5505	-	7471A	PPB	27 Jun 2017 11:13:31	U	SMPL
I3882-08 - 1	G43-1.5-3	16152	3.2930	-	7471A	PPB	27 Jun 2017 11:15:48	U	SMPL
I3882-09 - 1	G47-0-1.5	10139	2.0631	-	7471A	PPB	27 Jun 2017 11:18:05	U	SMPL
I3882-10 - 1	G47-1.5-3	8488	1.7254	-	7471A	PPB	27 Jun 2017 11:20:22	U	SMPL
I3882-11 - 1	G56-0-1.5	9430	1.9181	-	7471A	PPB	27 Jun 2017 11:22:39	U	SMPL
I3882-12 - 1	G56-1.5-3	10714	2.1807	-	7471A	PPB	27 Jun 2017 11:24:57	U	SMPL
I3882-13 - 1	G54-0-1.5	24554	5.0115	-	7471A	PPB	27 Jun 2017 11:27:14	U	SMPL
CCV65 - 1	CCV65	24317	4.9630	-	7471A	PPB	27 Jun 2017 11:29:31	U	SMPL
CCB65 - 1	CCB65	-75	-0.0260	-	7471A	PPB	27 Jun 2017 11:31:47	U	SMPL
I3882-14 - 1	G54-1.5-3	4661	0.9426	-	7471A	PPB	27 Jun 2017 11:34:04	U	SMPL
I3894-01DLX10 - 1	1706075476-77-89-90-COMP	158	0.0216	-	7471A	PPB	27 Jun 2017 11:36:21	U	SMPL
I3905-02 - 1	CAN-SB-05-11-50	549	0.1016	-	7471A	PPB	27 Jun 2017 11:38:38	U	SMPL
PB100179BL - 1	PBS	-21	-0.0150	-	7471A	PPB	27 Jun 2017 11:40:54	U	SMPL
PB100179BS - 1	LCSS	17137	3.4944	-	7471A	PPB	27 Jun 2017 11:43:11	U	SMPL
I3883-01 - 1	G45-0-1.5	4572	0.9244	-	7471A	PPB	27 Jun 2017 11:45:28	U	SMPL
I3883-02 - 1	G45-1.5-3	3097	0.6227	-	7471A	PPB	27 Jun 2017 11:47:45	U	SMPL
I3883-03 - 1	G41-0-1.5	9717	1.9768	-	7471A	PPB	27 Jun 2017 11:50:03	U	SMPL
I3883-04 - 1	G41-1.5-3	1812	0.3599	-	7471A	PPB	27 Jun 2017 11:52:21	U	SMPL
I3883-05 - 1	G49-0-1.5	3015	0.6060	-	7471A	PPB	27 Jun 2017 11:54:38	U	SMPL
CCV66 - 1	CCV66	24465	4.9933	-	7471A	PPB	27 Jun 2017 11:56:55	U	SMPL
CCB66 - 1	CCB66	-54	-0.0218	-	7471A	PPB	27 Jun 2017 11:59:12	U	SMPL
I3883-06 - 1	G49-1.5-3	5667	1.1484	-	7471A	PPB	27 Jun 2017 12:01:28	U	SMPL
I3883-07 - 1	G42-0-1.5	4754	0.9617	-	7471A	PPB	27 Jun 2017 12:03:45	U	SMPL
I3883-08 - 1	G42-1.5-3	5542	1.1228	-	7471A	PPB	27 Jun 2017 12:06:02	U	SMPL
I3883-09 - 1	G44-0-1.5	2275	0.4546	-	7471A	PPB	27 Jun 2017 12:08:19	U	SMPL
I3883-10 - 1	G44-1.5-3	1934	0.3849	-	7471A	PPB	27 Jun 2017 12:10:36	U	SMPL
I3883-11 - 1	G53-0-1.5	7278	1.4779	-	7471A	PPB	27 Jun 2017 12:12:53	U	SMPL
I3883-12 - 1	G53-1.5-3	6067	1.2302	-	7471A	PPB	27 Jun 2017 12:15:11	U	SMPL
I3883-13 - 1	G40-0-1.5	3105	0.6244	-	7471A	PPB	27 Jun 2017 12:17:28	U	SMPL
I3883-14 - 1	G40-1.5-3	1193	0.2333	-	7471A	PPB	27 Jun 2017 12:19:46	U	SMPL
I3883-15 - 1	G29-0-1.5	7131	1.4478	-	7471A	PPB	27 Jun 2017 12:22:03	U	SMPL
CCV67 - 1	CCV67	24337	4.9671	-	7471A	PPB	27 Jun 2017 12:24:20	U	SMPL

LB88366 INSTRUMENT ID :CV1

Method: 7471A		Operator: Admin		Date of Analysis: 27 Jun 2017 08:47:08															
Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type										
CCB67 - 1	CCB67	-68	-0.0246	-7471A	PPB		27 Jun 2017 12:26:37	U	SMPL										
I3883-16 - 1	G29-1.5-3	13460	2.7424	-7471A	PPB		27 Jun 2017 12:28:54	U	SMPL										
I3883-16DUP - 1	G29-1.5-3DUP	11378	2.3165	-7471A	PPB		27 Jun 2017 12:31:11	U	SMPL										
I3883-16MS - 1	G29-1.5-3MS	29232	5.9683	-7471A	PPB		27 Jun 2017 12:37:17	U	SMPL										
I3883-16MSD - 1	G29-1.5-3MSD	31256	6.3823	-7471A	PPB		27 Jun 2017 12:39:33	U	SMPL										
PB100187BL - 1	PBS	-93	-0.0297	-7471A	PPB		27 Jun 2017 12:41:50	U	SMPL										
PB100187BS - 1	LCSS	17726	3.6149	-7471A	PPB		27 Jun 2017 12:44:08	U	SMPL										
I3871-01 - 1	POSTEX-50-20170622	20915	4.2672	-7471A	PPB		27 Jun 2017 12:46:25	U	SMPL										
I3871-02 - 1	POSTEX-DUP3-20170622	26875	5.4862	-7471A	PPB		27 Jun 2017 12:48:42	U	SMPL										
I3871-03 - 1	POSTEX-51-20170622	22259	4.5421	-7471A	PPB		27 Jun 2017 12:51:00	U	SMPL										
I3871-03DUP - 1	POSTEX-51-20170622DUP	22088	4.5071	-7471A	PPB		27 Jun 2017 12:53:17	U	SMPL										
CCV68 - 1	CCV68	24699	5.0412	-7471A	PPB		27 Jun 2017 12:55:34	U	SMPL										
CCB68 - 1	CCB68	-34	-0.0177	-7471A	PPB		27 Jun 2017 12:57:51	U	SMPL										
I3871-04 - 1	I3871-03MS	37089	7.5754	-7471A	PPB		27 Jun 2017 13:00:08	U	SMPL										
I3871-05 - 1	I3871-03MSD	39288	8.0252	-7471A	PPB		27 Jun 2017 13:02:25	U	SMPL										
I3871-06 - 1	POSTEX-52-20170622	11790	2.4008	-7471A	PPB		27 Jun 2017 13:04:42	U	SMPL										
I3871-07 - 1	POSTEX-53-20170622	19313	3.9395	-7471A	PPB		27 Jun 2017 13:07:00	U	SMPL										
I3871-08 - 1	POSTEX-54-20170622	13808	2.8135	-7471A	PPB		27 Jun 2017 13:09:17	U	SMPL										
I3920-01 - 1	OB-1	25034	5.1097	-7471A	PPB		27 Jun 2017 13:11:34	U	SMPL										
I3920-02 - 1	OB-2	15321	3.1230	-7471A	PPB		27 Jun 2017 13:13:51	U	SMPL										
I3920-03 - 1	OB-3	12414	2.5284	-7471A	PPB		27 Jun 2017 13:16:08	U	SMPL										
I3920-04 - 1	OB-4	12619	2.5703	-7471A	PPB		27 Jun 2017 13:18:26	U	SMPL										
I3923-01DLX10 - 1	1706165525	-67	-0.0244	-7471A	PPB		27 Jun 2017 13:20:43	U	SMPL										
CCV69 - 1	CCV69	24596	5.0201	-7471A	PPB		27 Jun 2017 13:23:01	U	SMPL										
CCB69 - 1	CCB69	-87	-0.0285	-7471A	PPB		27 Jun 2017 13:25:18	U	SMPL										
I3923-02DLX10 - 1	17K2758	-25	-0.0158	-7471A	PPB		27 Jun 2017 13:27:36	U	SMPL										
I3924-01DLX10 - 1	1706195526-27-COMP	-4	-0.0115	-7471A	PPB		27 Jun 2017 13:29:52	U	SMPL										
I3924-02DLX10 - 1	69234M	14	-0.0078	-7471A	PPB		27 Jun 2017 13:32:10	U	SMPL										
I3928-01 - 1	E2-Q6-VSW	11745	2.3916	-7471A	PPB		27 Jun 2017 13:34:26	U	SMPL										
I3928-02 - 1	E2-Q7-BASE	1153319	HIGH	-7471A	PPB		27 Jun 2017 13:36:44	X											
I3928-03 - 1	E2-Q8-BASE	1235527	HIGH	-7471A	PPB		27 Jun 2017 13:38:35	X											
I3928-04 - 1	E2-CONCRETE-BATCH-14	17899	3.6503	-7471A	PPB		27 Jun 2017 13:41:31	U	SMPL										
I3879-01DLX2 - 1		26467	5.4028	-7471A	PPB		27 Jun 2017 13:45:33	U	SMPL										
I3909-03DLX10 - 1		44651	9.1221	-7471A	PPB		27 Jun 2017 13:50:41	U	SMPL										
I3909-04DLX10 - 1		67007	13.6947	-7471A	PPB		27 Jun 2017 13:52:57	U	SMPL										
CCV70 - 1	CCV70	25408	5.1862	-7471A	PPB		27 Jun 2017 13:55:16	U	SMPL										
CCB70 - 1	CCB70	-300	-0.0721	-7471A	PPB		27 Jun 2017 13:57:35	U	SMPL										
I3913-01LX5 - 1		58	0.0012	-7471A	PPB		27 Jun 2017 13:59:53	U	SMPL										
I3913-01A - 1		20191	4.1191	-7471A	PPB		27 Jun 2017 14:02:11	U	SMPL										
I3876-04LX5 - 1		2224	0.4442	-7471A	PPB		27 Jun 2017 14:04:28	U	SMPL										
I3876-04A - 1		31537	6.4398	-7471A	PPB		27 Jun 2017 14:06:46	U	SMPL										
I3883-16LX5 - 1		2612	0.5235	-7471A	PPB		27 Jun 2017 14:09:03	U	SMPL										
I3883-16A - 1		33070	6.7533	-7471A	PPB		27 Jun 2017 14:11:21	U	SMPL										
I3871-03LX5 - 1		4403	0.8899	-7471A	PPB		27 Jun 2017 14:13:39	U	SMPL										
I3871-03A - 1		41972	8.5741	-7471A	PPB		27 Jun 2017 14:15:57	U	SMPL										
I3928-02DLX10 - 1		466827	HIGH	-7471A	PPB		27 Jun 2017 14:18:15	X											
I3928-03DLX10 - 1		638972	HIGH	-7471A	PPB		27 Jun 2017 14:20:09	X											
CCV71 - 1	CCV71	24690	5.0393	-7471A	PPB		27 Jun 2017 14:22:05	U	SMPL										
CCB71 - 1	CCB71	-261	-0.0641	-7471A	PPB		27 Jun 2017 14:24:30	U	SMPL										
I3909-04DL2X50 - 1		13467	2.7438	-7471A	PPB		27 Jun 2017 14:26:49	U	SMPL										
I3928-02DL2X200 - 1		38584	7.8812	-7471A	PPB		27 Jun 2017 14:29:07	U	SMPL										
I3928-03DL2X500 - 1		25440	5.1927	-7471A	PPB		27 Jun 2017 14:39:00	U	SMPL										
CCV72 - 1	CCV72	25652	5.2361	-7471A	PPB		27 Jun 2017 14:41:17	U	SMPL										
CCB72 - 1	CCB72	-155	-0.0424	-7471A	PPB		27 Jun 2017 14:43:35	U	SMPL										