



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB88368

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	06/27/2017	09:00
2	I3808-05		1	>100.0	>212.0	06/27/2017	09:30
3	I3808-05DUP		1	>100.0	>212.0	06/27/2017	10:00
4	I3867-02		1	>100.0	>212.0	06/27/2017	10:20
5	I3885-02		1	>100.0	>212.0	06/27/2017	10:45
6	I3898-02		1	>100.0	>212.0	06/27/2017	11:10

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP62617

WorkList ID : 100100

Date : 6/26/2017 7:27:52 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/26/2017	Water	I3808-05	Flash Point	Cool 4 deg C	PSEG01	N43	DRUM-001-003-004-COMP	06/21/2017	1010A
	Water	I3867-02	Flash Point	Cool 4 deg C	PSEG01	N61	A6229	06/22/2017	1010A
	Water	I3885-02	Flash Point	Cool 4 deg C	PARS15	N53	CAN-WC-L-1	06/22/2017	1010A
06/28/2017	Water	I3898-02	Flash Point	Cool 4 deg C	PSEG01	L51	A6987MDW	06/23/2017	1010A

Date/Time

Received by:

Relinquished by:

Date/Time

Received by:

Relinquished by: