



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88381

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/27/2017	15:30
2	CAL2	1	Water			7.00	06/27/2017	15:31
3	CAL3	1	Water			10.02	06/27/2017	15:32
4	ICV	1	Water			7.00	06/27/2017	15:34
5	CCV1	1	Water			2.01	06/27/2017	15:35
6	I3930-13	1	Solid	20.03	20	7.44	06/27/2017	15:36
7	I3930-13DUP	1	Solid	20.04	20	7.45	06/27/2017	15:37
8	I3930-14	1	Solid	20.01	20	7.09	06/27/2017	15:39
9	I3930-15	1	Solid	20.03	20	6.87	06/27/2017	15:40
10	I3930-16	1	Solid	20.01	20	7.20	06/27/2017	15:41
11	I3930-17	1	Solid	20.03	20	7.89	06/27/2017	15:42
12	I3930-18	1	Solid	20.05	20	6.45	06/27/2017	15:43
13	I3930-19	1	Solid	20.04	20	6.52	06/27/2017	15:44
14	I3930-20	1	Solid	20.02	20	8.46	06/27/2017	15:45
15	I3930-21	1	Solid	20.03	20	7.63	06/27/2017	15:46
16	CCV2	1	Water			12.02	06/27/2017	15:47
17	I3930-22	1	Solid	20.02	20	7.83	06/27/2017	15:48
18	I3930-23	1	Solid	20.03	20	7.35	06/27/2017	15:49
19	I3930-24	1	Solid	20.04	20	7.62	06/27/2017	15:50
20	I3935-03	1	Solid	20.02	20	6.52	06/27/2017	15:51

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I3947-02	1	Solid	20.03	20	8.90	06/27/2017	15:52
22	I3949-05	1	Solid	20.02	20	7.02	06/27/2017	15:53
23	I3949-11	1	Solid	20.04	20	5.39	06/27/2017	15:54
24	I3952-05	1	Solid	20.01	20	8.10	06/27/2017	15:55
25	I3952-05DUP	1	Solid	20.03	20	8.11	06/27/2017	15:56
26	CCV3	1	Water			12.02	06/27/2017	15:57

WORKLIST(Hardcopy Internal Chain)

1388381

WorkList Name : CORROSIONITY I3930 WorkList ID : 100131

Date : 6/27/2017 11:35:46 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3930-13	Corrosivity	Cool 4 deg C	TECT01	N63	B2-0-2	06/26/2017	9045C
	Solid	I3930-14	Corrosivity	Cool 4 deg C	TECT01	N63	B2-2-4	06/26/2017	9045C
	Solid	I3930-15	Corrosivity	Cool 4 deg C	TECT01	N63	B2-4-6	06/26/2017	9045C
	Solid	I3930-16	Corrosivity	Cool 4 deg C	TECT01	N63	B2-6-8	06/26/2017	9045C
	Solid	I3930-17	Corrosivity	Cool 4 deg C	TECT01	N63	B2-8-10	06/26/2017	9045C
	Solid	I3930-18	Corrosivity	Cool 4 deg C	TECT01	N63	B2-10-12.5	06/26/2017	9045C
	Solid	I3930-19	Corrosivity	Cool 4 deg C	TECT01	N63	B2-12.5-15	06/26/2017	9045C
	Solid	I3930-20	Corrosivity	Cool 4 deg C	TECT01	N63	B3-0-2	06/26/2017	9045C
	Solid	I3930-21	Corrosivity	Cool 4 deg C	TECT01	N63	B3-2-4	06/26/2017	9045C
	Solid	I3930-22	Corrosivity	Cool 4 deg C	TECT01	N63	B3-4-6	06/26/2017	9045C
	Solid	I3930-23	Corrosivity	Cool 4 deg C	TECT01	N63	B3-6-8.5	06/26/2017	9045C
	Solid	I3930-24	Corrosivity	Cool 4 deg C	TECT01	N63	B3-8.5-11	06/26/2017	9045C
	Solid	I3935-03	Corrosivity	Cool 4 deg C	DVIR01	N41	JFK-SOIL-06-26-17(COMPOS	06/26/2017	9045C
06/30/2017	Solid	I3947-02	Corrosivity	Cool 4 deg C	PSEG01	P31	TP1-TP2-COMP	06/27/2017	9045C
06/30/2017	Solid	I3949-05	Corrosivity	Cool 4 deg C	PSEG01	P31	TP-H1-(0-8)	06/27/2017	9045C
06/30/2017	Solid	I3949-11	Corrosivity	Cool 4 deg C	PSEG01	P31	TP-C2-(0-8)	06/27/2017	9045C
	Solid	I3952-05	Corrosivity	Cool 4 deg C	DEWB01	N62	23MC-WC-2(COMP)	06/27/2017	9045C

Date/Time 06/27/15 11:40 AM

Received by: HL

Relinquished by: HL

Date/Time 06/27/15 7:10 PM

Received by: HL

Relinquished by: SP