


7470A

LB 88397
INSTRUMENT ID: CVA

Linear

 μ Abs.:

41708



A= 0.0000e+000

B= 2.4136e-004

C= -5.9601e-003

Rho= 0.9998766

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	-26	0.000	-26				
0.2	0.200	0.197	-0.003	842	0.0 %	842				
2.5	2.500	2.542	0.042	10556	0.0 %	10556				
5.0	5.000	5.030	0.030	20864	0.0 %	20864				
7.5	7.500	7.383	-0.117	30614	0.0 %	30614				
10.0	10.000	10.061	0.061	41708	0.0 %	41708				

LB88397 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Jun 2017 10:05:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-26	-	0.00007470A	PPB		28 Jun 2017 10:42:13	S	Std
0.2 - 1	50.2	842	-	0.20007470A	PPB		28 Jun 2017 10:44:21	S	Std
2.5 - 1	52.5	10556	-	2.50007470A	PPB		28 Jun 2017 10:46:29	S	Std
5.0 - 1	55	20864	-	5.00007470A	PPB		28 Jun 2017 10:48:38	S	Std
7.5 - 1	57.5	30614	-	7.50007470A	PPB		28 Jun 2017 10:50:46	S	Std
10.0 - 1	510	41708	-	10.00007470A	PPB		28 Jun 2017 10:52:54	S	Std
ICV42 - 1	ICV42	15562	3.7500	-7470A	PPB		28 Jun 2017 10:55:39	U	SMPL
ICB42 - 1	ICB42	-97	-0.0294	-7470A	PPB		28 Jun 2017 10:57:46	U	SMPL
CCV73 - 1	CCV73	19913	4.8002	-7470A	PPB		28 Jun 2017 10:59:52	U	SMPL
CCB73 - 1	CCB73	-81	-0.0255	-7470A	PPB		28 Jun 2017 11:01:59	U	SMPL
CRA - 1	CRA	862	0.2021	-7470A	PPB		28 Jun 2017 11:04:06	U	SMPL
HighStd - 1	HighStd	40395	9.7436	-7470A	PPB		28 Jun 2017 11:06:12	U	SMPL
ChkStd - 1	ChkStd	27292	6.5811	-7470A	PPB		28 Jun 2017 11:08:19	U	SMPL
PB100217BL - 1	PBW	-119	-0.0347	-7470A	PPB		28 Jun 2017 11:10:27	U	SMPL
PB100217BS - 1	LCSW	16943	4.0833	-7470A	PPB		28 Jun 2017 11:12:36	U	SMPL
PB100217TB - 1	PB100177TB	-73	-0.0236	-7470A	PPB		28 Jun 2017 11:14:43	U	SMPL
I3783-09 - 1	CS-03	-26	-0.0122	-7470A	PPB		28 Jun 2017 11:16:52	U	SMPL
I3908-13 - 1	G-55-0-1.5	-34	-0.0142	-7470A	PPB		28 Jun 2017 11:19:00	U	SMPL
I3908-14 - 1	G-55-1.5-3	-18	-0.0103	-7470A	PPB		28 Jun 2017 11:21:09	U	SMPL
I3908-15 - 1	G-57-0-1.5	-30	-0.0132	-7470A	PPB		28 Jun 2017 11:23:16	U	SMPL
CCV74 - 1	CCV74	20152	4.8579	-7470A	PPB		28 Jun 2017 11:25:22	U	SMPL
CCB74 - 1	CCB74	-74	-0.0238	-7470A	PPB		28 Jun 2017 11:27:29	U	SMPL
I3908-16 - 1	G-57-1.5-3	-10	-0.0084	-7470A	PPB		28 Jun 2017 11:29:36	U	SMPL
I3908-17 - 1	G-60-0-1.5	-4	-0.0069	-7470A	PPB		28 Jun 2017 11:31:43	U	SMPL
I3908-18 - 1	G-60-1.5-3	-13	-0.0091	-7470A	PPB		28 Jun 2017 11:33:50	U	SMPL
I3908-19 - 1	G-59-0-1.5	-19	-0.0105	-7470A	PPB		28 Jun 2017 11:35:57	U	SMPL
I3908-20 - 1	G-59-1.5-3	-18	-0.0103	-7470A	PPB		28 Jun 2017 11:38:04	U	SMPL
I3908-21 - 1	G-58-0-1.5	-18	-0.0103	-7470A	PPB		28 Jun 2017 11:40:12	U	SMPL
I3908-22 - 1	G-58-1.5-3	-26	-0.0122	-7470A	PPB		28 Jun 2017 11:42:20	U	SMPL
I3908-23 - 1	G-61-0-1.5	-29	-0.0130	-7470A	PPB		28 Jun 2017 11:44:28	U	SMPL
I3908-24 - 1	G-61-1.5-3	-27	-0.0125	-7470A	PPB		28 Jun 2017 11:46:37	U	SMPL
I3925-03 - 1	HD-01-062617	-19	-0.0105	-7470A	PPB		28 Jun 2017 11:48:43	U	SMPL
CCV75 - 1	CCV75	19880	4.7922	-7470A	PPB		28 Jun 2017 11:50:50	U	SMPL
CCB75 - 1	CCB75	-86	-0.0267	-7470A	PPB		28 Jun 2017 11:52:57	U	SMPL
I3931-03 - 1	OR-01-062617	-53	-0.0188	-7470A	PPB		28 Jun 2017 11:55:05	U	SMPL
I3932-05 - 1	TP-C1-(0-8)	-28	-0.0127	-7470A	PPB		28 Jun 2017 11:57:12	U	SMPL
I3932-05DUP - 1	TP-C1-(0-8)DUP	-16	-0.0098	-7470A	PPB		28 Jun 2017 11:59:19	U	SMPL
I3932-05MS - 1	TP-C1-(0-8)MS	14860	3.5806	-7470A	PPB		28 Jun 2017 12:01:26	U	SMPL
I3932-05MSD - 1	TP-C1-(0-8)MSD	14613	3.5210	-7470A	PPB		28 Jun 2017 12:03:33	U	SMPL
I3932-05LX5 - 1		-83	-0.0260	-7470A	PPB		28 Jun 2017 12:05:41	U	SMPL
I3932-05A - 1		15785	3.8039	-7470A	PPB		28 Jun 2017 12:07:49	U	SMPL
CCV76 - 1	CCV76	20257	4.8832	-7470A	PPB		28 Jun 2017 12:09:58	U	SMPL
CCB76 - 1	CCB76	187	0.0392	-7470A	PPB		28 Jun 2017 12:12:06	U	SMPL
I3884-01DLX10 - 1		30130	7.2661	-7470A	PPB		28 Jun 2017 12:14:14	U	SMPL
I3884-01DUPDLX10 - 1		29435	7.0984	-7470A	PPB		28 Jun 2017 12:16:21	U	SMPL
I3884-01MSDLX10 - 1		30440	7.3409	-7470A	PPB		28 Jun 2017 12:18:29	U	SMPL
I3884-01MSDDLX10 - 1		30188	7.2801	-7470A	PPB		28 Jun 2017 12:20:37	U	SMPL
I3884-01LDLX50 - 1		6094	1.4649	-7470A	PPB		28 Jun 2017 12:22:45	U	SMPL
I3884-01ADLX10 - 1		31424	7.5784	-7470A	PPB		28 Jun 2017 12:24:53	U	SMPL
I3885-02DLX5 - 1		28939	6.9787	-7470A	PPB		28 Jun 2017 12:27:00	U	SMPL
I3898-01DLX2 - 1		21409	5.1612	-7470A	PPB		28 Jun 2017 12:29:08	U	SMPL
CCV77 - 1	CCV77	19772	4.7661	-7470A	PPB		28 Jun 2017 12:31:16	U	SMPL
CCB77 - 1	CCB77	-96	-0.0291	-7470A	PPB		28 Jun 2017 12:33:24	U	SMPL