

7471A


LB88407

INSTRUMENT ID: CV1

Linear

 μ Abs.:

48945



A= 0.0000e+000

B= 2.0535e-004

C= -4.8453e-002

Rho= 0.9999641

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	284	0.000	284				
0.2	0.200	0.226	0.026	1338	0.0 %	1338				
2.5	2.500	2.477	-0.023	12300	0.0 %	12300				
5.0	5.000	4.945	-0.055	24314	0.0 %	24314				
7.5	7.500	7.538	0.038	36945	0.0 %	36945	0			
10.0	10.000	10.003	0.003	48949	0.0 %	48949				

LB88407 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 28 Jun 2017 15:45:51			
Sample ID	Extended ID	μ Abs	Conc.	Std ConcMethod	Units	Date	Type
0.0 - 1	50	284	-	0.0007471A	PPB	28 Jun 2017 15:50:03	S Std
0.2 - 1	50.2	1338	-	0.20007471A	PPB	28 Jun 2017 15:52:21	S Std
2.5 - 1	52.5	12300	-	2.50007471A	PPB	28 Jun 2017 15:54:39	S Std
5.0 - 1	55	24314	-	5.00007471A	PPB	28 Jun 2017 15:56:57	S Std
7.5 - 1	57.5	36945	-	7.50007471A	PPB	28 Jun 2017 15:59:14	S Std
10.0 - 1	510	48949	-	10.00007471A	PPB	28 Jun 2017 16:01:31	S Std
ICV43 - 1	ICV43	18754	3.8028	-7471A	PPB	28 Jun 2017 16:04:13	U SMPL
ICB43 - 1	ICB43	205	-0.0064	-7471A	PPB	28 Jun 2017 16:06:29	U SMPL
CCV78 - 1	CCV78	22902	4.6546	-7471A	PPB	28 Jun 2017 16:08:45	U SMPL
CCB78 - 1	CCB78	217	-0.0039	-7471A	PPB	28 Jun 2017 16:11:00	U SMPL
CRA - 1	CRA	1357	0.2302	-7471A	PPB	28 Jun 2017 16:13:17	U SMPL
HighStd - 1	HighStd	46988	9.6008	-7471A	PPB	28 Jun 2017 16:15:33	U SMPL
ChkStd - 1	ChkStd	33266	6.7829	-7471A	PPB	28 Jun 2017 16:17:49	U SMPL
PB100221BL - 1	PBS	126	-0.0226	-7471A	PPB	28 Jun 2017 16:20:07	U SMPL
PB100221BS - 1	LCSS	17230	3.4898	-7471A	PPB	28 Jun 2017 16:22:24	U SMPL
I3908-01 - 1	G-55-0-1.5	3367	0.6430	-7471A	PPB	28 Jun 2017 16:24:41	U SMPL
I3908-02 - 1	G-55-1.5-3	12779	2.5758	-7471A	PPB	28 Jun 2017 16:26:58	U SMPL
I3908-03 - 1	G-57-0-1.5	4638	0.9040	-7471A	PPB	28 Jun 2017 16:29:16	U SMPL
I3908-04 - 1	G-57-1.5-3	6165	1.2176	-7471A	PPB	28 Jun 2017 16:31:34	U SMPL
I3908-05 - 1	G-60-0-1.5	6299	1.2451	-7471A	PPB	28 Jun 2017 16:33:50	U SMPL
CCV79 - 1	CCV79	23732	4.8250	-7471A	PPB	28 Jun 2017 16:36:06	U SMPL
CCB79 - 1	CCB79	209	-0.0055	-7471A	PPB	28 Jun 2017 16:38:21	U SMPL
I3908-06 - 1	G-60-1.5-3	1119	0.1813	-7471A	PPB	28 Jun 2017 16:40:37	U SMPL
I3908-07 - 1	G-59-0-1.5	3314	0.6321	-7471A	PPB	28 Jun 2017 16:42:53	U SMPL
I3908-07DUP - 1	G-59-0-1.5DUP	3102	0.5886	-7471A	PPB	28 Jun 2017 16:45:09	U SMPL
I3908-07MS - 1	G-59-0-1.5MS	20691	4.2005	-7471A	PPB	28 Jun 2017 16:47:26	U SMPL
I3908-07MSD - 1	G-59-0-1.5MSD	20832	4.2295	-7471A	PPB	28 Jun 2017 16:50:04	U SMPL
I3908-08 - 1	G-59-1.5-3	3320	0.6333	-7471A	PPB	28 Jun 2017 16:58:20	U SMPL
I3908-09 - 1	G-58-0-1.5	12369	2.4916	-7471A	PPB	28 Jun 2017 17:00:38	U SMPL
I3908-10 - 1	G-58-1.5-3	15931	3.2231	-7471A	PPB	28 Jun 2017 17:02:55	U SMPL
I3908-11 - 1	G-61-0-1.5	625	0.0799	-7471A	PPB	28 Jun 2017 17:05:13	U SMPL
I3908-12 - 1	G-61-1.5-3	10797	2.1688	-7471A	PPB	28 Jun 2017 17:07:28	U SMPL
CCV80 - 1	CCV80	23591	4.7961	-7471A	PPB	28 Jun 2017 17:09:44	U SMPL
CCB80 - 1	CCB80	211	-0.0051	-7471A	PPB	28 Jun 2017 17:12:00	U SMPL
I3929-01 - 1	EX-25	3868	0.7459	-7471A	PPB	28 Jun 2017 17:14:16	U SMPL
I3929-02 - 1	EX-26	12693	2.5581	-7471A	PPB	28 Jun 2017 17:16:33	U SMPL
I3929-03 - 1	EX-27	5657	1.1132	-7471A	PPB	28 Jun 2017 17:18:49	U SMPL
I3929-04 - 1	EX-28	2917	0.5506	-7471A	PPB	28 Jun 2017 17:21:05	U SMPL
I3929-05 - 1	EX-29	865	0.1292	-7471A	PPB	28 Jun 2017 17:23:22	U SMPL
I3929-06 - 1	EX-30	2125	0.3879	-7471A	PPB	28 Jun 2017 17:25:39	U SMPL
I3929-07 - 1	EX-31	551	0.0647	-7471A	PPB	28 Jun 2017 17:27:56	U SMPL
I3929-08 - 1	EX-32	1504	0.2604	-7471A	PPB	28 Jun 2017 17:30:14	U SMPL
PB100222BL - 1	PBS	54	-0.0374	-7471A	PPB	28 Jun 2017 17:32:32	U SMPL
PB100222BS - 1	LCSS	16917	3.4255	-7471A	PPB	28 Jun 2017 17:34:48	U SMPL
CCV81 - 1	CCV81	24172	4.9154	-7471A	PPB	28 Jun 2017 17:39:55	U SMPL
CCB81 - 1	CCB81	224	-0.0025	-7471A	PPB	28 Jun 2017 17:42:11	U SMPL
I3925-01 - 1	HD-01-062617	5074	0.9935	-7471A	PPB	28 Jun 2017 17:44:27	U SMPL
I3925-01MS - 1	HD-01-062617MS	21374	4.3408	-7471A	PPB	28 Jun 2017 17:49:00	U SMPL
I3925-01MSD - 1	HD-01-062617MSD	23996	4.8792	-7471A	PPB	28 Jun 2017 17:51:16	U SMPL
I3925-01DUP - 1	HD-01-062617DUP	5026	0.9837	-7471A	PPB	28 Jun 2017 17:54:06	U SMPL
I3926-01 - 1	SP1-A-COMP	1628	0.2859	-7471A	PPB	28 Jun 2017 17:56:44	U SMPL
I3926-06 - 1	SP1-B-COMP	1333	0.2253	-7471A	PPB	28 Jun 2017 17:59:00	U SMPL
I3931-01 - 1	OR-01-062617	2107	0.3842	-7471A	PPB	28 Jun 2017 18:01:17	U SMPL
I3932-01 - 1	TP-C1-(0-8)	8500	1.6971	-7471A	PPB	28 Jun 2017 18:03:34	U SMPL
I3935-02 - 1	JFK-SOIL-06-26-17(COMPOSITE)	3927	0.7580	-7471A	PPB	28 Jun 2017 18:05:52	U SMPL
I3941-01 - 1	LSS-DEBRIS-COMP	14	-0.0456	-7471A	PPB	28 Jun 2017 18:08:09	U SMPL
CCV82 - 1	CCV82	23717	4.8220	-7471A	PPB	28 Jun 2017 18:10:25	U SMPL
CCB82 - 1	CCB82	229	-0.0014	-7471A	PPB	28 Jun 2017 18:12:41	U SMPL
I3941-03 - 1	LSS-STONE-COMP-1	312	0.0156	-7471A	PPB	28 Jun 2017 18:14:58	U SMPL
I3941-05 - 1	LSS-STONE-COMP-2	149	-0.0179	-7471A	PPB	28 Jun 2017 18:17:14	U SMPL
I3947-01 - 1	TP1-TP2-COMP	11242	2.2601	-7471A	PPB	28 Jun 2017 18:21:47	U SMPL
I3949-01 - 1	TP-H1-(0-8)	13360	2.6951	-7471A	PPB	28 Jun 2017 18:24:04	U SMPL
I3949-07 - 1	TP-C2-(0-8)	7530	1.4979	-7471A	PPB	28 Jun 2017 18:26:21	U SMPL
I3951-01 - 1	DUCT-BANK	77336	15.8329	-7471A	PPB	28 Jun 2017 18:28:38	U SMPL
I3952-01 - 1	23MC-WC-2(COMP)	28282	5.7594	-7471A	PPB	28 Jun 2017 18:30:55	U SMPL
PB100223BL - 1	PBS	-85	-0.0659	-7471A	PPB	28 Jun 2017 18:33:14	U SMPL
PB100223BS - 1	LCSS	16782	3.3978	-7471A	PPB	28 Jun 2017 18:35:31	U SMPL
CCV83 - 1	CCV83	22927	4.6597	-7471A	PPB	28 Jun 2017 18:37:47	U SMPL
CCB83 - 1	CCB83	288	0.0107	-7471A	PPB	28 Jun 2017 18:40:04	U SMPL
I3930-01 - 1	B2-0-2	5770	1.1364	-7471A	PPB	28 Jun 2017 18:42:20	U SMPL
I3930-02 - 1	B2-2-4	3294	0.6280	-7471A	PPB	28 Jun 2017 18:44:37	U SMPL
I3930-03 - 1	B2-4-6	37407	7.6333	-7471A	PPB	28 Jun 2017 18:46:54	U SMPL
I3930-04 - 1	B2-6-8	371120	76.1629	-7471A	PPB	28 Jun 2017 18:49:11	U SMPL
I3930-05 - 1	B2-8-10	183338	37.6009	-7471A	PPB	28 Jun 2017 18:51:28	U SMPL
I3930-06 - 1	B2-10-12.5	12803	2.5807	-7471A	PPB	28 Jun 2017 18:53:52	U SMPL
I3930-07 - 1	B2-12.5-15	9394	1.8807	-7471A	PPB	28 Jun 2017 18:56:12	U SMPL
I3930-08 - 1	B3-0-2	7826	1.5587	-7471A	PPB	28 Jun 2017 18:58:29	U SMPL
I3930-09 - 1	B3-2-4	6210	1.2268	-7471A	PPB	28 Jun 2017 19:00:47	U SMPL
I3930-09DUP - 1	B3-2-4DUP	6262	1.2375	-7471A	PPB	28 Jun 2017 19:03:04	U SMPL
CCV84 - 1	CCV84	23768	4.8324	-7471A	PPB	28 Jun 2017 19:05:21	U SMPL
CCB84 - 1	CCB84	336	0.0205	-7471A	PPB	28 Jun 2017 19:07:38	U SMPL
I3930-09MS - 1	B3-2-4MS	24233	4.9279	-7471A	PPB	28 Jun 2017 19:09:55	U SMPL
I3930-09MSD - 1	B3-2-4MSD	25106	5.1072	-7471A	PPB	28 Jun 2017 19:12:12	U SMPL
I3930-10 - 1	B3-4-6	7134	1.4165	-7471A	PPB	28 Jun 2017 19:14:29	U SMPL
I3930-11 - 1	B3-6-8.5	8944	1.7882	-7471A	PPB	28 Jun 2017 19:16:46	U SMPL
I3930-12 - 1	B3-8.5-11	6728	1.3332	-7471A	PPB	28 Jun 2017 19:19:03	U SMPL
I3946-01DLX10 - 1	170695528-29-COMP	335	0.0203	-7471A	PPB	28 Jun 2017 19:21:20	U SMPL
CCV85 - 1	CCV85	23737	4.8261	-7471A	PPB	28 Jun 2017 19:23:38	U SMPL
CCB85 - 1	CCB85	335	0.0203	-7471A	PPB	28 Jun 2017 19:25:55	U SMPL
I3908-07LX5 - 1		860	0.1282	-7471A	PPB	28 Jun 2017 19:28:13	U SMPL
I3908-07A - 1		22232	4.5170	-7471A	PPB	28 Jun 2017 19:30:30	U SMPL
I3925-01LX5 - 1		1098	0.1770	-7471A	PPB	28 Jun 2017 19:32:47	U SMPL
I3925-01A - 1		23208	4.7174	-7471A	PPB	28 Jun 2017 19:35:04	U SMPL

LB88407 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 28 Jun 2017 15:45:51

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
I3930-09LX5 - 1		1268	0.2119		-7471A	PPB	28 Jun 2017 19:37:21	U	SMPL
I3930-09A - 1		24762	5.0365		-7471A	PPB	28 Jun 2017 19:39:38	U	SMPL
I3930-04DLX10 - 1		40857	8.3417		-7471A	PPB	28 Jun 2017 19:41:55	U	SMPL
I3930-05DLX10 - 1		18840	3.8204		-7471A	PPB	28 Jun 2017 19:44:12	U	SMPL
CCV85 - 1	CCV85	23256	4.7273		-7471A	PPB	28 Jun 2017 19:46:30	U	SMPL
CCB85 - 1	CCB85	226	-0.0020		-7471A	PPB	28 Jun 2017 19:48:47	U	SMPL
I3951-01DLX5 - 1		15737	3.1832		-7471A	PPB	28 Jun 2017 20:02:13	U	SMPL
CCV86 - 1	CCV86	24577	4.9986		-7471A	PPB	28 Jun 2017 20:04:30	U	SMPL
CCB86 - 1	CCB86	261	0.0051		-7471A	PPB	28 Jun 2017 20:06:48	U	SMPL

ns
7/10/17