



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88409

Temp.: 22.2 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

True Value of CCV4 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/28/2017	13:10
2	CAL2	1	Water			7.01	06/28/2017	13:11
3	CAL3	1	Water			10.03	06/28/2017	13:12
4	ICV	1	Water			7.01	06/28/2017	13:15
5	CCV1	1	Water			2.01	06/28/2017	13:16
6	I3950-20	1	Solid	20.03	20	9.41	06/28/2017	13:17
7	I3950-20DUP	1	Solid	20.04	20	9.42	06/28/2017	13:19
8	I3950-21	1	Solid	20.05	20	7.80	06/28/2017	13:20
9	I3950-22	1	Solid	20.02	20	7.78	06/28/2017	13:21
10	I3950-23	1	Solid	20.04	20	8.02	06/28/2017	13:22
11	I3950-24	1	Solid	20.02	20	7.52	06/28/2017	13:24
12	I3950-25	1	Solid	20.01	20	6.70	06/28/2017	13:25
13	I3950-26	1	Solid	20.02	20	9.75	06/28/2017	13:26
14	I3950-27	1	Solid	20.05	20	9.25	06/28/2017	13:27
15	I3950-28	1	Solid	20.04	20	8.60	06/28/2017	13:29
16	CCV2	1	Water			12.02	06/28/2017	13:30
17	I3950-29	1	Solid	20.02	20	8.63	06/28/2017	13:31
18	I3950-30	1	Solid	20.04	20	7.85	06/28/2017	13:32
19	I3950-31	1	Solid	20.01	20	8.20	06/28/2017	13:33
20	I3950-32	1	Solid	20.03	20	10.32	06/28/2017	13:35

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I3950-33	1	Solid	20.04	20	9.73	06/28/2017	13:36
22	I3950-34	1	Solid	20.02	20	9.01	06/28/2017	13:37
23	I3950-35	1	Solid	20.05	20	8.60	06/28/2017	13:39
24	I3950-36	1	Solid	20.04	20	7.90	06/28/2017	13:40
25	I3950-37	1	Solid	20.02	20	7.58	06/28/2017	13:42
26	I3950-38	1	Solid	20.03	20	6.86	06/28/2017	13:44
27	CCV3	1	Water			12.02	06/28/2017	13:45
28	I3950-38DUP	1	Solid	20.03	20	6.87	06/28/2017	13:46
29	CCV4	1	Water			12.01	06/28/2017	13:47

LB 88409

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3950

WorkList ID : 100181

Date : 6/28/2017 12:49:56 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3950-20	Corrosivity	Cool 4 deg C	TECT01	N63	B4-0-2	06/27/2017	9045C
	Solid	I3950-21	Corrosivity	Cool 4 deg C	TECT01	N63	B4-2-4	06/27/2017	9045C
	Solid	I3950-22	Corrosivity	Cool 4 deg C	TECT01	N63	B4-4-6	06/27/2017	9045C
	Solid	I3950-23	Corrosivity	Cool 4 deg C	TECT01	N63	B4-6-8	06/27/2017	9045C
	Solid	I3950-24	Corrosivity	Cool 4 deg C	TECT01	N63	B4-8-10	06/27/2017	9045C
	Solid	I3950-25	Corrosivity	Cool 4 deg C	TECT01	N63	B4-10-12	06/27/2017	9045C
	Solid	I3950-26	Corrosivity	Cool 4 deg C	TECT01	N63	B5-0-2	06/27/2017	9045C
	Solid	I3950-27	Corrosivity	Cool 4 deg C	TECT01	N63	B5-2-4	06/27/2017	9045C
	Solid	I3950-28	Corrosivity	Cool 4 deg C	TECT01	N63	B5-4-6	06/27/2017	9045C
	Solid	I3950-29	Corrosivity	Cool 4 deg C	TECT01	N63	B5-6-8	06/27/2017	9045C
	Solid	I3950-30	Corrosivity	Cool 4 deg C	TECT01	N63	B5-8-10.5	06/27/2017	9045C
	Solid	I3950-31	Corrosivity	Cool 4 deg C	TECT01	N63	B5-10.5-13	06/27/2017	9045C
	Solid	I3950-32	Corrosivity	Cool 4 deg C	TECT01	N63	B6-0-2	06/27/2017	9045C
	Solid	I3950-33	Corrosivity	Cool 4 deg C	TECT01	N63	B6-2-4	06/27/2017	9045C
	Solid	I3950-34	Corrosivity	Cool 4 deg C	TECT01	N63	B6-4-6	06/27/2017	9045C
	Solid	I3950-35	Corrosivity	Cool 4 deg C	TECT01	N63	B6-6-8	06/27/2017	9045C
	Solid	I3950-36	Corrosivity	Cool 4 deg C	TECT01	N63	B6-8-10	06/27/2017	9045C
	Solid	I3950-37	Corrosivity	Cool 4 deg C	TECT01	N63	B6-10-12	06/27/2017	9045C
	Solid	I3950-38	Corrosivity	Cool 4 deg C	TECT01	N63	B6-12-14	06/27/2017	9045C

Date/Time 6/28/17 12:50 PM

Received by: JO

Relinquished by:

Date/Time 6/28/17 6:30 PM

Received by:

Relinquished by: SO