



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88416

Temp.: 22.5 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/28/2017	16:00
2	CAL2	1	Water			7.00	06/28/2017	16:01
3	CAL3	1	Water			10.03	06/28/2017	16:02
4	ICV	1	Water			7.01	06/28/2017	16:03
5	CCV1	1	Water			2.01	06/28/2017	16:05
6	I3956-05	1	Solid	20.03	20	8.10	06/28/2017	16:06
7	I3956-05DUP	1	Solid	20.04	20	8.11	06/28/2017	16:07
8	I3959-03	1	Solid	20.03	20	6.72	06/28/2017	16:08
9	I3959-06	1	Solid	20.04	20	6.20	06/28/2017	16:10
10	I3959-09	1	Solid	20.05	20	6.28	06/28/2017	16:11
11	I3959-12	1	Solid	20.02	20	10.97	06/28/2017	16:12
12	I3974-03	1	Solid	20.01	20	6.75	06/28/2017	16:13
13	I3977-03	1	Solid	20.03	20	8.75	06/28/2017	16:15
14	CCV2	1	Water			12.02	06/28/2017	16:17

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3956

WorkList ID : 100178

Date : 6/28/2017 11:39:58 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/30/2017	Solid	I3956-05	Corrosivity	Cool 4 deg C	PSEG01	P51	MH-2093-COMP	06/27/2017	9045C
06/30/2017	Solid	I3959-03	Corrosivity	Cool 4 deg C	PSEG05	P61	BU-03-062717-A	06/27/2017	9045C
06/30/2017	Solid	I3959-06	Corrosivity	Cool 4 deg C	PSEG05	P61	BU-03-062717-B	06/27/2017	9045C
06/30/2017	Solid	I3959-09	Corrosivity	Cool 4 deg C	PSEG05	P61	BU-03-062717-C	06/27/2017	9045C
	Solid	I3959-12	Corrosivity	Cool 4 deg C	PSEG05	P61	BU-03-062717-D	06/27/2017	9045C
	Solid	I3974-03	Corrosivity	Cool 4 deg C	PSEG05	N32	OK-04-062817	06/28/2017	9045C
	Solid	I3977-03	Corrosivity	Cool 4 deg C	PSEG05	N31	SU-03-062817	06/28/2017	9045C

Date/Time 06/28/17 11:39 AM
 Received by: JO
 Relinquished by: [Signature]

Date/Time 06/28/17 7:00 PM
 Received by: CP
 Relinquished by: JO