



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88418

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/28/2017	18:10
2	CAL2	1	Water			7.01	06/28/2017	18:11
3	CAL3	1	Water			10.03	06/28/2017	18:12
4	ICV	1	Water			7.00	06/28/2017	18:15
5	CCV1	1	Water			2.01	06/28/2017	18:19
6	I3968-01	1	Solid	20.03	20	5.69	06/28/2017	18:20
7	I3968-01DUP	1	Solid	20.05	20	5.70	06/28/2017	18:21
8	CCV2	1	Water			12.01	06/28/2017	18:22

WORKLIST(Hardcopy Internal Chain)

2088418

WorkList Name : PH 13968

WorkList ID : 100192

Date : 6/28/2017 5:51:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I3968-01-X	pH	Cool 4 deg C	TULL02	N42	FK-PS-01-009	06/28/2017	9045C

Date/Time 06/28/17 6:00 PM
 Received by: JP
 Relinquished by: JP

Date/Time 06/28/17 7:20 PM
 Received by: JP
 Relinquished by: JP