



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88444

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	06/29/2017	13:30
2	CAL2	1	Water			7.01	06/29/2017	13:31
3	CAL3	1	Water			10.03	06/29/2017	13:32
4	ICV	1	Water			7.00	06/29/2017	13:33
5	CCV1	1	Water			2.01	06/29/2017	13:35
6	I3979-05	1	Solid	20.03	20	8.47	06/29/2017	13:36
7	I3979-05DUP	1	Solid	20.04	20	8.48	06/29/2017	13:37
8	I3979-10	1	Solid	20.02	20	7.93	06/29/2017	13:39
9	I3980-03	1	Solid	20.01	20	8.73	06/29/2017	13:40
10	I3980-06	1	Solid	20.02	20	8.24	06/29/2017	13:41
11	I3980-09	1	Solid	20.03	20	8.61	06/29/2017	13:42
12	I3980-12	1	Solid	20.05	20	8.73	06/29/2017	13:45
13	I3982-03	1	Solid	20.01	20	8.43	06/29/2017	13:46
14	I3982-08	1	Solid	20.04	20	7.90	06/29/2017	13:47
15	I3982-12	1	Solid	20.02	20	8.60	06/29/2017	13:48
16	CCV2	1	Water			12.02	06/29/2017	13:50
17	I3982-17	1	Solid	20.03	20	9.05	06/29/2017	13:51
18	I3983-03	1	Solid	20.04	20	8.52	06/29/2017	13:52
19	I3991-03	1	Solid	20.02	20	7.94	06/29/2017	13:53
20	I3991-06	1	Solid	20.05	20	8.77	06/29/2017	13:54

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
21	I3991-09	1	Solid	20.03	20	9.40	06/29/2017	14:00
22	I3996-02	1	Solid	20.03	20	9.22	06/29/2017	14:01
23	I3996-02DUP	1	Solid	20.05	20	9.23	06/29/2017	14:02
24	CCV3	1	Water			12.02	06/29/2017	14:05

28388444

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I3980 WorkList ID : 100252 Date : 6/29/2017 12:32:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/03/2017	Solid	I3979-05	Corrosivity	Cool 4 deg C	PSEG01	P61	TP3-MH2095-COMP	06/28/2017	9045C
07/03/2017	Solid	I3979-10	Corrosivity	Cool 4 deg C	PSEG01	P61	TP4-MH2096-COMP	06/28/2017	9045C
07/03/2017	Solid	I3980-03	Corrosivity	Cool 4 deg C	PSEG01	P62	SP-A-COMP	06/28/2017	9045C
07/03/2017	Solid	I3980-06	Corrosivity	Cool 4 deg C	PSEG01	P62	SP-B-COMP	06/28/2017	9045C
07/03/2017	Solid	I3980-09	Corrosivity	Cool 4 deg C	PSEG01	P62	SP-C-COMP	06/28/2017	9045C
07/03/2017	Solid	I3980-12	Corrosivity	Cool 4 deg C	PSEG01	P62	SP-D-COMP	06/28/2017	9045C
07/03/2017	Solid	I3982-03	Corrosivity	Cool 4 deg C	PSEG01	P62	339-COMP	06/28/2017	9045C
07/03/2017	Solid	I3982-08	Corrosivity	Cool 4 deg C	PSEG01	P62	348-343-COMP	06/28/2017	9045C
07/03/2017	Solid	I3982-12	Corrosivity	Cool 4 deg C	PSEG01	P62	277-COMP	06/28/2017	9045C
07/03/2017	Solid	I3982-17	Corrosivity	Cool 4 deg C	PSEG01	P62	163-72-11929-COMP	06/28/2017	9045C
07/05/2017	Solid	I3983-03	Corrosivity	Cool 4 deg C	PSEG01	P62	OILY-DRUM-COMP	06/28/2017	9045C
07/04/2017	Solid	I3991-03	Corrosivity	Cool 4 deg C	PSEG05	P63	HR-06-062917-A	06/29/2017	9045C
07/04/2017	Solid	I3991-06	Corrosivity	Cool 4 deg C	PSEG05	P63	HR-06-062917-B	06/29/2017	9045C
07/04/2017	Solid	I3991-09	Corrosivity	Cool 4 deg C	PSEG05	P63	HR-06-062917-C	06/29/2017	9045C
07/04/2017	Solid	I3996-02	Corrosivity	Cool 4 deg C	PSEG01	Q12	I-5-(A-H)-COMP	06/29/2017	9045C

Date/Time 06/29/17 12:32 PM
 Received by: JP
 Relinquished by: CP

Date/Time 06/29/17- 7:00 PM
 Received by: CP
 Relinquished by: JP