


7470A

LB88452
INSTRUMENT ID: CX2

Linear

 μ Abs.:

42209



A= 0.0000e+000

B= 2.4080e-004

C= -5.9996e-002

Rho= 0.9997472

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	287	0.000	287				
0.2	0.200	0.196	-0.004	1062	0.0 %	1062				
2.5	2.500	2.595	0.095	11024	0.0 %	11024				
5.0	5.000	4.916	-0.084	20666	0.0 %	20666				
7.5	7.500	7.380	-0.120	30896	0.0 %	30896				
10.0	10.000	10.104	0.104	42209	0.0 %	42209				

LB88452 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Jun 2017 09:24:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	287	-	0.00007470A	PPB		30 Jun 2017 09:33:30	S	Std
0.2 - 1	50.2	1062	-	0.20007470A	PPB		30 Jun 2017 09:35:38	S	Std
2.5 - 1	52.5	11024	-	2.50007470A	PPB		30 Jun 2017 09:37:47	S	Std
5.0 - 1	55	20666	-	5.00007470A	PPB		30 Jun 2017 09:39:55	S	Std
7.5 - 1	52.5	30896	-	7.50007470A	PPB		30 Jun 2017 09:42:03	S	Std
10.0 - 1	50	42209	-	10.00007470A	PPB		30 Jun 2017 09:44:12	S	Std
ICV45 - 1	ICV45	16912	4.0125	-7470A	PPB		30 Jun 2017 09:46:49	U	SMPL
ICB45 - 1	ICB45	276	0.0065	-7470A	PPB		30 Jun 2017 09:48:55	U	SMPL
CCV08 - 1	CCV08	20214	4.8076	-7470A	PPB		30 Jun 2017 09:51:01	U	SMPL
CCB08 - 1	CCB08	240	-0.0022	-7470A	PPB		30 Jun 2017 09:53:08	U	SMPL
CRA - 1	CRA	981	0.1762	-7470A	PPB		30 Jun 2017 09:55:16	U	SMPL
HighStd - 1	HighStd	41428	9.9161	-7470A	PPB		30 Jun 2017 09:57:22	U	SMPL
ChkStd - 1	ChkStd	28236	6.7394	-7470A	PPB		30 Jun 2017 09:59:29	U	SMPL
PB100301BL - 1	PBW	175	-0.0179	-7470A	PPB		30 Jun 2017 10:04:06	U	SMPL
PB100301BS - 1	LCSW	15548	3.6840	-7470A	PPB		30 Jun 2017 10:06:12	U	SMPL
PB100301TB - 1	PB100236TB	14	-0.0566	-7470A	PPB		30 Jun 2017 10:08:20	U	SMPL
I3956-05 - 1	MH-2093-COMP	-40	-0.0696	-7470A	PPB		30 Jun 2017 10:10:28	U	SMPL
I3956-05DUP - 1	MH-2093-COMPDUP	-82	-0.0797	-7470A	PPB		30 Jun 2017 10:12:37	U	SMPL
I3956-05MS - 1	MH-2093-COMPMS	15709	3.7228	-7470A	PPB		30 Jun 2017 10:14:45	U	SMPL
I3956-05MSD - 1	MH-2093-COMPMSD	16108	3.8189	-7470A	PPB		30 Jun 2017 10:16:52	U	SMPL
CCV09 - 1	CCV09	20206	4.8057	-7470A	PPB		30 Jun 2017 10:18:59	U	SMPL
CCB09 - 1	CCB09	196	-0.0128	-7470A	PPB		30 Jun 2017 10:21:07	U	SMPL
I3959-03 - 1	BU-03-062717-A	-18	-0.0643	-7470A	PPB		30 Jun 2017 10:23:14	U	SMPL
I3959-06 - 1	BU-03-062717-B	-25	-0.0660	-7470A	PPB		30 Jun 2017 10:25:21	U	SMPL
I3959-09 - 1	BU-03-062717-C	-18	-0.0643	-7470A	PPB		30 Jun 2017 10:27:27	U	SMPL
I3959-12 - 1	BU-03-062717-D	-9	-0.0622	-7470A	PPB		30 Jun 2017 10:29:34	U	SMPL
I3974-03 - 1	OK-04-062817	-17	-0.0641	-7470A	PPB		30 Jun 2017 10:31:42	U	SMPL
I3977-03 - 1	SU-03-062817	-18	-0.0643	-7470A	PPB		30 Jun 2017 10:33:49	U	SMPL
PB100302BL - 1	PBW	-28	-0.0667	-7470A	PPB		30 Jun 2017 10:35:57	U	SMPL
PB100302BS - 1	LCSW	15196	3.5993	-7470A	PPB		30 Jun 2017 10:38:06	U	SMPL
PB100302TB - 1	PB100252TB	-51	-0.0723	-7470A	PPB		30 Jun 2017 10:40:15	U	SMPL
I3950-20 - 1	B4-0-2	-17	-0.0641	-7470A	PPB		30 Jun 2017 10:42:22	U	SMPL
CCV10 - 1	CCV10	20521	4.8816	-7470A	PPB		30 Jun 2017 10:44:29	U	SMPL
CCB10 - 1	CCB10	214	-0.0085	-7470A	PPB		30 Jun 2017 10:46:35	U	SMPL
I3950-20DUP - 1	B4-0-2DUP	-20	-0.0648	-7470A	PPB		30 Jun 2017 10:48:43	U	SMPL
I3950-20MS - 1	B4-0-2MS	15370	3.6412	-7470A	PPB		30 Jun 2017 10:50:50	U	SMPL
I3950-20MSD - 1	B4-0-2MSD	15435	3.6568	-7470A	PPB		30 Jun 2017 10:52:56	U	SMPL
I3950-21 - 1	B4-2-4	-49	-0.0718	-7470A	PPB		30 Jun 2017 10:55:03	U	SMPL
I3950-22 - 1	B4-4-6	20	-0.0552	-7470A	PPB		30 Jun 2017 10:57:11	U	SMPL
I3950-23 - 1	B4-6-8	32	-0.0523	-7470A	PPB		30 Jun 2017 10:59:18	U	SMPL
I3950-24 - 1	B4-8-10	35	-0.0516	-7470A	PPB		30 Jun 2017 11:01:26	U	SMPL
I3950-25 - 1	B4-10-12	56	-0.0465	-7470A	PPB		30 Jun 2017 11:03:34	U	SMPL
I3950-26 - 1	B5-0-2	44	-0.0494	-7470A	PPB		30 Jun 2017 11:05:43	U	SMPL
I3950-27 - 1	B5-2-4	35	-0.0516	-7470A	PPB		30 Jun 2017 11:07:50	U	SMPL
CCV11 - 1	CCV11	20704	4.9256	-7470A	PPB		30 Jun 2017 11:09:57	U	SMPL
CCB11 - 1	CCB11	145	-0.0251	-7470A	PPB		30 Jun 2017 11:12:04	U	SMPL
I3950-28 - 1	B5-4-6	24	-0.0542	-7470A	PPB		30 Jun 2017 11:14:12	U	SMPL
I3950-29 - 1	B5-6-8	20	-0.0552	-7470A	PPB		30 Jun 2017 11:16:19	U	SMPL
I3950-30 - 1	B5-8-10.5	43	-0.0496	-7470A	PPB		30 Jun 2017 11:18:27	U	SMPL
I3950-31 - 1	B5-10.5-13	2	-0.0595	-7470A	PPB		30 Jun 2017 11:20:34	U	SMPL
I3950-32 - 1	B6-0-2	-4	-0.0610	-7470A	PPB		30 Jun 2017 11:22:41	U	SMPL
I3950-33 - 1	B6-2-4	-6	-0.0614	-7470A	PPB		30 Jun 2017 11:24:49	U	SMPL
I3950-34 - 1	B6-4-6	-10	-0.0624	-7470A	PPB		30 Jun 2017 11:26:57	U	SMPL
I3950-35 - 1	B6-6-8	36	-0.0513	-7470A	PPB		30 Jun 2017 11:29:05	U	SMPL
I3950-36 - 1	B6-8-10	51	-0.0477	-7470A	PPB		30 Jun 2017 11:31:14	U	SMPL
I3950-37 - 1	B6-10-12	249	-0.0000	-7470A	PPB		30 Jun 2017 11:33:21	U	SMPL
CCV12 - 1	CCV12	20982	4.9926	-7470A	PPB		30 Jun 2017 11:35:28	U	SMPL
CCB12 - 1	CCB12	218	-0.0075	-7470A	PPB		30 Jun 2017 11:37:36	U	SMPL
I3950-38 - 1	B6-12-14	62	-0.0451	-7470A	PPB		30 Jun 2017 11:39:44	U	SMPL
I3956-05LX5 - 1		-20	-0.0648	-7470A	PPB		30 Jun 2017 11:41:51	U	SMPL
I3956-05A - 1		16758	3.9754	-7470A	PPB		30 Jun 2017 11:43:58	U	SMPL
I3950-20LX5 - 1		-69	-0.0766	-7470A	PPB		30 Jun 2017 11:46:05	U	SMPL
I3950-20A - 1		16275	3.8591	-7470A	PPB		30 Jun 2017 11:48:13	U	SMPL
CCV13 - 1	CCV13	20680	4.9198	-7470A	PPB		30 Jun 2017 11:50:20	U	SMPL
CCB13 - 1	CCB13	138	-0.0268	-7470A	PPB		30 Jun 2017 11:52:28	U	SMPL