



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88470

Temp.: 22.3 °C

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/30/2017	13:35
2	CAL2	1	Water			7.00	06/30/2017	13:36
3	CAL3	1	Water			10.02	06/30/2017	13:37
4	ICV	1	Water			7.01	06/30/2017	13:38
5	CCV1	1	Water			2.01	06/30/2017	13:40
6	I4000-01	1	Water			6.60	06/30/2017	13:41
7	I4000-01DUP	1	Water			6.61	06/30/2017	13:42
8	I4006-02	1	Water			6.62	06/30/2017	13:43
9	I4011-01	1	Water			5.72	06/30/2017	13:45
10	I4011-02	1	Water			5.43	06/30/2017	13:46
11	I4011-03	1	Water			5.40	06/30/2017	13:47
12	I4020-01	1	Water			6.71	06/30/2017	13:48
13	I4020-04	1	Water			6.83	06/30/2017	13:49
14	CCV2	1	Water			12.02	06/30/2017	13:50

LD 88470

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4000

WorkList ID : 100309

Date : 6/30/2017 1:20:48 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/29/2017	Water	I4000-01 E	pH	Cool 4 deg C	PSEG01	P32	A1273-FT	06/29/2017	SM 4500-PH B
	Water	I4006-02 E	pH	Cool 4 deg C	PARS15		CIV-WC-L-1	06/29/2017	SM 4500-PH B
	Water	I4011-01 A	pH	Cool 4 deg C	STER07	P13	CELL-A	06/27/2017	SM 4500-PH B
	Water	I4011-02 A	pH	Cool 4 deg C	STER07	P13	CELL-B-C	06/27/2017	SM 4500-PH B
	Water	I4011-03 A	pH	Cool 4 deg C	STER07	P13	CELL-D	06/27/2017	SM 4500-PH B
06/30/2017	Water	I4020-01 C	pH	Cool 4 deg C	PSEG01	D41	HDD-WATER	06/30/2017	SM 4500-PH B
06/30/2017	Water	I4020-04 C	pH	Cool 4 deg C	PSEG01	D41	HDD-SHEEN	06/30/2017	SM 4500-PH B

Date/Time 06/30/17 1:20 PM
 Received by: JS
 Relinquished by: JS

Date/Time 06/30/17 4:00 PM
 Received by: JS
 Relinquished by: JS