



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88474

Temp.: 23.7 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	06/30/2017	15:35
2	CAL2	1	Water			6.99	06/30/2017	15:36
3	CAL3	1	Water			10.03	06/30/2017	15:37
4	ICV	1	Water			7.00	06/30/2017	15:39
5	CCV1	1	Water			2.01	06/30/2017	15:40
6	I4007-05	1	Solid	20.03	20	7.84	06/30/2017	15:41
7	I4007-10	1	Solid	20.04	20	7.90	06/30/2017	15:42
8	I4025-05	1	Solid	20.03	20	6.75	06/30/2017	15:43
9	I4031-03	1	Solid	20.04	20	7.57	06/30/2017	15:44
10	I4031-03DUP	1	Solid	20.02	20	7.58	06/30/2017	15:45
11	CCV2	1	Water			12.01	06/30/2017	15:46

1588474

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4007

WorkList ID : 100304

Date : 6/30/2017 12:12:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/04/2017	Solid	I4007-05	Corrosivity	Cool 4 deg C	PSEG01	D31	TP2-MH2094-COMP	06/29/2017	9045C
07/04/2017	Solid	I4007-10	Corrosivity	Cool 4 deg C	PSEG01	D31	TP5-MH2097-COMP	06/29/2017	9045C
07/05/2017	Solid	I4025-05	Corrosivity	Cool 4 deg C	PSEG01	D41	TP6-MH2098-COMP	06/30/2017	9045C
07/05/2017	Solid	I4031-03	Corrosivity	Cool 4 deg C	PSEG05	D32	OR-02-063017	06/30/2017	9045C

Date/Time 06/30/17 12:12 PM
 Received by: JP
 Relinquished by: JP

Date/Time 06/30/17 7:00 PM
 Received by: CJ
 Relinquished by: JP