

7470A


LB88476

INSTRUMENT ID: CV2

Linear

 μ Abs.:

45960



A= 0.0000e+000

B= 2.2215e-004

C= 3.7022e-002

Rho= 0.9984980

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.029	0.029	-38	0.000	-38				
0.2	0.200	0.219	0.019	818	0.0 %	818				
2.5	2.500	2.722	0.222	12085	0.0 %	12085				
5.0	5.000	4.761	-0.239	21266	0.0 %	21266				
7.5	7.500	7.223	-0.277	32346	0.0 %	32346				
10.0	10.000	10.247	0.247	45960	0.0 %	45960				

LB88476 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Jun 2017 19:40:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-38	-	0.00007470A	PPB		30 Jun 2017 19:44:53	S	Std
0.2 - 1	50.2	818	-	0.20007470A	PPB		30 Jun 2017 19:47:02	S	Std
2.5 - 1	52.5	12085	-	2.50007470A	PPB		30 Jun 2017 19:49:10	S	Std
5.0 - 1	55	21266	-	5.00007470A	PPB		30 Jun 2017 19:51:18	S	Std
7.5 - 1	57.5	32346	-	7.50007470A	PPB		30 Jun 2017 19:55:54	S	Std
10.0 - 1	510	45960	-	10.00007470A	PPB		30 Jun 2017 20:05:37	S	Std
ICV47 - 1	ICV47	17923	4.0186	-7470A	PPB		30 Jun 2017 20:10:50	U	SMPL
ICB47 - 1	ICB47	-100	0.0148	-7470A	PPB		30 Jun 2017 20:12:56	U	SMPL
CCV26 - 1	CCV26	20735	4.6433	-7470A	PPB		30 Jun 2017 20:15:02	U	SMPL
CCB26 - 1	CCB26	-104	0.0139	-7470A	PPB		30 Jun 2017 20:17:09	U	SMPL
CRA - 1	CRA	804	0.2156	-7470A	PPB		30 Jun 2017 20:19:16	U	SMPL
HighStd - 1	HighStd	45538	10.1533	-7470A	PPB		30 Jun 2017 20:21:23	U	SMPL
ChkStd - 1	ChkStd	28289	6.3214	-7470A	PPB		30 Jun 2017 20:23:29	U	SMPL
PB100348BL - 1	PBW	-152	0.0033	-7470A	PPB		30 Jun 2017 20:25:38	U	SMPL
PB100348BS - 1	LCSW	15445	3.4681	-7470A	PPB		30 Jun 2017 20:27:47	U	SMPL
I3955-01 - 1	A1022C	1694	0.4133	-7470A	PPB		30 Jun 2017 20:29:54	U	SMPL
I3955-01DUP - 1	A1022CDUP	1646	0.4027	-7470A	PPB		30 Jun 2017 20:32:02	U	SMPL
I3955-01MS - 1	A1022CMS	16435	3.6881	-7470A	PPB		30 Jun 2017 20:34:10	U	SMPL
I3955-01MSD - 1	A1022CMSD	16753	3.7587	-7470A	PPB		30 Jun 2017 20:36:19	U	SMPL
I3962-19 - 1	FL-EB-2017-1	-154	0.0028	-7470A	PPB		30 Jun 2017 20:38:26	U	SMPL
CCV27 - 1	CCV27	20633	4.6206	-7470A	PPB		30 Jun 2017 20:40:34	U	SMPL
CCB27 - 1	CCB27	-112	0.0121	-7470A	PPB		30 Jun 2017 20:42:41	U	SMPL
I4000-01 - 1	A1273-FT	4116	0.9514	-7470A	PPB		30 Jun 2017 20:44:48	U	SMPL
I4004-06 - 1	CIY-FB062917	-91	0.0168	-7470A	PPB		30 Jun 2017 20:46:54	U	SMPL
I4005-01 - 1	CIY-GW-01	369359	HIGH	-7470A	PPB		30 Jun 2017 20:49:01	X	
I4006-01 - 1	CIY-WC-L-1	246834	54.8712	-7470A	PPB		30 Jun 2017 20:50:53	U	SMPL
I4018-01 - 1	GROUNDWATER	3800	0.8812	-7470A	PPB		30 Jun 2017 20:53:03	U	SMPL
I4020-01 - 1	HDD-WATER	-105	0.0137	-7470A	PPB		30 Jun 2017 20:55:20	U	SMPL
I4020-04 - 1	HDD-SHEEN	-73	0.0208	-7470A	PPB		30 Jun 2017 20:57:28	U	SMPL
PB100349BL - 1	PBW	-55	0.0248	-7470A	PPB		30 Jun 2017 20:59:36	U	SMPL
PB100349BS - 1	LCSW	15509	3.4823	-7470A	PPB		30 Jun 2017 21:01:45	U	SMPL
PB100349TB - 1	PB100295TB	-87	0.0177	-7470A	PPB		30 Jun 2017 21:03:52	U	SMPL
CCV28 - 1	CCV28	21511	4.8157	-7470A	PPB		30 Jun 2017 21:05:59	U	SMPL
CCB28 - 1	CCB28	-105	0.0137	-7470A	PPB		30 Jun 2017 21:08:06	U	SMPL
I3979-05 - 1	TP3-MH2095-COMP	-48	0.0264	-7470A	PPB		30 Jun 2017 21:10:14	U	SMPL
I3979-05DUP - 1	TP3-MH2095-COMPDUP	-40	0.0281	-7470A	PPB		30 Jun 2017 21:12:21	U	SMPL
I3979-05MS - 1	TP3-MH2095-COMPMS	15452	3.4697	-7470A	PPB		30 Jun 2017 21:14:27	U	SMPL
I3979-05MSD - 1	TP3-MH2095-COMPMSD	14627	3.2864	-7470A	PPB		30 Jun 2017 21:16:34	U	SMPL
I3979-10 - 1	TP4-MH2096-COMP	-29	0.0306	-7470A	PPB		30 Jun 2017 21:18:41	U	SMPL
I3980-03 - 1	SP-A-COMP	-24	0.0317	-7470A	PPB		30 Jun 2017 21:20:49	U	SMPL
I3980-06 - 1	SP-B-COMP	-36	0.0290	-7470A	PPB		30 Jun 2017 21:22:57	U	SMPL
I3980-09 - 1	SP-C-COMP	-50	0.0259	-7470A	PPB		30 Jun 2017 21:25:05	U	SMPL
I3980-12 - 1	SP-D-COMP	-44	0.0272	-7470A	PPB		30 Jun 2017 21:27:14	U	SMPL
I3982-03 - 1	339-COMP	-32	0.0299	-7470A	PPB		30 Jun 2017 21:29:21	U	SMPL
CCV29 - 1	CCV29	21290	4.7666	-7470A	PPB		30 Jun 2017 21:31:28	U	SMPL
CCB29 - 1	CCB29	-106	0.0135	-7470A	PPB		30 Jun 2017 21:33:36	U	SMPL
I3982-08 - 1	348-343-COMP	-40	0.0281	-7470A	PPB		30 Jun 2017 21:35:44	U	SMPL
I3982-12 - 1	277-COMP	0	0.0370	-7470A	PPB		30 Jun 2017 21:37:51	U	SMPL
I3982-17 - 1	163-72-11929-COMP	-22	0.0321	-7470A	PPB		30 Jun 2017 21:39:58	U	SMPL
I3983-03 - 1	OILY-DRUM-COMP	-9	0.0350	-7470A	PPB		30 Jun 2017 21:42:05	U	SMPL
I3986-01 - 1	TP-10	-25	0.0315	-7470A	PPB		30 Jun 2017 21:44:13	U	SMPL
I3986-02 - 1	TP-11	-11	0.0346	-7470A	PPB		30 Jun 2017 21:46:21	U	SMPL
I3986-03 - 1	TP-12	-31	0.0301	-7470A	PPB		30 Jun 2017 21:48:29	U	SMPL
I3991-03 - 1	HR-06-062917-A	-21	0.0324	-7470A	PPB		30 Jun 2017 21:50:37	U	SMPL
I3991-06 - 1	HR-06-062917-B	-48	0.0264	-7470A	PPB		30 Jun 2017 21:52:45	U	SMPL
I3991-09 - 1	HR-06-062917-C	-45	0.0270	-7470A	PPB		30 Jun 2017 21:54:53	U	SMPL
CCV30 - 1	CCV30	20978	4.6973	-7470A	PPB		30 Jun 2017 21:57:00	U	SMPL
CCB30 - 1	CCB30	-101	0.0146	-7470A	PPB		30 Jun 2017 21:59:07	U	SMPL
I3996-02 - 1	I-5-(A-H)-COMP	-23	0.0319	-7470A	PPB		30 Jun 2017 22:01:15	U	SMPL
I4005-01DLX10 - 1		49919	11.1265	-7470A	PPB		30 Jun 2017 22:03:22	U	SMPL
I4006-01DLX10 - 1		24377	5.4524	-7470A	PPB		30 Jun 2017 22:05:29	U	SMPL
I3955-01LX5 - 1		209	0.0835	-7470A	PPB		30 Jun 2017 22:07:38	U	SMPL
I3955-01A - 1		17467	3.9173	-7470A	PPB		30 Jun 2017 22:09:47	U	SMPL
I3979-05LX5 - 1		-94	0.0161	-7470A	PPB		30 Jun 2017 22:11:54	U	SMPL
I3979-05A - 1		16817	3.7729	-7470A	PPB		30 Jun 2017 22:14:03	U	SMPL
I4005-01DL2X50 - 1		9704	2.1928	-7470A	PPB		30 Jun 2017 22:16:11	U	SMPL
CCV31 - 1	CCV31	21589	4.8330	-7470A	PPB		30 Jun 2017 22:18:20	U	SMPL
CCB31 - 1	CCB31	-128	0.0086	-7470A	PPB		30 Jun 2017 22:20:27	U	SMPL