



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB88491

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	07/01/2017	09:02
2	I3955-02		1	>100.0	>212.0	07/01/2017	09:35
3	I3955-02DUP		1	>100.0	>212.0	07/01/2017	10:02
4	I4000-02		1	>100.0	>212.0	07/01/2017	10:40
5	I4006-02		1	>100.0	>212.0	07/01/2017	11:05
6	I4018-01		1	>100.0	>212.0	07/01/2017	11:45
7	I4020-02		1	>100.0	>212.0	07/01/2017	12:15
8	I4020-05		1	>100.0	>212.0	07/01/2017	12:40

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

LB 88491

WorkList Name : FP7117

WorkList ID : 100331

Date : 7/1/2017 08:30AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/30/2017	Water	I3955-02	Flash Point	Cool 4 deg C	PSEG01	P51	A1022C	06/27/2017	1010A
07/04/2017	Water	I4000-02	Flash Point	Cool 4 deg C	PSEG01	P32	A1273-FT	06/29/2017	1010A
	Water	I4006-02	Flash Point	Cool 4 deg C	PARS15		CIY-WC-L-1	06/29/2017	1010A
	Water	I4018-01	Flash Point	Cool 4 deg C	SEAC01	P32	GROUNDWATER	06/30/2017	1010A
07/05/2017	Water	I4020-02	Flash Point	Cool 4 deg C	PSEG01	D41	HDD-WATER	06/30/2017	1010A
07/05/2017	Water	I4020-05	Flash Point	Cool 4 deg C	PSEG01	D41	HDD-SHEEN	06/30/2017	1010A

Date/Time 7/1/17 8:40 AM
 Received by: KS
 Relinquished by: KS

Date/Time 7/1/17 12:45 PM
 Received by: KS
 Relinquished by: KS