


7470A

LB 88503
INSTRUMENT ID: CV1

Linear

 μ Abs.:

45300



A= 0.0000e+000

B= 2.2433e-004

C= -5.5114e-003

Rho= 0.9991371

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.019	-0.019	-61	0.000	-61				
0.2	0.200	0.168	-0.032	773	0.0 %	773				
2.5	2.500	2.688	0.188	12005	0.0 %	12005				
5.0	5.000	4.989	-0.011	22265	0.0 %	22265				
7.5	7.500	7.218	-0.282	32199	0.0 %	32199				
10.0	10.000	10.157	0.157	45300	0.0 %	45300				

LB88503 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 03 Jul 2017 15:35:19					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-61	-	0.0000	7470A	PPB	03 Jul 2017 15:39:27	S	Std
0.2 - 1	50.2	773	-	0.2000	7470A	PPB	03 Jul 2017 15:41:35	S	Std
2.5 - 1	52.5	12005	-	2.5000	7470A	PPB	03 Jul 2017 15:48:59	S	Std
5.0 - 1	55	22265	-	5.0000	7470A	PPB	03 Jul 2017 15:51:07	S	Std
7.5 - 1	57.5	32199	-	7.5000	7470A	PPB	03 Jul 2017 15:53:15	S	Std
10.0 - 1	518	45300	-	10.0000	7470A	PPB	03 Jul 2017 15:55:23	S	Std
ICV48 - 1	ICV48	16923	3.7909	-	7470A	PPB	03 Jul 2017 16:02:50	U	SMPL
ICB48 - 1	ICB48	-141	-0.0371	-	7470A	PPB	03 Jul 2017 16:04:56	U	SMPL
CCV32 - 1	CCV32	21356	4.7853	-	7470A	PPB	03 Jul 2017 16:09:25	U	SMPL
CCB32 - 1	CCB32	-149	-0.0389	-	7470A	PPB	03 Jul 2017 16:11:31	U	SMPL
CRA - 1	CRA	827	0.1800	-	7470A	PPB	03 Jul 2017 16:13:39	U	SMPL
HighStd - 1	HighStd	44986	10.0863	-	7470A	PPB	03 Jul 2017 16:15:45	U	SMPL
ChkStd - 1	ChkStd	29049	6.5111	-	7470A	PPB	03 Jul 2017 16:17:52	U	SMPL
PB100379BL - 1	PBW	-173	-0.0443	-	7470A	PPB	03 Jul 2017 16:20:01	U	SMPL
PB100379BS - 1	LCSW	15389	3.4467	-	7470A	PPB	03 Jul 2017 16:22:09	U	SMPL
PB100379TB - 1	PB100346TB	-135	-0.0358	-	7470A	PPB	03 Jul 2017 16:24:17	U	SMPL
I3975-15 - 1	B7-0-2	-84	-0.0244	-	7470A	PPB	03 Jul 2017 16:26:25	U	SMPL
I3975-15DUP - 1	B7-0-2DUP	-72	-0.0217	-	7470A	PPB	03 Jul 2017 16:28:33	U	SMPL
I3975-15MS - 1	B7-0-2MS	14830	3.3213	-	7470A	PPB	03 Jul 2017 16:30:42	U	SMPL
I3975-15MSD - 1	B7-0-2MSD	16484	3.6924	-	7470A	PPB	03 Jul 2017 16:32:49	U	SMPL
CCV33 - 1	CCV33	21881	4.9031	-	7470A	PPB	03 Jul 2017 16:34:55	U	SMPL
CCB33 - 1	CCB33	-150	-0.0392	-	7470A	PPB	03 Jul 2017 16:37:03	U	SMPL
I3975-16 - 1	B7-2-4	-87	-0.0250	-	7470A	PPB	03 Jul 2017 16:39:10	U	SMPL
I3975-17 - 1	B7-4-6	-86	-0.0248	-	7470A	PPB	03 Jul 2017 16:41:17	U	SMPL
I3975-18 - 1	B7-6-8	-69	-0.0210	-	7470A	PPB	03 Jul 2017 16:43:24	U	SMPL
I3975-19 - 1	B7-8-10	-64	-0.0199	-	7470A	PPB	03 Jul 2017 16:45:31	U	SMPL
I3975-20 - 1	B7-10-12	-59	-0.0187	-	7470A	PPB	03 Jul 2017 16:47:38	U	SMPL
I3975-21 - 1	B7-12-14	-63	-0.0196	-	7470A	PPB	03 Jul 2017 16:49:46	U	SMPL
I3975-22 - 1	B8-0-2	-62	-0.0194	-	7470A	PPB	03 Jul 2017 16:51:54	U	SMPL
I3975-23 - 1	B8-2-4	-67	-0.0205	-	7470A	PPB	03 Jul 2017 16:54:03	U	SMPL
I3975-24 - 1	B8-4-6	-64	-0.0199	-	7470A	PPB	03 Jul 2017 16:56:11	U	SMPL
I3975-25 - 1	B8-6-8	-76	-0.0226	-	7470A	PPB	03 Jul 2017 16:58:18	U	SMPL
CCV34 - 1	CCV34	21811	4.8874	-	7470A	PPB	03 Jul 2017 17:00:25	U	SMPL
CCB34 - 1	CCB34	-140	-0.0369	-	7470A	PPB	03 Jul 2017 17:02:32	U	SMPL
I3975-26 - 1	B8-8-10	44	0.0044	-	7470A	PPB	03 Jul 2017 17:04:40	U	SMPL
I3975-27 - 1	B8-10-12	-59	-0.0187	-	7470A	PPB	03 Jul 2017 17:06:47	U	SMPL
I3975-28 - 1	B8-12-14	-69	-0.0210	-	7470A	PPB	03 Jul 2017 17:08:54	U	SMPL
I3976-09 - 1	B9-0-2	-70	-0.0212	-	7470A	PPB	03 Jul 2017 17:11:01	U	SMPL
I3976-10 - 1	B9-2-4	-61	-0.0192	-	7470A	PPB	03 Jul 2017 17:13:08	U	SMPL
I3976-11 - 1	B9-4-6	-56	-0.0181	-	7470A	PPB	03 Jul 2017 17:15:16	U	SMPL
I3976-12 - 1	B9-6-8	-36	-0.0136	-	7470A	PPB	03 Jul 2017 17:17:24	U	SMPL
I3976-13 - 1	B9-8-10	-73	-0.0219	-	7470A	PPB	03 Jul 2017 17:19:32	U	SMPL
I3976-14 - 1	B10-0-2	-64	-0.0199	-	7470A	PPB	03 Jul 2017 17:21:41	U	SMPL
I3976-15 - 1	B10-2-4	-54	-0.0176	-	7470A	PPB	03 Jul 2017 17:23:49	U	SMPL
CCV35 - 1	CCV35	21184	4.7467	-	7470A	PPB	03 Jul 2017 17:25:56	U	SMPL
CCB35 - 1	CCB35	-133	-0.0353	-	7470A	PPB	03 Jul 2017 17:28:03	U	SMPL
I3976-16 - 1	B10-4-6	-74	-0.0221	-	7470A	PPB	03 Jul 2017 17:30:10	U	SMPL
PB100380BL - 1	PBW	-61	-0.0192	-	7470A	PPB	03 Jul 2017 17:32:18	U	SMPL
PB100380BS - 1	LCSW	16119	3.6105	-	7470A	PPB	03 Jul 2017 17:34:25	U	SMPL
PB100380TB - 1	PB100351TB	-140	-0.0369	-	7470A	PPB	03 Jul 2017 17:36:32	U	SMPL
4000-02 - 1	A1273-FT	-81	-0.0237	-	7470A	PPB	03 Jul 2017 17:38:40	U	SMPL
4000-02DUP - 1	A1273-FTDUP	-97	-0.0273	-	7470A	PPB	03 Jul 2017 17:40:47	U	SMPL
4000-02MS - 1	A1273-FTMS	15026	3.3653	-	7470A	PPB	03 Jul 2017 17:42:55	U	SMPL
4000-02MSD - 1	A1273-FTMSD	14805	3.3157	-	7470A	PPB	03 Jul 2017 17:45:04	U	SMPL
4007-05 - 1	TP2-MH2094-COMP	-101	-0.0282	-	7470A	PPB	03 Jul 2017 17:47:13	U	SMPL
4007-10 - 1	TP5-MH2097-COMP	-71	-0.0214	-	7470A	PPB	03 Jul 2017 17:49:20	U	SMPL
CCV36 - 1	CCV36	21019	4.7097	-	7470A	PPB	03 Jul 2017 17:51:27	U	SMPL
CCB36 - 1	CCB36	-124	-0.0333	-	7470A	PPB	03 Jul 2017 17:53:34	U	SMPL
4017-01 - 1	PURGE-WATER-A	2781	0.6184	-	7470A	PPB	03 Jul 2017 17:55:42	U	SMPL
4020-02 - 1	HDD-WATER	-87	-0.0250	-	7470A	PPB	03 Jul 2017 17:57:49	U	SMPL
4020-05 - 1	HDD-SHEEN	-88	-0.0253	-	7470A	PPB	03 Jul 2017 17:59:56	U	SMPL
4025-05 - 1	TP6-MH2098-COMP	-57	-0.0183	-	7470A	PPB	03 Jul 2017 18:02:03	U	SMPL
4028-02 - 1	REACTOR-1-CA	-79	-0.0232	-	7470A	PPB	03 Jul 2017 18:04:11	U	SMPL
4028-04 - 1	REACTOR-1-CB	-57	-0.0183	-	7470A	PPB	03 Jul 2017 18:06:18	U	SMPL
4031-03 - 1	OR-02-063017	-64	-0.0199	-	7470A	PPB	03 Jul 2017 18:08:26	U	SMPL
CCV37 - 1	CCV37	21195	4.7492	-	7470A	PPB	03 Jul 2017 18:10:35	U	SMPL
CCB37 - 1	CCB37	-116	-0.0315	-	7470A	PPB	03 Jul 2017 18:12:43	U	SMPL
3975-15LX5 - 1		-65	-0.0201	-	7470A	PPB	03 Jul 2017 18:14:50	U	SMPL
3975-15A - 1		16071	3.5997	-	7470A	PPB	03 Jul 2017 18:16:57	U	SMPL
4000-02LX5 - 1		-125	-0.0336	-	7470A	PPB	03 Jul 2017 18:19:05	U	SMPL
4000-02A - 1		15966	3.5762	-	7470A	PPB	03 Jul 2017 18:21:12	U	SMPL
CCV38 - 1	CCV38	21167	4.7429	-	7470A	PPB	03 Jul 2017 18:23:19	U	SMPL
CCB38 - 1	CCB38	-134	-0.0356	-	7470A	PPB	03 Jul 2017 18:25:26	U	SMPL