



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88515

Temp.: 24.2 °C

BalanceID: WC SC-4

SLOPE.: 99.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/05/2017	10:00
2	CAL2	1	Water			7.01	07/05/2017	10:01
3	CAL3	1	Water			10.02	07/05/2017	10:02
4	ICV	1	Water			7.00	07/05/2017	10:03
5	CCV1	1	Water			2.02	07/05/2017	10:05
6	I4035-01	1	Solid	20.03	20	7.46	07/05/2017	10:06
7	I4035-01DUP	1	Solid	20.05	20	7.47	07/05/2017	10:07
8	I4049-03	1	Solid	20.02	20	6.61	07/05/2017	10:09
9	I4049-06	1	Solid	20.03	20	8.90	07/05/2017	10:10
10	I4052-03	1	Solid	20.03	20	8.01	07/05/2017	10:11
11	I4052-06	1	Solid	20.02	20	7.71	07/05/2017	10:12
12	I4052-09	1	Solid	20.01	20	8.76	07/05/2017	10:14
13	I4053-03	1	Solid	20.03	20	8.46	07/05/2017	10:15
14	I4053-06	1	Solid	20.04	20	5.60	07/05/2017	10:16
15	I4053-09	1	Solid	20.05	20	5.28	07/05/2017	10:17
16	CCV2	1	Water			12.01	07/05/2017	10:19
17	I4053-12	1	Solid	20.03	20	5.67	07/05/2017	10:20
18	I4053-15	1	Solid	20.05	20	8.00	07/05/2017	10:21
19	CCV3	1	Water			12.02	07/05/2017	10:22

LB88515

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4052 WorkList ID : 100384 Date : 7/5/2017 9:31:41 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4035-01	Corrosivity	Cool 4 deg C	PARS15	D41	CY-SB-WC-S-2-10-50	06/30/2017	9045C
07/06/2017	Solid	I4049-03	Corrosivity	Cool 4 deg C	PSEG01	E31	RO-72-12000	07/03/2017	9045C
07/06/2017	Solid	I4049-06	Corrosivity	Cool 4 deg C	PSEG01	E31	DRUM-1-070317	07/03/2017	9045C
07/06/2017	Solid	I4052-03	Corrosivity	Cool 4 deg C	PSEG05	E32	HR-0-070317-A	07/03/2017	9045C
07/06/2017	Solid	I4052-06	Corrosivity	Cool 4 deg C	PSEG05	E32	HR-0-070317-B	07/03/2017	9045C
07/06/2017	Solid	I4052-09	Corrosivity	Cool 4 deg C	PSEG05	E32	HR-0-070317-C	07/03/2017	9045C
07/06/2017	Solid	I4053-03	Corrosivity	Cool 4 deg C	PSEG05	E32	CL-02-070317-A	07/03/2017	9045C
07/06/2017	Solid	I4053-06	Corrosivity	Cool 4 deg C	PSEG05	E32	CL-02-070317-B	07/03/2017	9045C
07/06/2017	Solid	I4053-09	Corrosivity	Cool 4 deg C	PSEG05	E32	CL-02-070317-C	07/03/2017	9045C
07/06/2017	Solid	I4053-12	Corrosivity	Cool 4 deg C	PSEG05	E32	CL-02-070317-D	07/03/2017	9045C
07/06/2017	Solid	I4053-15	Corrosivity	Cool 4 deg C	PSEG05	E32	CL-02-070317-E	07/03/2017	9045C

Date/Time 07/05/17 9:31 AM

Received by: JP

Relinquished by: UN

Date/Time 07/05/17 12:30 PM

Received by: CL

Relinquished by: SP