



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB88517
Analysis Date: 07/05/2017
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: JIGNESH
REVIEWED BY: apatel
Extraction Date: 07/05/2017
Extraction IN Time: 12:45
Extraction OUT Time: 13:40

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB88517BL	LB88517BL	WATER	2	1000	100	2.7411	2.7411	3.02	2.7411	2.7411	0	0
2	LB88517BS	LB88517BS	WATER	2	1000	100	2.9833	2.9833	3.04	3.0024	3.0024	0.0191	19.1
3	LB88517BSD	LB88517BSD	WATER	2	1000	100	2.9128	2.9128	3.03	2.9321	2.9321	0.0193	19.3
4	I4000-01	A1273-FT	WATER	2	980	100	2.9152	2.9152	3.04	2.9206	2.9206	0.0054	5.51
5	I4020-01	HDD-WATER	WATER	2	1000	100	2.909	2.909	3.05	2.9094	2.9094	0.0004	0.4
6	I4020-04	HDD-SHEEN	WATER	2	800	100	3.0531	3.0531	3.03	3.0538	3.0538	0.0007	0.87



QC Batch# LB88517 Test: TPH Analysis Date: 07/05/2017

Chemicals Used:

Chemical Name	Chemical Lot #
Hexane	W2243
pH Paper 0-14	W2186
Sodium Sulfate	EP1754
1:1 HCL	WP54437
Silica Gel	W2199
Sand	

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP52823
LCSWD	10.00 ML	WP54439
MS/MSD		

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance: 0.0018 In OVEN TEMP : 70 Dessicator Time In : 15:00
1.0000 gram Balance: 1.0005 In Time: 14:20
Bal Check Time: 13:15

After Analysis

0.0020 gram Balance: 0.0021 Out OVEN TEMP: 71 Dessicator Time Out: 16:00
1.0000 gram Balance: 1.0003 Out Time: 15:25
Bal Check Time: 16:30

WORKLIST(Hardcopy Internal Chain)

288517

WorkList Name : O&G I4000

WorkList ID : 100388

Date : 7/5/2017 12:02:30 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/04/2017	Water	I4000-01-11	TPH	Conc H2SO4 tk	PSEG01	P32	A1273-FT	06/29/2017	1664A
07/05/2017	Water	I4020-01-12	TPH	Conc H2SO4 tk	PSEG01	D41	HDD-WATER	06/30/2017	1664A
07/05/2017	Water	I4020-04-12	TPH	Conc H2SO4 tk	PSEG01	D41	HDD-SHEEN	06/30/2017	1664A

Date/Time 07/05/17 12:01 PM
 Received by: JP
 Relinquished by: CL

Date/Time 07/05/17 3:30 PM
 Received by: CP
 Relinquished by: JP