

LB 88518


ISM02.4

INSTRUMENT ID: CV4

Linear

 μ Abs.:

48785



A= 0.0000e+000

B= 2.0463e-004

C= 1.1273e-002

Rho= 0.9998827

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.002	0.002	-44	0.000	-44					1
0.2	0.200	0.213	0.013	987	0.0 %	987					6
2.5	2.500	2.535	0.035	12334	0.0 %	12334					1
5.0	5.000	4.886	-0.114	23822	0.0 %	23822					-2
7.5	7.500	7.569	0.069	36932	0.0 %	36932					1
10.0	10.000	9.994	-0.006	48785	0.0 %	48785					0

LB
715117

LB88518
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-44	-		0 ISM02.4	PPB	7/5/2017 14:39	Std
	0.2 S0.2	987	-		0.2 ISM02.4	PPB	7/5/2017 14:41	Std
	2.5 S2.5	12334	-		2.5 ISM02.4	PPB	7/5/2017 14:43	Std
	5 S5	23822	-		5 ISM02.4	PPB	7/5/2017 14:45	Std
	7.5 S7.5	36932	-		7.5 ISM02.4	PPB	7/5/2017 14:48	Std
	10 S10	48785	-		10 ISM02.4	PPB	7/5/2017 14:50	Std
ICV	ICV	18307	3.7575 -		ISM02.4	PPB	7/5/2017 14:53	SMPL
ICB	ICB	-155	-0.0204 -		ISM02.4	PPB	7/5/2017 14:55	SMPL
CCV045	CCV045	23562	4.8329 -		ISM02.4	PPB	7/5/2017 14:57	SMPL
CCB045	CCB045	-167	-0.0229 -		ISM02.4	PPB	7/5/2017 15:00	SMPL
PB100413BL	PBW003	147	0.0414 -		ISM02.4	PPB	7/5/2017 15:02	SMPL
I3855-01	MBDBS6	126	0.0371 -		ISM02.4	PPB	7/5/2017 15:04	SMPL
I3855-02	MBDBS6D	139	0.0397 -		ISM02.4	PPB	7/5/2017 15:06	SMPL
I3855-03	MBDBS6S	5082	1.0512 -		ISM02.4	PPB	7/5/2017 15:09	SMPL
PB100414BL	PBW003	143	0.0405 -		ISM02.4	PPB	7/5/2017 15:11	SMPL
I3862-02	MDADY7	23975	4.9174 -		ISM02.4	PPB	7/5/2017 15:13	SMPL
I3862-04	MDADZ8	45	0.0205 -		ISM02.4	PPB	7/5/2017 15:16	SMPL
I3862-05	MDADZ9	134	0.0387 -		ISM02.4	PPB	7/5/2017 15:18	SMPL
I3862-06	MDADZ9D	129	0.0377 -		ISM02.4	PPB	7/5/2017 15:20	SMPL
I3862-07	MDADZ9S	5108	1.0565 -		ISM02.4	PPB	7/5/2017 15:22	SMPL
PB100415BL	PBW003	147	0.0414 -		ISM02.4	PPB	7/5/2017 15:25	SMPL
I3891-01	MA4XC8	146	0.0411 -		ISM02.4	PPB	7/5/2017 15:27	SMPL
I3891-02	MA4XC8D	159	0.0438 -		ISM02.4	PPB	7/5/2017 15:29	SMPL
I3891-03	MA4XC8S	4913	1.0166 -		ISM02.4	PPB	7/5/2017 15:31	SMPL
I3891-04	MA4XC9	181	0.0483 -		ISM02.4	PPB	7/5/2017 15:34	SMPL
I3891-05	MA4XD1	144	0.0407 -		ISM02.4	PPB	7/5/2017 15:36	SMPL
I3891-06	MA4XD2	149	0.0418 -		ISM02.4	PPB	7/5/2017 15:38	SMPL
I3891-07	MA4XD3	149	0.0418 -		ISM02.4	PPB	7/5/2017 15:41	SMPL
I3891-08	MA4XD4	147	0.0414 -		ISM02.4	PPB	7/5/2017 15:43	SMPL
I3891-09	MA4XD5	150	0.042 -		ISM02.4	PPB	7/5/2017 15:45	SMPL
CCV046	CCV046	23064	4.731 -		ISM02.4	PPB	7/5/2017 15:47	SMPL
CCB046	CCB046	-164	-0.0223 -		ISM02.4	PPB	7/5/2017 15:50	SMPL
I3891-10	MA4XD6	132	0.0383 -		ISM02.4	PPB	7/5/2017 15:52	SMPL
I3891-11	MA4XD7	139	0.0397 -		ISM02.4	PPB	7/5/2017 15:54	SMPL
I3891-12	MA4XD8	187	0.0495 -		ISM02.4	PPB	7/5/2017 15:57	SMPL
I3891-13	MA4XD9	136	0.0391 -		ISM02.4	PPB	7/5/2017 15:59	SMPL
I3891-14	MA4XE1	142	0.0403 -		ISM02.4	PPB	7/5/2017 16:01	SMPL
I3891-15	MA4XE2	137	0.0393 -		ISM02.4	PPB	7/5/2017 16:03	SMPL
I3891-16	MA4XE4	131	0.0381 -		ISM02.4	PPB	7/5/2017 16:06	SMPL
I3891-17	MA4XE5	126	0.0371 -		ISM02.4	PPB	7/5/2017 16:08	SMPL
I3891-18	MA4XE6	136	0.0391 -		ISM02.4	PPB	7/5/2017 16:10	SMPL
I3891-19	MA4XE7	143	0.0405 -		ISM02.4	PPB	7/5/2017 16:13	SMPL
I3891-20	MA4XE8	168	0.0457 -		ISM02.4	PPB	7/5/2017 16:15	SMPL
I3891-21	MA4XE9	116	0.035 -		ISM02.4	PPB	7/5/2017 16:17	SMPL
I3891-22	MA4XF1	128	0.0375 -		ISM02.4	PPB	7/5/2017 16:19	SMPL
CCV047	CCV047	23745	4.8703 -		ISM02.4	PPB	7/5/2017 16:22	SMPL

LB88518
INSTRUMENT ID : CV1

CCB047	CCB047	-171	-0.0237 -	ISM02.4	PPB	7/5/2017 16:24	SMPL
PB100417BL	PBW003	136	0.0391 -	ISM02.4	PPB	7/5/2017 16:26	SMPL
I3896-10	ME7FD5	197	0.0516 -	ISM02.4	PPB	7/5/2017 16:28	SMPL
I3896-11	ME7FD5D	171	0.0463 -	ISM02.4	PPB	7/5/2017 16:31	SMPL
I3896-12	ME7FD5S	5006	1.0357 -	ISM02.4	PPB	7/5/2017 16:33	SMPL
I3896-13	ME7FD6	105	0.0328 -	ISM02.4	PPB	7/5/2017 16:35	SMPL
I3896-14	ME7FD7	142	0.0403 -	ISM02.4	PPB	7/5/2017 16:38	SMPL
I3896-15	ME7FD8	171	0.0463 -	ISM02.4	PPB	7/5/2017 16:40	SMPL
I3896-16	ME7FD9	125	0.0369 -	ISM02.4	PPB	7/5/2017 16:42	SMPL
I3896-17	ME7FE0	132	0.0383 -	ISM02.4	PPB	7/5/2017 16:44	SMPL
I3896-18	ME7FE1	128	0.0375 -	ISM02.4	PPB	7/5/2017 16:47	SMPL
I3896-19	ME7FE2	29	0.0172 -	ISM02.4	PPB	7/5/2017 16:49	SMPL
PB100418BL	PBW003	120	0.0358 -	ISM02.4	PPB	7/5/2017 16:51	SMPL
I3918-01	MA4XC3	134	0.0387 -	ISM02.4	PPB	7/5/2017 16:54	SMPL
I3918-02	MA4XG0	16082	3.3022 -	ISM02.4	PPB	7/5/2017 16:56	SMPL
I3918-03	MA4XG1	15912	3.2674 -	ISM02.4	PPB	7/5/2017 16:58	SMPL
I3918-08	MA4XH2	97	0.0311 -	ISM02.4	PPB	7/5/2017 17:00	SMPL
I3918-09	MA4XH3	107	0.0332 -	ISM02.4	PPB	7/5/2017 17:03	SMPL
I3918-10	MA4XJ5	117	0.0352 -	ISM02.4	PPB	7/5/2017 17:05	SMPL
CCV048	CCV048	23013	4.7205 -	ISM02.4	PPB	7/5/2017 17:07	SMPL
CCB048	CCB048	-154	-0.0202 -	ISM02.4	PPB	7/5/2017 17:10	SMPL
I3918-11	MA4XK8	127	0.0373 -	ISM02.4	PPB	7/5/2017 17:12	SMPL
I3918-12	MA4XK9	162	0.0444 -	ISM02.4	PPB	7/5/2017 17:14	SMPL
I3918-13	MA4XL0	158	0.0436 -	ISM02.4	PPB	7/5/2017 17:16	SMPL
I3918-14	MA4XL6	139	0.0397 -	ISM02.4	PPB	7/5/2017 17:19	SMPL
I3918-15	MA4XL7	172	0.0465 -	ISM02.4	PPB	7/5/2017 17:21	SMPL
I3918-16	MA4XL8	146	0.0411 -	ISM02.4	PPB	7/5/2017 17:23	SMPL
I3918-17	MA4XL8D	144	0.0407 -	ISM02.4	PPB	7/5/2017 17:26	SMPL
I3918-18	MA4XL8S	5054	1.0455 -	ISM02.4	PPB	7/5/2017 17:28	SMPL
I3918-19	MA4XM2	148	0.0416 -	ISM02.4	PPB	7/5/2017 17:30	SMPL
PB100425BL	PBW003	131	0.0381 -	ISM02.4	PPB	7/5/2017 17:32	SMPL
I3919-01	MA4XK6	122	0.0362 -	ISM02.4	PPB	7/5/2017 17:35	SMPL
I3919-02	MA4XK7	105	0.0328 -	ISM02.4	PPB	7/5/2017 17:37	SMPL
I3919-03	MA4XL1	118	0.0354 -	ISM02.4	PPB	7/5/2017 17:39	SMPL
I3919-04	MA4XL2	133	0.0385 -	ISM02.4	PPB	7/5/2017 17:42	SMPL
I3919-05	MA4XL3	156	0.0432 -	ISM02.4	PPB	7/5/2017 17:44	SMPL
I3919-06	MA4XM1	190	0.0502 -	ISM02.4	PPB	7/5/2017 17:46	SMPL
I3919-07	MA4XM1D	203	0.0528 -	ISM02.4	PPB	7/5/2017 17:48	SMPL
I3919-08	MA4XM1S	5161	1.0674 -	ISM02.4	PPB	7/5/2017 17:51	SMPL
I3943-01	MBDC79	94401	19.329 -	ISM02.4	PPB	7/5/2017 17:53	SMPL
CCV049	CCV049	23403	4.8003 -	ISM02.4	PPB	7/5/2017 17:58	SMPL
CCB049	CCB049	-176	-0.0247 -	ISM02.4	PPB	7/5/2017 18:00	SMPL