



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88533

Temp.: 23.1 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.00	07/05/2017	16:40
2	CAL2	1	Water			7.01	07/05/2017	16:41
3	CAL3	1	Water			10.02	07/05/2017	16:42
4	ICV	1	Water			7.00	07/05/2017	16:44
5	CCV1	1	Water			2.02	07/05/2017	16:45
6	I3992-01	1	Solid	20.03	20	8.87	07/05/2017	16:46
7	I3992-01DUP	1	Solid	20.04	20	8.88	07/05/2017	16:47
8	CCV2	1	Water			12.02	07/05/2017	16:48

LB88533

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I3992

WorkList ID : 100394

Date : 7/5/2017 4:29:16 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/29/2017	Solid	I3992-01	pH	Cool 4 deg C	GLOB02	P63	CLINTON-NJ-QUARRY-STON	06/29/2017	9045C

Date/Time 07/05/17 4:30 PM
Received by: JS
Relinquished by: CS

Date/Time 07/05/17 6:30 PM
Received by: JS
Relinquished by: JS