



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88537

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.6

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/05/2017	18:05
2	CAL2	1	Water			6.99	07/05/2017	18:06
3	CAL3	1	Water			10.03	07/05/2017	18:07
4	ICV	1	Water			7.00	07/05/2017	18:09
5	CCV1	1	Water			2.02	07/05/2017	18:10
6	I4061-03	1	Solid	20.03	20	10.39	07/05/2017	18:11
7	I4061-03DUP	1	Solid	20.05	20	10.40	07/05/2017	18:12
8	I4061-06	1	Solid	20.04	20	9.65	07/05/2017	18:15
9	I4061-09	1	Solid	20.01	20	10.93	07/05/2017	18:20
10	I4061-12	1	Solid	20.02	20	8.44	07/05/2017	18:21
11	I4062-03	1	Solid	20.04	20	9.20	07/05/2017	18:22
12	I4063-03	1	Solid	20.05	20	8.80	07/05/2017	18:25
13	I4063-06	1	Solid	20.06	20	8.78	07/05/2017	18:26
14	I4063-09	1	Solid	20.04	20	8.85	07/05/2017	18:27
15	CCV2	1	Water			12.01	07/05/2017	18:30

288537

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4061

WorkList ID : 100398

Date : 7/5/2017 5:59:40 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/12/2017	Solid	I4061-03	Corrosivity	Cool 4 deg C	PSEG01	E41	SP-1-A	07/05/2017	9045C
07/12/2017	Solid	I4061-06	Corrosivity	Cool 4 deg C	PSEG01	E41	SP-1-B	07/05/2017	9045C
07/12/2017	Solid	I4061-09	Corrosivity	Cool 4 deg C	PSEG01	E41	SP-1-C	07/05/2017	9045C
07/12/2017	Solid	I4061-12	Corrosivity	Cool 4 deg C	PSEG01	E41	SP-1-D	07/05/2017	9045C
07/10/2017	Solid	I4062-03	Corrosivity	Cool 4 deg C	PSEG05	E41	SU-02-070517	07/05/2017	9045C
07/10/2017	Solid	I4063-03	Corrosivity	Cool 4 deg C	PSEG05	E42	NB-01-070517-A	07/05/2017	9045C
07/10/2017	Solid	I4063-06	Corrosivity	Cool 4 deg C	PSEG05	E42	NB-01-070517-B	07/05/2017	9045C
07/10/2017	Solid	I4063-09	Corrosivity	Cool 4 deg C	PSEG05	E42	NB-01-070517-C	07/05/2017	9045C

Date/Time 07/05/17 6:00 PM

Received by: 70

Relinquished by:

Date/Time 07/05/17 7:30 PM

Received by: CP

Relinquished by: JP