



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB88538

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP56166
COD calibration std. 100 ppm	WP56164
COD calibration std. 10 ppm	WP56162
COD calibration std. 150 ppm	WP56165
COD calibration std. 50 ppm	WP56163
COD calibration std. 0 ppm	WP56161
COD CCV std, 50ppm	WP56167
COD Digestion Vials Low Level 0-150Mg/L	W2235

Temp In (C): <u>150</u>	Date In: <u>07/05/2017</u>	Time In: <u>13:10</u>
Temp Out (C): <u>150</u>	Date Out: <u>07/05/2017</u>	Time Out: <u>15:12</u>

Intercept: 0.2985

Slope: 1.0016

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	07/05/2017	16:00
2	CAL2	10	1	Water	10.000	07/05/2017	16:00
3	CAL3	50	1	Water	51.000	07/05/2017	16:01
4	CAL4	100	1	Water	101.000	07/05/2017	16:01
5	CAL5	150	1	Water	150.000	07/05/2017	16:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB88538

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.622	07/05/2017	16:02
2	ICB				1	Water	0.000	-0.298	07/05/2017	16:03
3	CCV1	50			1	Water	51.000	50.621	07/05/2017	16:03
4	CCB1				1	Water	0.000	-0.298	07/05/2017	16:04
5	LB88538BL				1	Water	0.000	-0.298	07/05/2017	16:04
6	LB88538BS	50			1	Water	50.000	49.622	07/05/2017	16:05
7	I4000-01				1	Water	12.000	11.683	07/05/2017	16:05
8	I4000-01DUP				1	Water	11.000	10.684	07/05/2017	16:06
9	I4000-01MS	50			1	Water	63.000	62.601	07/05/2017	16:06
10	I4000-01MSD	50			1	Water	62.000	61.603	07/05/2017	16:07
11	I4020-01				1	Water	3.000	2.697	07/05/2017	16:07
12	I4020-04				1	Water	5.000	4.694	07/05/2017	16:08
13	CCV2	50			1	Water	52.000	51.619	07/05/2017	16:08
14	CCB2				1	Water	0.000	-0.298	07/05/2017	16:08

2388538

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD070517

WorkList ID : 100390

Date : 7/5/2017 12:29:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/04/2017	Water	I4000-01	COD	Conc H2SO4 4	PSEG01	P32	A1273-FT	06/29/2017	SM5220 D
07/05/2017	Water	I4020-01	COD	Conc H2SO4 4	PSEG01	D41	HDD-WATER	06/30/2017	SM5220 D
07/05/2017	Water	I4020-04	COD	Conc H2SO4 4	PSEG01	D41	HDD-SHEEN	06/30/2017	SM5220 D

Date/Time 07/05/17 12:40
 Received by: IB
 Relinquished by: [Signature]

Date/Time 07/05/17 14:00
 Received by: [Signature]
 Relinquished by: IB