


ISM02.4

LB 88545
INSTRUMENT ID: CV2

Linear

 μ Abs.:

36447



A= 0.0000e+000

B= 2.7321e-004

C= 3.9569e-002

Rho= 0.9999168

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.024	-0.024	-232	0.000	-232				
0.2	0.200	0.216	0.016	645	0.0 %	645				
2.5	2.500	2.558	0.058	9218	0.0 %	9218				
5.0	5.000	4.912	-0.088	17834	0.0 %	17834				
7.5	7.500	7.541	0.041	27456	0.0 %	27456				
10.0	10.000	9.997	-0.003	36447	0.0 %	36447				

900
1
8
2
2
1
0NB
716115

LB88545
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-232	-		0 ISM02.4	PPB	7/6/2017 11:37	Std
	0.2 S0.2	645	-		0.2 ISM02.4	PPB	7/6/2017 11:39	Std
	2.5 S2.5	9218	-		2.5 ISM02.4	PPB	7/6/2017 11:41	Std
	5 S5	17834	-		5 ISM02.4	PPB	7/6/2017 11:43	Std
	7.5 S7.5	27456	-		7.5 ISM02.4	PPB	7/6/2017 11:46	Std
	10 S10	36447	-		10 ISM02.4	PPB	7/6/2017 11:48	Std
ICV	ICV	13332	3.682	-	ISM02.4	PPB	7/6/2017 11:51	SMPL
ICB	ICB	-314	-0.0462	-	ISM02.4	PPB	7/6/2017 11:53	SMPL
CCV050	CCV050	18003	4.9582	-	ISM02.4	PPB	7/6/2017 11:55	SMPL
CCB050	CCB050	-301	-0.0427	-	ISM02.4	PPB	7/6/2017 11:58	SMPL
PB100426BL	PBW003	-191	-0.0126	-	ISM02.4	PPB	7/6/2017 12:00	SMPL
I4015-01	MBDWS9	4859	1.3671	-	ISM02.4	PPB	7/6/2017 12:02	SMPL
I4015-02	MBDWS9D	4845	1.3633	-	ISM02.4	PPB	7/6/2017 12:04	SMPL
I4015-03	MBDWS9S	9182	2.5482	-	ISM02.4	PPB	7/6/2017 12:07	SMPL
I4015-04DLX100	MBDWT0	343447	HIGH	-	ISM02.4	PPB	7/6/2017 12:09	
I4015-05DLX100	MBDWT1	346671	HIGH	-	ISM02.4	PPB	7/6/2017 12:11	
I4015-06DLX100	MBDWT2	757043	HIGH	-	ISM02.4	PPB	7/6/2017 12:13	
I4015-07DLX100	MBDWT3	1182729	HIGH	-	ISM02.4	PPB	7/6/2017 12:15	
I4015-08DLX100	MBDWT4	230980	63.1454	-	ISM02.4	PPB	7/6/2017 12:18	SMPL
I4015-09DLX100	MBDWT5	315093	86.1259	-	ISM02.4	PPB	7/6/2017 12:25	SMPL
I4015-10DLX100	MBDWT6	85835	23.4905	-	ISM02.4	PPB	7/6/2017 12:28	SMPL
I4015-09DL2X2000	MBDWT5	19740	5.4327	-	ISM02.4	PPB	7/6/2017 12:38	SMPL
CCV051	CCV051	18093	4.9827	-	ISM02.4	PPB	7/6/2017 12:41	SMPL
CCB051	CCB051	-422	-0.0757	-	ISM02.4	PPB	7/6/2017 12:43	SMPL
I4015-11DLX100	MBDWT7	162157	44.3424	-	ISM02.4	PPB	7/6/2017 12:46	SMPL
I4015-12DLX100	MBDWT8	39230	10.7576	-	ISM02.4	PPB	7/6/2017 12:48	SMPL
I4015-15DLX100	MBDWW1	112060	30.6554	-	ISM02.4	PPB	7/6/2017 12:55	SMPL
I4015-17DLX100	MBDWW3	14733	4.0648	-	ISM02.4	PPB	7/6/2017 12:59	SMPL
I4015-19DLX100	MBDWW5	14804	4.0842	-	ISM02.4	PPB	7/6/2017 13:04	SMPL
CCV052	CCV052	18676	5.142	-	ISM02.4	PPB	7/6/2017 13:06	SMPL
CCB052	CCB052	-225	-0.0219	-	ISM02.4	PPB	7/6/2017 13:09	SMPL
I4015-04DLX1000	MBDWT0	47094	12.9061	-	ISM02.4	PPB	7/6/2017 13:11	SMPL
I4015-05DLX1000	MBDWT1	48110	13.1837	-	ISM02.4	PPB	7/6/2017 13:13	SMPL
I4015-06DLX1000	MBDWT2	159339	43.5725	-	ISM02.4	PPB	7/6/2017 13:15	SMPL
I4015-07DLX1000	MBDWT3	578014	HIGH	-	ISM02.4	PPB	7/6/2017 13:18	
I4015-08DL2X1000	MBDWT4	27114	7.4474	-	ISM02.4	PPB	7/6/2017 13:20	SMPL
I4015-10DL2X1000	MBDWT6	9738	2.7001	-	ISM02.4	PPB	7/6/2017 13:25	SMPL
I4015-11DL2X1000	MBDWT7	16723	4.6084	-	ISM02.4	PPB	7/6/2017 13:27	SMPL
I4015-12DL2X500	MBDWT8	7839	2.1813	-	ISM02.4	PPB	7/6/2017 13:29	SMPL
I4015-15DL2X1000	MBDWW1	11418	3.1591	-	ISM02.4	PPB	7/6/2017 13:32	SMPL
I4015-04DL2X5000	MBDWT0	9917	2.749	-	ISM02.4	PPB	7/6/2017 13:34	SMPL
I4015-05DL2X5000	MBDWT1	9492	2.6329	-	ISM02.4	PPB	7/6/2017 13:36	SMPL
I4015-06DL2X10000	MBDWT2	16898	4.6563	-	ISM02.4	PPB	7/6/2017 13:39	SMPL
I4015-07DL2X100000	MBDWT3	10289	2.8506	-	ISM02.4	PPB	7/6/2017 13:41	SMPL
CCV053	CCV053	18065	4.9751	-	ISM02.4	PPB	7/6/2017 13:43	SMPL
CCB053	CCB053	-137	0.0021	-	ISM02.4	PPB	7/6/2017 13:45	SMPL

LB88545

INSTRUMENT ID : CV1

I4015-13DLX50	MBDWT9	25292	6.9496 -	ISM02.4	PPB	7/6/2017 13:56	SMPL
I4015-14DLX25	MBDWW0	18124	4.9912 -	ISM02.4	PPB	7/6/2017 13:58	SMPL
I4015-16DLX10	MBDWW2	20672	5.6874 -	ISM02.4	PPB	7/6/2017 14:00	SMPL
I4015-18DLX10	MBDWW4	10354	2.8684 -	ISM02.4	PPB	7/6/2017 14:03	SMPL
I3943-01DLX5	MBDC79	14372	3.9661 -	ISM02.4	PPB	7/6/2017 14:05	SMPL
CCV054	CCV054	17943	4.9418 -	ISM02.4	PPB	7/6/2017 14:07	SMPL
CCB054	CCB054	-207	-0.017 -	ISM02.4	PPB	7/6/2017 14:09	SMPL