

LB 88563


7470A

INSTRUMENT ID : CV8

Linear

 μ Abs.:

34126



A= 0.0000e+000

B= 2.9503e-004

C= 6.1206e-002

Rho= 0.9997237

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	-119	0.000	-119				
0.2	0.200	0.240	0.040	605	0.0 %	605				
2.5	2.500	2.529	0.029	8365	0.0 %	8365				
5.0	5.000	4.895	-0.105	16384	0.0 %	16384				
7.5	7.500	7.381	-0.119	24809	0.0 %	24809				
10.0	10.000	10.129	0.129	34126	0.0 %	34126				

LB88563 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Jul 2017 16:05:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-119	-	0.00007470A	PPB		06 Jul 2017 16:20:03	S	Std
0.2 - 1	50.2	605	-	0.20007470A	PPB		06 Jul 2017 16:22:12	S	Std
2.5 - 1	52.5	8365	-	2.50007470A	PPB		06 Jul 2017 16:24:20	S	Std
5.0 - 1	55	16384	-	5.00007470A	PPB		06 Jul 2017 16:26:28	S	Std
7.5 - 1	59.5	24809	-	7.50007470A	PPB		06 Jul 2017 16:28:36	S	Std
10.0 - 1	810	34126	-	10.00007470A	PPB		06 Jul 2017 16:35:33	S	Std
ICV50 - 1	ICV50	12381	3.7140	-7470A	PPB		06 Jul 2017 16:38:04	U	SMPL
ICB50 - 1	ICB50	-176	0.0093	-7470A	PPB		06 Jul 2017 16:40:11	U	SMPL
CCV55 - 1	CCV55	16615	4.9631	-7470A	PPB		06 Jul 2017 16:42:17	U	SMPL
CCB55 - 1	CCB55	-171	0.0108	-7470A	PPB		06 Jul 2017 16:44:23	U	SMPL
CRA - 1	CRA	623	0.2450	-7470A	PPB		06 Jul 2017 16:46:30	U	SMPL
HighStd - 1	HighStd	34119	10.1274	-7470A	PPB		06 Jul 2017 16:48:36	U	SMPL
ChkStd - 1	ChkStd	22443	6.6826	-7470A	PPB		06 Jul 2017 16:50:43	U	SMPL
PB100431BL - 1	PBW	-203	0.0013	-7470A	PPB		06 Jul 2017 16:52:52	U	SMPL
PB100431BS - 1	LCSW	12138	3.6423	-7470A	PPB		06 Jul 2017 16:55:00	U	SMPL
PB100431TB - 1	PB100391TB	-147	0.0178	-7470A	PPB		06 Jul 2017 16:57:07	U	SMPL
I4035-01 - 1	CIY-SB-WC-S-2-10-50	-68	0.0411	-7470A	PPB		06 Jul 2017 16:59:15	U	SMPL
I4049-03 - 1	RO-72-12000	-45	0.0479	-7470A	PPB		06 Jul 2017 17:01:23	U	SMPL
I4049-06 - 1	DRUM-1-070317	-44	0.0482	-7470A	PPB		06 Jul 2017 17:03:32	U	SMPL
I4052-03 - 1	HR-03-070317-A	-42	0.0488	-7470A	PPB		06 Jul 2017 17:05:39	U	SMPL
CCV56 - 1	CCV56	16668	4.9788	-7470A	PPB		06 Jul 2017 17:07:46	U	SMPL
CCB56 - 1	CCB56	-164	0.0128	-7470A	PPB		06 Jul 2017 17:09:52	U	SMPL
I4052-06 - 1	HR-03-070317-B	-61	0.0432	-7470A	PPB		06 Jul 2017 17:11:59	U	SMPL
I4052-09 - 1	HR-03-070317-C	-35	0.0509	-7470A	PPB		06 Jul 2017 17:14:05	U	SMPL
I4053-03 - 1	CL-02-070317-A	-41	0.0491	-7470A	PPB		06 Jul 2017 17:16:11	U	SMPL
I4053-06 - 1	CL-02-070317-B	-40	0.0494	-7470A	PPB		06 Jul 2017 17:18:18	U	SMPL
I4053-09 - 1	CL-02-070317-C	-30	0.0524	-7470A	PPB		06 Jul 2017 17:20:25	U	SMPL
I4053-12 - 1	CL-02-070317-D	-12	0.0577	-7470A	PPB		06 Jul 2017 17:22:33	U	SMPL
I4053-15 - 1	CL-02-070317-E	-31	0.0521	-7470A	PPB		06 Jul 2017 17:24:41	U	SMPL
I4053-15DUP - 1	CL-02-070317-EDUP	-32	0.0518	-7470A	PPB		06 Jul 2017 17:26:49	U	SMPL
I4053-15MS - 1	CL-02-070317-EMS	12062	3.6199	-7470A	PPB		06 Jul 2017 17:28:57	U	SMPL
I4053-15MSD - 1	CL-02-070317-EMSD	12187	3.6567	-7470A	PPB		06 Jul 2017 17:31:03	U	SMPL
CCV57 - 1	CCV57	16765	5.0074	-7470A	PPB		06 Jul 2017 17:33:10	U	SMPL
CCB57 - 1	CCB57	-170	0.0111	-7470A	PPB		06 Jul 2017 17:35:17	U	SMPL
I4053-15LX5 - 1		-65	0.0420	-7470A	PPB		06 Jul 2017 17:37:24	U	SMPL
I4053-15A - 1		12654	3.7945	-7470A	PPB		06 Jul 2017 17:46:20	U	SMPL
CCV58 - 1	CCV58	16623	4.9655	-7470A	PPB		06 Jul 2017 17:48:26	U	SMPL
CCB58 - 1	CCB58	-172	0.0105	-7470A	PPB		06 Jul 2017 17:50:33	U	SMPL