



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88570

Temp.: 23.3 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/06/2017	16:10
2	CAL2	1	Water			6.99	07/06/2017	16:11
3	CAL3	1	Water			10.02	07/06/2017	16:12
4	ICV	1	Water			7.00	07/06/2017	16:13
5	CCV1	1	Water			2.02	07/06/2017	16:15
6	I4069-07	1	Solid	20.03	20	8.90	07/06/2017	16:16
7	I4069-08	1	Solid	20.04	20	9.28	07/06/2017	16:17
8	I4069-09	1	Solid	20.02	20	10.87	07/06/2017	16:18
9	I4070-03	1	Solid	20.03	20	5.55	07/06/2017	16:20
10	I4070-06	1	Solid	20.04	20	5.58	07/06/2017	16:21
11	I4070-09	1	Solid	20.02	20	6.81	07/06/2017	16:23
12	I4070-12	1	Solid	20.04	20	7.02	07/06/2017	16:25
13	I4070-15	1	Solid	20.01	20	8.45	07/06/2017	16:26
14	I4072-03	1	Solid	20.03	20	6.42	07/06/2017	16:27
15	I4072-06	1	Solid	20.04	20	8.61	07/06/2017	16:30
16	CCV2	1	Water			12.01	07/06/2017	16:31
17	I4075-05	1	Solid	20.02	20	7.36	07/06/2017	16:32
18	I4075-10	1	Solid	20.02	20	7.98	07/06/2017	16:34
19	I4075-10DUP	1	Solid	20.03	20	7.99	07/06/2017	16:36
20	CCV3	1	Water			12.02	07/06/2017	16:38

1988570

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I4070

WorkList ID : 100437

Date : 7/6/2017 2:47:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4069-07	Corrosivity	Cool 4 deg C	DIST01		SB-1	07/05/2017	9045C
	Solid	I4069-08	Corrosivity	Cool 4 deg C	DIST01		SB-2	07/05/2017	9045C
	Solid	I4069-09	Corrosivity	Cool 4 deg C	DIST01		SB-3	07/05/2017	9045C
07/10/2017	Solid	I4070-03	Corrosivity	Cool 4 deg C	PSEG05	E51	AU-05-070517-A	07/05/2017	9045C
07/10/2017	Solid	I4070-06	Corrosivity	Cool 4 deg C	PSEG05	E51	AU-05-070517-B	07/05/2017	9045C
07/10/2017	Solid	I4070-09	Corrosivity	Cool 4 deg C	PSEG05	E51	AU-05-070517-C	07/05/2017	9045C
07/10/2017	Solid	I4070-12	Corrosivity	Cool 4 deg C	PSEG05	E51	AU-05-070517-D	07/05/2017	9045C
07/10/2017	Solid	I4070-15	Corrosivity	Cool 4 deg C	PSEG05	E51	AU-05-070517-E	07/05/2017	9045C
07/10/2017	Solid	I4072-03	Corrosivity	Cool 4 deg C	PSEG05	E52	TR-05-070517-A	07/05/2017	9045C
07/10/2017	Solid	I4072-06	Corrosivity	Cool 4 deg C	PSEG05	E52	TR-05-070517-B	07/05/2017	9045C
07/10/2017	Solid	I4075-05	Corrosivity	Cool 4 deg C	PSEG01	E52	TP8-MH2104-COMP	07/05/2017	9045C
07/10/2017	Solid	I4075-10	Corrosivity	Cool 4 deg C	PSEG01	E52	TP7-MH2103-COMP	07/05/2017	9045C

Date/Time 07/06/17 2:17:11PM

Received by: JP

Relinquished by:

Date/Time 07/06/17 6:30PM

Received by: CP

Relinquished by: JP