



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88571

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/06/2017	18:15
2	CAL2	1	Water			7.00	07/06/2017	18:16
3	CAL3	1	Water			10.03	07/06/2017	18:17
4	ICV	1	Water			7.01	07/06/2017	18:19
5	CCV1	1	Water			2.01	07/06/2017	18:20
6	I4092-01	1	Solid	20.03	20	5.71	07/06/2017	18:21
7	I4092-01DUP	1	Solid	20.04	20	5.72	07/06/2017	18:22
8	I4095-01	1	Solid	20.05	20	5.93	07/06/2017	18:25
9	CCV2	1	Water			12.01	07/06/2017	18:26

LB88571

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4092

WorkList ID : 100452

Date : 7/6/2017 PM 5:59:10 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4092-01-A	pH	Cool 4 deg C	TULL02	D61	FK-PS-01-010	07/06/2017	9045C
	Solid	I4095-01-A	pH	Cool 4 deg C	TULL02	D61	FK-PS-01-011	07/06/2017	9045C

Date/Time 07/06/17 6:00 AM
Received by: JP
Relinquished by:

Date/Time 07/06/17 7:20 AM
Received by: Q
Relinquished by: JP