



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88579

Temp.: 21.2 °C

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/07/2017	10:00
2	CAL2	1	Water			7.01	07/07/2017	10:01
3	CAL3	1	Water			10.03	07/07/2017	10:02
4	ICV	1	Water			6.99	07/07/2017	10:03
5	CCV1	1	Water			2.02	07/07/2017	10:05
6	I4086-01	1	Water			7.36	07/07/2017	10:06
7	I4086-01DUP	1	Water			7.37	07/07/2017	10:07
8	CCV2	1	Water			12.01	07/07/2017	10:08

1888579

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4086

WorkList ID : 100476

Date : 7/7/2017 9:47:50 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/06/2017	Water	I4086-01-5	pH		PSEG01		A2882	07/06/2017	9040C

Date/Time 07/07/17 9:46 AM
Received by: JP
Relinquished by:

Date/Time 07/07/17 11:00 AM
Received by: CP
Relinquished by: JP