

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB88581

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP55256
calibration std. phosphate 1 ppm	WP56225
calibration std. phosphate 0.5 ppm	WP56224
calibration std. phosphate 0.3 ppm	WP56223
calibration std. 0 ppm	WP56220
phosphate ICV std.	WP56226
phosphate CCV std.	WP56227
calibration std. phosphate 0.1 ppm	WP56222
calibration std. phosphate 0.05 ppm	WP56221
Combined reagent	WP56229
11 N sulfuric acid	WP52745
Phenolphthalein indicator	WP52748
Sodium hydroxide, 1N	WP52749
Ascorbic Acid	WP56228

Intercept: -0.0014Slope: 0.6511Regression: 0.999835

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	07/07/2017	11:01
2	CAL2	0.05	1	50	50	0.032	07/07/2017	11:01
3	CAL3	0.10	1	50	50	0.064	07/07/2017	11:02
4	CAL4	0.30	1	50	50	0.186	07/07/2017	11:02
5	CAL5	0.50	1	50	50	0.330	07/07/2017	11:03
6	CAL6	1.00	1	50	50	0.649	07/07/2017	11:03



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB88581

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.327	0.504	07/07/2017	11:04
2	ICB		1	50	50	0.000	0.002	07/07/2017	11:04
3	CCV1	0.50	1	50	50	0.328	0.506	07/07/2017	11:05
4	CCB1		1	50	50	0.000	0.002	07/07/2017	11:05
5	PB100491BL		1	50	50	0.000	0.002	07/07/2017	11:05
6	PB100491BS	0.50	1	50	50	0.326	0.503	07/07/2017	11:06
7	I3967-01		1	50	50	0.012	0.021	07/07/2017	11:07
8	I3967-01DUP		1	50	50	0.013	0.022	07/07/2017	11:07
9	I3967-01MS	0.50	1	50	50	0.386	0.595	07/07/2017	11:08
10	I3967-01MSD	0.50	1	50	50	0.388	0.598	07/07/2017	11:08
11	I3967-02		1	50	50	0.013	0.022	07/07/2017	11:09
12	I4058-03		1	50	50	1.488	2.288	07/07/2017	11:09
13	I4058-03		10	50	50	0.570	0.878	07/07/2017	11:20
14	CCV2	0.50	1	50	50	0.330	0.509	07/07/2017	11:21
15	CCB2		1	50	50	0.000	0.002	07/07/2017	11:21