



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88588

Temp.: 21.2 °C

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/07/2017	13:00
2	CAL2	1	Water			7.00	07/07/2017	13:01
3	CAL3	1	Water			10.02	07/07/2017	13:02
4	ICV	1	Water			7.00	07/07/2017	13:05
5	CCV1	1	Water			2.02	07/07/2017	13:06
6	I4111-01	1	Water			4.01	07/07/2017	13:07
7	I4111-01DUP	1	Water			4.02	07/07/2017	13:08
8	I4112-01	1	Water			4.22	07/07/2017	13:10
9	CCV2	1	Water			12.01	07/07/2017	13:11

1B88588

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i4111

WorkList ID : 100489

Date : 7/7/2017 12:49:32 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Water	I4111-01 - 1A	pH	Cool 4 deg C	STER14	D21	MINI-CELL	07/05/2017	SM 4500-PH B
	Water	I4112-01 - 1A	pH	Cool 4 deg C	STER04	D12	POOLWATER	07/06/2017	SM 4500-PH B

Date/Time 07/07/17 12:50 PM
Received by: SP
Relinquished by: C8

Date/Time 07/07/17 2:00 PM
Received by: CD
Relinquished by: SP