


ISM02.4

LB88589
INSTRUMENT ID: CV2

Linear

 μ Abs.:

39777



A= 0.0000e+000

B= 2.5112e-004

C= -2.1507e-002

Rho= 0.9999788

Accept=Accepted

Handwritten notes:
 slope
 y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/P
0.00	0.000	-0.025	-0.025	-12	0.000	-12					-3
0.2	0.200	0.194	-0.006	858	0.0 %	858					1
2.5	2.500	2.528	0.028	10152	0.0 %	10152					0
5.0	5.000	5.004	0.004	20013	0.0 %	20013					0
7.5	7.500	7.531	0.031	30075	0.0 %	30075					0
10.0	10.000	9.967	-0.033	39777	0.0 %	39777					

Handwritten notes:
 103
 717117

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-12	-		0 ISM02.4	PPB	7/7/2017 14:38	Std
	0.2 S0.2	858	-		0.2 ISM02.4	PPB	7/7/2017 14:40	Std
	2.5 S2.5	10152	-		2.5 ISM02.4	PPB	7/7/2017 14:42	Std
	5 S5	20013	-		5 ISM02.4	PPB	7/7/2017 14:45	Std
	7.5 S7.5	30075	-		7.5 ISM02.4	PPB	7/7/2017 14:47	Std
	10 S10	39777	-		10 ISM02.4	PPB	7/7/2017 14:49	Std
ICV	ICV	15443	3.8566 -		ISM02.4	PPB	7/7/2017 14:52	SMPL
ICB	ICB	-175	-0.0655 -		ISM02.4	PPB	7/7/2017 14:54	SMPL
CCV066	CCV066	20057	5.0153 -		ISM02.4	PPB	7/7/2017 14:57	SMPL
CCB066	CCB066	-164	-0.0627 -		ISM02.4	PPB	7/7/2017 14:59	SMPL
PB100460BL	PBS004	-41	-0.0318 -		ISM02.4	PPB	7/7/2017 15:01	SMPL
I3896-01	ME7FC7	13839	3.4538 -		ISM02.4	PPB	7/7/2017 15:03	SMPL
I3896-02	ME7FC8	582	0.1246 -		ISM02.4	PPB	7/7/2017 15:06	SMPL
I3896-03	ME7FC9	482	0.0995 -		ISM02.4	PPB	7/7/2017 15:08	SMPL
I3896-04	ME7FD0	721	0.1596 -		ISM02.4	PPB	7/7/2017 15:10	SMPL
I3896-05	ME7FD0D	711	0.157 -		ISM02.4	PPB	7/7/2017 15:16	SMPL
I3896-06	ME7FD0S	12384	3.0884 -		ISM02.4	PPB	7/7/2017 15:18	SMPL
I3896-07	ME7FD1	501	0.1043 -		ISM02.4	PPB	7/7/2017 15:20	SMPL
I3896-08	ME7FD2	776	0.1734 -		ISM02.4	PPB	7/7/2017 15:23	SMPL
I3896-09	ME7FD3	901	0.2048 -		ISM02.4	PPB	7/7/2017 15:25	SMPL
PB100461BL	PBS004	-16	-0.0255 -		ISM02.4	PPB	7/7/2017 15:27	SMPL
I3964-01	ME7FB1	6132	1.5184 -		ISM02.4	PPB	7/7/2017 15:29	SMPL
I3964-02	ME7FB2	762	0.1698 -		ISM02.4	PPB	7/7/2017 15:32	SMPL
I3964-03	ME7FB3	828	0.1864 -		ISM02.4	PPB	7/7/2017 15:34	SMPL
I3964-04	ME7FB4	1192	0.2778 -		ISM02.4	PPB	7/7/2017 15:36	SMPL
I3964-05	ME7FB4D	1057	0.2439 -		ISM02.4	PPB	7/7/2017 15:41	SMPL
I3964-06	ME7FB4S	11768	2.9337 -		ISM02.4	PPB	7/7/2017 15:43	SMPL
I3964-07	ME7FB5	762	0.1698 -		ISM02.4	PPB	7/7/2017 15:46	SMPL
I3964-08	ME7FB6	68	-0.0044 -		ISM02.4	PPB	7/7/2017 15:48	SMPL
I3964-09	ME7FB7	103	0.0044 -		ISM02.4	PPB	7/7/2017 15:50	SMPL
CCV067	CCV067	20481	5.1217 -		ISM02.4	PPB	7/7/2017 15:52	SMPL
CCB067	CCB067	-149	-0.0589 -		ISM02.4	PPB	7/7/2017 15:55	SMPL
I3964-10	ME7FB8	540	0.1141 -		ISM02.4	PPB	7/7/2017 15:57	SMPL
I3964-11	ME7FB9	524	0.1101 -		ISM02.4	PPB	7/7/2017 15:59	SMPL
I3964-12	ME7FC0	40	-0.0115 -		ISM02.4	PPB	7/7/2017 16:02	SMPL
I3964-13	ME7FC1	346	0.0654 -		ISM02.4	PPB	7/7/2017 16:04	SMPL
I3964-14	ME7FC2	733	0.1626 -		ISM02.4	PPB	7/7/2017 16:06	SMPL
I3964-15	ME7FC3	3365	0.8235 -		ISM02.4	PPB	7/7/2017 16:08	SMPL
I3964-16	ME7FC4	6688	1.658 -		ISM02.4	PPB	7/7/2017 16:11	SMPL
I3964-17	ME7FC5	370	0.0714 -		ISM02.4	PPB	7/7/2017 16:13	SMPL
I3964-18	ME7FC6	372	0.0719 -		ISM02.4	PPB	7/7/2017 16:15	SMPL
CCV068	CCV068	20487	5.1233 -		ISM02.4	PPB	7/7/2017 16:18	SMPL
CCB068	CCB068	-162	-0.0622 -		ISM02.4	PPB	7/7/2017 16:20	SMPL