


7470A

LB885-97
INSTRUMENT ID: CVL

Linear

 μ Abs.:

34027



A= 0.0000e+000

B= 2.9620e-004

C= -7.8720e-002

Rho= 0.9999711

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.047	-0.047	107	0.000	107				
0.2	0.200	0.233	0.033	1052	0.0 %	1052				
2.5	2.500	2.533	0.033	8817	0.0 %	8817				
5.0	5.000	4.981	-0.019	17083	0.0 %	17083				
7.5	7.500	7.500	-0.000	25585	0.0 %	25585				
10.0	10.000	10.000	0.000	34027	0.0 %	34027				

LB88597 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Jul 2017 16:53:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	107	-	0.00007470A	PPB		07 Jul 2017 17:00:48	S	Std
0.2 - 1	50.2	1052	-	0.20007470A	PPB		07 Jul 2017 17:02:57	S	Std
2.5 - 1	52.5	8817	-	2.50007470A	PPB		07 Jul 2017 17:05:05	S	Std
5.0 - 1	55	17083	-	5.00007470A	PPB		07 Jul 2017 17:07:13	S	Std
7.5 - 1	57.5	25585	-	7.50007470A	PPB		07 Jul 2017 17:09:21	S	Std
10.0 - 1	510	34027	-	10.00007470A	PPB		07 Jul 2017 17:19:00	S	Std
ICV52 - 1	ICV52	13167	3.8214	-7470A	PPB		07 Jul 2017 17:21:36	U	SMPL
ICB52 - 1	ICB52	152	-0.0337	-7470A	PPB		07 Jul 2017 17:23:42	U	SMPL
CCV69 - 1	CCV69	17132	4.9959	-7470A	PPB		07 Jul 2017 17:25:49	U	SMPL
CCB69 - 1	CCB69	172	-0.0278	-7470A	PPB		07 Jul 2017 17:27:55	U	SMPL
CRA - 1	CRA	1033	0.2273	-7470A	PPB		07 Jul 2017 17:30:03	U	SMPL
HighStd - 1	HighStd	33722	9.9099	-7470A	PPB		07 Jul 2017 17:32:10	U	SMPL
ChkStd - 1	ChkStd	23760	6.9591	-7470A	PPB		07 Jul 2017 17:34:17	U	SMPL
PB100464BL - 1	PBW	186	-0.0236	-7470A	PPB		07 Jul 2017 17:43:07	U	SMPL
PB100464BS - 1	LCSW	12187	3.5311	-7470A	PPB		07 Jul 2017 17:45:14	U	SMPL
PB100464TB - 1	PB100432TB	167	-0.0293	-7470A	PPB		07 Jul 2017 17:47:22	U	SMPL
I4061-03 - 1	SP-1-A	21	-0.0725	-7470A	PPB		07 Jul 2017 17:49:30	U	SMPL
I4061-06 - 1	SP-1-B	15	-0.0743	-7470A	PPB		07 Jul 2017 17:51:38	U	SMPL
I4061-09 - 1	SP-1-C	25	-0.0713	-7470A	PPB		07 Jul 2017 17:53:46	U	SMPL
I4061-12 - 1	SP-1-D	51	-0.0636	-7470A	PPB		07 Jul 2017 17:55:53	U	SMPL
CCV70 - 1	CCV70	17308	5.0480	-7470A	PPB		07 Jul 2017 17:57:59	U	SMPL
CCB70 - 1	CCB70	184	-0.0242	-7470A	PPB		07 Jul 2017 18:00:06	U	SMPL
I4062-03 - 1	SU-02-070517	54	-0.0627	-7470A	PPB		07 Jul 2017 18:02:14	U	SMPL
I4063-03 - 1	NB-01-070517-A	72	-0.0574	-7470A	PPB		07 Jul 2017 18:04:20	U	SMPL
I4063-06 - 1	NB-01-070517-B	99	-0.0494	-7470A	PPB		07 Jul 2017 18:06:27	U	SMPL
I4063-09 - 1	NB-01-070517-C	125	-0.0417	-7470A	PPB		07 Jul 2017 18:08:34	U	SMPL
I4063-09DUP - 1	NB-01-070517-CDUP	135	-0.0387	-7470A	PPB		07 Jul 2017 18:10:42	U	SMPL
I4063-09MS - 1	NB-01-070517-CMS	12570	3.6446	-7470A	PPB		07 Jul 2017 18:12:49	U	SMPL
I4063-09MSD - 1	NB-01-070517-CMSD	12367	3.5844	-7470A	PPB		07 Jul 2017 18:14:58	U	SMPL
PB100475BL - 1	PBW	202	-0.0189	-7470A	PPB		07 Jul 2017 18:19:31	U	SMPL
PB100475BS - 1	LCSW	12395	3.5927	-7470A	PPB		07 Jul 2017 18:21:39	U	SMPL
I3757-05 - 1	LOD-WATER-05-QT3-2017	354	0.0261	-7470A	PPB		07 Jul 2017 18:28:38	U	SMPL
CCV71 - 1	CCV71	16293	4.7473	-7470A	PPB		07 Jul 2017 18:30:45	U	SMPL
CCB71 - 1	CCB71	142	-0.0367	-7470A	PPB		07 Jul 2017 18:32:52	U	SMPL
I3757-06 - 1	LOQ-WATER-06-QT3-2017	1272	0.2981	-7470A	PPB		07 Jul 2017 18:37:43	U	SMPL
I4014-17 - 1	FL-0500	9	-0.0761	-7470A	PPB		07 Jul 2017 18:39:49	U	SMPL
I4086-01 - 1	A2882	7952	2.2767	-7470A	PPB		07 Jul 2017 18:41:56	U	SMPL
I4086-01DUP - 1	A2882DUP	8122	2.3271	-7470A	PPB		07 Jul 2017 18:44:03	U	SMPL
I4086-01MS - 1	A2882MS	21249	6.2153	-7470A	PPB		07 Jul 2017 18:46:11	U	SMPL
I4086-01MSD - 1	A2882MSD	20768	6.0729	-7470A	PPB		07 Jul 2017 18:48:18	U	SMPL
I4063-09LX5 - 1		167	-0.0293	-7470A	PPB		07 Jul 2017 18:56:26	U	SMPL
I4063-09A - 1		13860	4.0267	-7470A	PPB		07 Jul 2017 18:58:34	U	SMPL
I4086-01LX5 - 1		1344	0.3194	-7470A	PPB		07 Jul 2017 19:00:42	U	SMPL
I4086-01A - 1		22014	6.4419	-7470A	PPB		07 Jul 2017 19:02:49	U	SMPL
CCV72 - 1	CCV72	17182	5.0107	-7470A	PPB		07 Jul 2017 19:04:56	U	SMPL
CCB72 - 1	CCB72	109	-0.0464	-7470A	PPB		07 Jul 2017 19:09:25	U	SMPL