



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB88602

Temp.: 24.3 °C

BalanceID: WC SC-4

SLOPE.: 98.7

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/07/2017	16:30
2	CAL2	1	Water			7.01	07/07/2017	16:31
3	CAL3	1	Water			10.02	07/07/2017	16:32
4	ICV	1	Water			7.00	07/07/2017	16:33
5	CCV1	1	Water			2.01	07/07/2017	16:35
6	I4118-01	1	Solid	20.03	20	9.40	07/07/2017	16:36
7	I4118-01DUP	1	Solid	20.02	20	9.41	07/07/2017	16:37
8	CCV2	1	Water			12.02	07/07/2017	16:38

LB88602

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I4114

WorkList ID : 100501

Date : 7/7/2017 4:16:54 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
	Solid	I4118-01-1A	pH	Cool 4 deg C	NORT14	D32	NAA-PIPE-SAND	07/07/2017	9045C

Date/Time 07/07/17 4:16 PM
 Received by: SO
 Relinquished by: UN

Date/Time 07/07/17 8:09 PM
 Received by: CD
 Relinquished by: SO