


245.1

LB 88609
INSTRUMENT ID: CV1

Linear

 μ Abs.:

29088



A= 0.0000e+000

B= 3.4548e-004

C= 2.2720e-003

Rho= 0.9998664

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.013	-0.013	-43	0.000	-43				
0.2	0.200	0.211	0.011	605	0.0 %	605				
2.5	2.500	2.497	-0.003	7222	0.0 %	7222				
5.0	5.000	5.070	0.070	14670	0.0 %	14670				
7.5	7.500	7.382	-0.118	21361	0.0 %	21361				
10.0	10.000	10.052	0.052	29088	0.0 %	29088				

LB88609 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 07 Jul 2017 19:40:42

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
0.0 - 1	50	-43	-	0.0000	245.1	PPB	07 Jul 2017 19:44:21	S	Std
0.2 - 1	50.2	605	-	0.2000	245.1	PPB	07 Jul 2017 19:46:29	S	Std
2.5 - 1	32.5	7222	-	2.5000	245.1	PPB	07 Jul 2017 19:48:37	S	Std
5.0 - 1	55	14670	-	5.0000	245.1	PPB	07 Jul 2017 19:50:46	S	Std
7.5 - 1	57.5	21361	-	7.5000	245.1	PPB	07 Jul 2017 19:52:53	S	Std
10.0 - 1	51.5	29088	-	10.0000	245.1	PPB	07 Jul 2017 19:55:02	S	Std
ICV53 - 1	ICV53	11583	4.0039	-245.1		PPB	07 Jul 2017 19:57:35	U	SMPL
ICB53 - 1	ICB53	-136	-0.0447	-245.1		PPB	07 Jul 2017 19:59:41	U	SMPL
CCV73 - 1	CCV73	14485	5.0065	-245.1		PPB	07 Jul 2017 20:01:48	U	SMPL
CCB73 - 1	CCB73	-150	-0.0495	-245.1		PPB	07 Jul 2017 20:03:55	U	SMPL
CRA - 1	CRA	605	0.2113	-245.1		PPB	07 Jul 2017 20:06:03	U	SMPL
HighStd - 1	HighStd	29414	10.1642	-245.1		PPB	07 Jul 2017 20:08:09	U	SMPL
ChkStd - 1	ChkStd	20474	7.0756	-245.1		PPB	07 Jul 2017 20:10:16	U	SMPL
PB100482BL - 1	PBW	-191	-0.0637	-245.1		PPB	07 Jul 2017 20:12:24	U	SMPL
PB100482BS - 1	LCSW	12344	4.2668	-245.1		PPB	07 Jul 2017 20:14:34	U	SMPL
I3757-07 - 1	LOD-WATER-07-QT3-2017	105	0.0385	-245.1		PPB	07 Jul 2017 20:19:06	U	SMPL
I3757-08 - 1	LOQ-WATER-08-QT3-2017	1188	0.4127	-245.1		PPB	07 Jul 2017 20:21:14	U	SMPL
I4016-01 - 1	KITCHEN	-99	-0.0319	-245.1		PPB	07 Jul 2017 20:23:21	U	SMPL
I4016-01DUP - 1	KITCHENDUP	-67	-0.0209	-245.1		PPB	07 Jul 2017 20:25:30	U	SMPL
I4016-01MS - 1	KITCHENMS	12430	4.2966	-245.1		PPB	07 Jul 2017 20:27:37	U	SMPL
CCV74 - 1	CCV74	14481	5.0051	-245.1		PPB	07 Jul 2017 20:29:43	U	SMPL
CCB74 - 1	CCB74	-157	-0.0520	-245.1		PPB	07 Jul 2017 20:31:50	U	SMPL
I4016-01MSD - 1	KITCHENMSD	12585	4.3501	-245.1		PPB	07 Jul 2017 20:33:58	U	SMPL
I4016-02 - 1	BATHROOM	-156	-0.0516	-245.1		PPB	07 Jul 2017 20:36:05	U	SMPL
I4055-01 - 1	COMP	107	0.0392	-245.1		PPB	07 Jul 2017 20:38:12	U	SMPL
I4073-01 - 1	32220008-3	-40	-0.0115	-245.1		PPB	07 Jul 2017 20:40:19	U	SMPL
I4080-01 - 1	15220032-1	88	0.0327	-245.1		PPB	07 Jul 2017 20:42:26	U	SMPL
I4016-01LX5 - 1		-61	-0.0188	-245.1		PPB	07 Jul 2017 20:44:33	U	SMPL
I4016-01A - 1		13078	4.5204	-245.1		PPB	07 Jul 2017 20:46:42	U	SMPL
CCV75 - 1	CCV75	14469	5.0010	-245.1		PPB	07 Jul 2017 20:48:50	U	SMPL
CCB75 - 1	CCB75	-186	-0.0620	-245.1		PPB	07 Jul 2017 20:50:58	U	SMPL