


7470A

LB 88620
INSTRUMENT ID: CY 1

Linear

 μ Abs.:

32840



A= 0.0000e+000

B= 3.0336e-004

C= -8.1522e-003

Rho= 0.9999643

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	-50	0.000	-50				
0.2	0.200	0.191	-0.009	655	0.0 %	655				
2.5	2.500	2.503	0.003	8277	0.0 %	8277				
5.0	5.000	5.049	0.049	16669	0.0 %	16669				
7.5	7.500	7.527	0.027	24839	0.0 %	24839				
10.0	10.000	9.954	-0.046	32840	0.0 %	32840				

LB88620 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Jul 2017 12:01:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-50	-	0.0000	7470A	PPB	10 Jul 2017 12:15:53	S	Std
0.2 - 1	50.2	655	-	0.2000	7470A	PPB	10 Jul 2017 12:18:01	S	Std
2.5 - 1	52.5	8277	-	2.5000	7470A	PPB	10 Jul 2017 12:20:09	S	Std
5.0 - 1	55	16669	-	5.0000	7470A	PPB	10 Jul 2017 12:22:18	S	Std
7.5 - 1	57.5	24839	-	7.5000	7470A	PPB	10 Jul 2017 12:24:26	S	Std
10.0 - 1	510	32840	-	10.0000	7470A	PPB	10 Jul 2017 12:29:05	S	Std
ICV54 - 1	ICV54	12542	3.7966	-	7470A	PPB	10 Jul 2017 12:31:38	U	SMPL
ICB54 - 1	ICB54	-163	-0.0576	-	7470A	PPB	10 Jul 2017 12:33:44	U	SMPL
CCV76 - 1	CCV76	16546	5.0113	-	7470A	PPB	10 Jul 2017 12:35:52	U	SMPL
CCB76 - 1	CCB76	-157	-0.0558	-	7470A	PPB	10 Jul 2017 12:37:58	U	SMPL
CRA - 1	CRA	654	0.1902	-	7470A	PPB	10 Jul 2017 12:40:05	U	SMPL
HighStd - 1	HighStd	32746	9.9258	-	7470A	PPB	10 Jul 2017 12:42:12	U	SMPL
ChkStd - 1	ChkStd	22924	6.9461	-	7470A	PPB	10 Jul 2017 12:44:19	U	SMPL
PB100506BL - 1	PBW	-236	-0.0797	-	7470A	PPB	10 Jul 2017 12:46:28	U	SMPL
PB100506BS - 1	LCSW	12694	3.8427	-	7470A	PPB	10 Jul 2017 12:48:37	U	SMPL
PB100506TB - 1	PB100454TB	-128	-0.0470	-	7470A	PPB	10 Jul 2017 12:50:44	U	SMPL
I4069-07 - 1	SB-1	-55	-0.0248	-	7470A	PPB	10 Jul 2017 12:52:53	U	SMPL
I4069-08 - 1	SB-2	-32	-0.0179	-	7470A	PPB	10 Jul 2017 12:55:01	U	SMPL
I4069-09 - 1	SB-3	-10	-0.0112	-	7470A	PPB	10 Jul 2017 12:57:09	U	SMPL
I4070-03 - 1	AU-05-070517-A	-23	-0.0151	-	7470A	PPB	10 Jul 2017 12:59:16	U	SMPL
CCV77 - 1	CCV77	16443	4.9800	-	7470A	PPB	10 Jul 2017 13:01:23	U	SMPL
CCB77 - 1	CCB77	-146	-0.0524	-	7470A	PPB	10 Jul 2017 13:03:30	U	SMPL
I4070-06 - 1	AU-05-070517-B	-50	-0.0233	-	7470A	PPB	10 Jul 2017 13:05:37	U	SMPL
I4070-09 - 1	AU-05-070517-C	-24	-0.0154	-	7470A	PPB	10 Jul 2017 13:07:44	U	SMPL
I4070-12 - 1	AU-05-070517-D	-12	-0.0118	-	7470A	PPB	10 Jul 2017 13:09:51	U	SMPL
I4070-15 - 1	AU-05-070517-E	-26	-0.0160	-	7470A	PPB	10 Jul 2017 13:11:58	U	SMPL
I4072-03 - 1	TR-05-070517-A	-22	-0.0148	-	7470A	PPB	10 Jul 2017 13:14:06	U	SMPL
I4072-06 - 1	TR-05-070517-B	-14	-0.0124	-	7470A	PPB	10 Jul 2017 13:16:13	U	SMPL
I4075-05 - 1	TP8-MH2104-COMP	-41	-0.0206	-	7470A	PPB	10 Jul 2017 13:18:22	U	SMPL
I4075-10 - 1	TP7-MH2103-COMP	-41	-0.0206	-	7470A	PPB	10 Jul 2017 13:20:30	U	SMPL
I4083-02 - 1	STOCKPILE	-15	-0.0127	-	7470A	PPB	10 Jul 2017 13:22:39	U	SMPL
I4083-02DUP - 1	STOCKPILEDUP	-21	-0.0145	-	7470A	PPB	10 Jul 2017 13:24:45	U	SMPL
CCV78 - 1	CCV78	16507	4.9995	-	7470A	PPB	10 Jul 2017 13:26:52	U	SMPL
CCB78 - 1	CCB78	-162	-0.0573	-	7470A	PPB	10 Jul 2017 13:28:59	U	SMPL
I4083-02MS - 1	STOCKPILEMS	12228	3.7014	-	7470A	PPB	10 Jul 2017 13:31:06	U	SMPL
I4083-02MSD - 1	STOCKPILEMSD	12164	3.6820	-	7470A	PPB	10 Jul 2017 13:33:13	U	SMPL
I4083-02LX5 - 1		-175	-0.0612	-	7470A	PPB	10 Jul 2017 13:35:19	U	SMPL
I4083-02A - 1		12889	3.9019	-	7470A	PPB	10 Jul 2017 13:37:26	U	SMPL
CCV79 - 1	CCV79	16793	5.0862	-	7470A	PPB	10 Jul 2017 13:39:34	U	SMPL
CCB79 - 1	CCB79	-189	-0.0655	-	7470A	PPB	10 Jul 2017 13:41:41	U	SMPL