

LB88628

CHEMTECH

284 Sheffield Street, Mountainside, NJ 07092 Phone: 908-789-8900 Fax: 908-789-892

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB88628
Instrument: Appolo-9000

M 9060-A-TOC-19

Analytical Summary Report

Analysis Method: 9060 TOC
Parameter: TOC
Run Number: LB88628
Instrument: Appolo-9000

Seq	Lab ID	Client ID	Sample Type	Inst Conc. ppm/ppb	Dil	Analytical Date
1	BLANK	BLANK	CAL	334110.000	1	7/10/17
2	250mg/L	250mg/L	CAL	1012653.000	1	7/10/17
3	500mg/L	500mg/L	CAL	1980117.000	1	7/10/17
4	1000mg/L	1000mg/L	CAL	3620980.000	1	7/10/17
5	2000mg/L	2000mg/L	CAL	6703403.000	1	7/10/17
6	ICV	ICV	ICV	1058.973	1	7/10/17
7	ICB	ICB	ICB	0.000	1	7/10/17
8	CCV	CCV	CCV	932.931	1	7/10/17
9	CCB	CCB	CCB	0.000	1	7/10/17
10	LB88628BLS	LB88628BLS	MB	4.000	1	7/10/17
11	LB88628BSS	LB88628BSS	LCS	962.000	1	7/10/17
12	I3757-01	LOD-SOIL-01-QT3-2017	SAM	69.000	1	7/10/17
13	I3757-02	LOQ-SOIL-02-QT3-2017	LCS	514.000	1	7/10/17
14	CCV	CCV	CCV	951.686	1	7/10/17
15	CCB	CCB	CCB	0.000	1	7/10/17

* conc are in raw and

Chemtech Consulting Group

Analytical Review Report

Date Printed : 7/10/17
Analyst : AP
Data File : LB88628.MDB

Approved By : 
Approved Date : 7-11-17
Worksheet # :

Lab Sample ID	Client ID		Raw Amt	Dil	Matrix	A. Date	Prep	Analysis				Line 1		
Parameter			PPB		Final Conc	%Rec	Method	Method	LCL	UCL	RPD	Max RPD	Units	Line 2
TOC														
BLANK	BLANK				W	7/10/17								
TOC		PASS	334110.000		334000								mg/L	
250mg/L	250mg/L				W	7/10/17								
TOC		PASS	012653.000		1010000								mg/L	
500mg/L	500mg/L				W	7/10/17								
TOC		PASS	980117.000		1980000								mg/L	
1000mg/L	1000mg/L				W	7/10/17								
TOC		PASS	620980.000		3620000								mg/L	
2000mg/L	2000mg/L				W	7/10/17								
TOC		PASS	703403.000		6700000								mg/L	
ICV1	ICV1				W	7/10/17								
TOC		PASS	1058.973		1060.00	106.0	90	110					mg/L	
ICB1	ICB1				W	7/10/17								
TOC		PASS	0.000		0.000			+/-0.0970					mg/L	
CCV1	CCV1				W	7/10/17								
TOC		PASS	932.931		933.00	93.0	90	110					mg/L	
CCB1	CCB1				W	7/10/17								
TOC		PASS	0.000		0.000			+/-0.0970					mg/L	
LB88628BLS	LB88628BLS				S	7/10/17								
TOC		PASS	4.000		4.000			+/-48.8490					mg/Kg	
LB88628BSS	LB88628BSS				S	7/10/17								
TOC		PASS	962.000		962.00	96.0	90.00	110.00					mg/Kg	
I3757-01	LOD-SOIL-01-QT3-2017			1	S	7/10/17								
TOC		PASS	69.000		69.0								mg/Kg	
I3757-02	LOQ-SOIL-02-QT3-2017				S	7/10/17								
TOC		PASS	514.000		514.00	103.0	50.00	150.00					mg/Kg	
CCV2	CCV2				W	7/10/17								
TOC		PASS	951.686		952.00	95.0	90	110					mg/L	
CCB2	CCB2				W	7/10/17								
TOC		PASS	0.000		0.000			+/-0.0970					mg/L	