



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB88630

Temp.: 23.4 °C

BalanceID: WC SC-4

SLOPE.: 98.3

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2183
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
BUFFER PH 7.00 GREEN 1PINT PK6	W2140
pH 2.00 Buffer	W2079
Buffer Solution, PH12 (500ml)	W2099

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	07/10/2017	15:00
2	CAL2	1	Water			7.01	07/10/2017	15:01
3	CAL3	1	Water			10.03	07/10/2017	15:02
4	ICV	1	Water			7.00	07/10/2017	15:05
5	CCV1	1	Water			2.01	07/10/2017	15:06
6	I4120-05	1	Solid	20.03	20	6.80	07/10/2017	15:07
7	I4120-05DUP	1	Solid	20.05	20	6.81	07/10/2017	15:09
8	I4127-05	1	Solid	20.02	20	8.32	07/10/2017	15:11
9	I4127-06	1	Solid	20.04	20	7.56	07/10/2017	15:12
10	I4132-03	1	Solid	20.02	20	7.96	07/10/2017	15:13
11	CCV2	1	Water			12.01	07/10/2017	15:14

1388630

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I4120			WorkList ID : 100526		Date : 7/10/2017 11:55:06 AM				
Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
07/12/2017	Solid	I4120-05	Corrosivity	Cool 4 deg C	PSEG01	F31	TP-12-COMP	07/07/2017	9045C
	Solid	I4127-05	Corrosivity	Cool 4 deg C	DIST01	D61	SB-7	07/07/2017	9045C
	Solid	I4127-06	Corrosivity	Cool 4 deg C	DIST01	D61	SB-8	07/07/2017	9045C
07/13/2017	Solid	I4132-03	Corrosivity	Cool 4 deg C	PSEG01	F61	CLINTON-AVE-69KV-COMP-4	07/10/2017	9045C

Date/Time 07-10-17 12:00 PM
Received by: [Signature]
Relinquished by: [Signature]

Date/Time 07-10-17 6:00 PM
Received by: [Signature]
Relinquished by: [Signature]