



Analytical Summary Report

Analysis Method: SM4500-H B

Analyst By : Ayul

Parameter: Field pH

Supervisor Review By : apatel

Run Number: LB88631

Temp.: _____ °C

SLOPE.: _____

Calibration Standards	Chemtech Log#
Nitric Acid, Instra-Analyzed (cs/4:	W2187
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2159
buffer solution pH 7 yellow	W2159

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 7.00 Control Limits[+/- 0.1].

True Value of CCV2 = 7.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4	06/28/2017	10:24
2	CAL2	1	Water			7.01	06/28/2017	10:26
3	CAL3	1	Water			10.00	06/28/2017	10:28
4	ICV	1	Water					
5	CCV1	1	Water					
6	I3967-01	1	Water					
7	I3967-02	1	Water					
8	CCV2	1	Water					