


7470A

LAB8635
INSTRUMENT ID: CV1

Linear

 μ Abs.:

32409



A= 0.0000e+000

B= 3.0878e-004

C= 1.5277e-002

Rho= 0.9999887

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.016	0.016	2	0.000	2				
0.2	0.200	0.213	0.013	639	0.0 %	639				
2.5	2.500	2.476	-0.024	7970	0.0 %	7970				
5.0	5.000	4.986	-0.014	16099	0.0 %	16099				
7.5	7.500	7.486	-0.014	24196	0.0 %	24196				
10.0	10.000	10.022	0.022	32409	0.0 %	32409				

LB88635 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Jul 2017 08:46:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	2	-	0.0000	7470A	PPB	11 Jul 2017 09:21:56	S	Std
0.2 - 1	50.2	639	-	0.2000	7470A	PPB	11 Jul 2017 09:24:04	S	Std
2.5 - 1	52.5	7970	-	2.5000	7470A	PPB	11 Jul 2017 09:26:13	S	Std
5.0 - 1	55	16099	-	5.0000	7470A	PPB	11 Jul 2017 09:28:21	S	Std
7.5 - 1	57.5	24196	-	7.5000	7470A	PPB	11 Jul 2017 09:30:29	S	Std
10.0 - 1	510	32409	-	10.0000	7470A	PPB	11 Jul 2017 09:37:26	S	Std
ICV55 - 1	ICV55	12370	3.8349	-	7470A	PPB	11 Jul 2017 09:42:00	U	SMPL
ICB55 - 1	ICB55	-369	-0.0987	-	7470A	PPB	11 Jul 2017 09:44:06	U	SMPL
CCV80 - 1	CCV80	16223	5.0246	-	7470A	PPB	11 Jul 2017 09:46:13	U	SMPL
CCB80 - 1	CCB80	-288	-0.0737	-	7470A	PPB	11 Jul 2017 09:48:20	U	SMPL
CRA - 1	CRA	691	0.2286	-	7470A	PPB	11 Jul 2017 09:50:27	U	SMPL
HighStd - 1	HighStd	32571	10.0725	-	7470A	PPB	11 Jul 2017 09:52:34	U	SMPL
ChkStd - 1	ChkStd	21916	6.7825	-	7470A	PPB	11 Jul 2017 09:54:41	U	SMPL
PB100524BL - 1	PBW	-341	-0.0900	-	7470A	PPB	11 Jul 2017 09:56:49	U	SMPL
PB100524BS - 1	LCSW	12700	3.9368	-	7470A	PPB	11 Jul 2017 10:06:24	U	SMPL
PB100524TB - 1	PB100492TB	-49	0.0001	-	7470A	PPB	11 Jul 2017 10:08:32	U	SMPL
I4071-03 - 1	HR-06-070617-A	-139	-0.0276	-	7470A	PPB	11 Jul 2017 10:10:40	U	SMPL
I4071-06 - 1	HR-06-070617-B	-134	-0.0261	-	7470A	PPB	11 Jul 2017 10:12:49	U	SMPL
I4071-09 - 1	HR-06-070617-C	-102	-0.0162	-	7470A	PPB	11 Jul 2017 10:14:57	U	SMPL
I4078-03 - 1	VNJ-228	-86	-0.0113	-	7470A	PPB	11 Jul 2017 10:17:04	U	SMPL
CCV81 - 1	CCV81	16550	5.1256	-	7470A	PPB	11 Jul 2017 10:19:11	U	SMPL
CCB81 - 1	CCB81	-184	-0.0415	-	7470A	PPB	11 Jul 2017 10:21:18	U	SMPL
I4078-07 - 1	VNJ-211	27	0.0236	-	7470A	PPB	11 Jul 2017 10:23:26	U	SMPL
I4081-02 - 1	VNJ-202	54	0.0320	-	7470A	PPB	11 Jul 2017 10:25:33	U	SMPL
I4082-08 - 1	BIN-6-COMP	-80	-0.0094	-	7470A	PPB	11 Jul 2017 10:27:40	U	SMPL
I4084-05 - 1	TP11-MH2111-COMP	-84	-0.0107	-	7470A	PPB	11 Jul 2017 10:29:47	U	SMPL
I4084-10 - 1	TP9-MH2108-COMP	-100	-0.0156	-	7470A	PPB	11 Jul 2017 10:31:55	U	SMPL
I4086-02 - 1	A2882	-44	0.0017	-	7470A	PPB	11 Jul 2017 10:34:03	U	SMPL
I4106-01 - 1	CIY-WC-S-1-0-10	-95	-0.0141	-	7470A	PPB	11 Jul 2017 10:36:11	U	SMPL
I4108-03 - 1	BU-01-070617-A	-65	-0.0048	-	7470A	PPB	11 Jul 2017 10:38:19	U	SMPL
I4108-06 - 1	BU-01-070617-B	-85	-0.0110	-	7470A	PPB	11 Jul 2017 10:40:28	U	SMPL
I4108-09 - 1	BU-01-070617-C	-65	-0.0048	-	7470A	PPB	11 Jul 2017 10:42:35	U	SMPL
CCV82 - 1	CCV82	16621	5.1475	-	7470A	PPB	11 Jul 2017 10:44:42	U	SMPL
CCB82 - 1	CCB82	-199	-0.0462	-	7470A	PPB	11 Jul 2017 10:46:49	U	SMPL
I4113-07 - 1	SB-4	-85	-0.0110	-	7470A	PPB	11 Jul 2017 10:48:56	U	SMPL
I4113-08 - 1	SB-5	83	0.0409	-	7470A	PPB	11 Jul 2017 10:51:03	U	SMPL
I4113-09 - 1	SB-6	-38	0.0035	-	7470A	PPB	11 Jul 2017 10:53:10	U	SMPL
I4119-03 - 1	EO-03-070717-A	-23	0.0082	-	7470A	PPB	11 Jul 2017 10:55:17	U	SMPL
I4119-06 - 1	EO-03-070717-B	-49	0.0001	-	7470A	PPB	11 Jul 2017 10:57:25	U	SMPL
I4119-09 - 1	EO-03-070717-C	-67	-0.0054	-	7470A	PPB	11 Jul 2017 10:59:32	U	SMPL
I4119-09DUP - 1	EO-03-070717-CDUP	-43	0.0020	-	7470A	PPB	11 Jul 2017 11:01:41	U	SMPL
I4119-09MS - 1	EO-03-070717-CMS	11961	3.7086	-	7470A	PPB	11 Jul 2017 11:03:49	U	SMPL
I4119-09MSD - 1	EO-03-070717-CMSD	12233	3.7926	-	7470A	PPB	11 Jul 2017 11:05:57	U	SMPL
CCV83 - 1	CCV83	16853	5.2191	-	7470A	PPB	11 Jul 2017 11:11:13	U	SMPL
CCB83 - 1	CCB83	-198	-0.0459	-	7470A	PPB	11 Jul 2017 11:13:20	U	SMPL
I4119-09LX5 - 1		-96	-0.0144	-	7470A	PPB	11 Jul 2017 11:15:28	U	SMPL
I4119-09A - 1		12905	4.0001	-	7470A	PPB	11 Jul 2017 11:17:35	U	SMPL
CCV84 - 1	CCV84	16985	5.2599	-	7470A	PPB	11 Jul 2017 11:19:43	U	SMPL
CCB84 - 1	CCB84	-205	-0.0480	-	7470A	PPB	11 Jul 2017 11:21:50	U	SMPL